



Radiological Physics Service, Inc.

3839 Napier Road
Plymouth, MI 48170
(734) 455-4730 Fax: (734) 453-8851

Nov. 11, 2015
~~October 21, 2015~~

PHYSICISTS:

MICHIGAN

Ray A. Carlson, M.S.
Cell: (734) 395-7361
rcarlson@att.net

Laura T. Smith, M.S.

Cell: (586) 808-3058
Pager: (313) 609-2038
lsphysics@att.net

Vincent McCormick, M.S.

Cell: (734) 395-9323
V_mccormick@comcast.net

GEORGIA

Diane Griffiths, M.S.

Cell: (770) 355-7709
dianegriffiths@comcast.net

U. S. NUCLEAR REGULATORY COMMISSION
Region III Materials Licensing Branch
2443 Warrenville Road, Ste 210
Lisle, IL 60532-4352

RE: Follow-up on the #21-17974-01 termination


Thank you for your email on the Tri-County license. Sorry, I got home so late last night

1. I have attached the inventories and leaks tests for before and after move are now attached, I must have either over looked that, or the fax did not go thru properly.
2. The license #21-26253-01 Radiological Physics Service.

If there are any further questions please contact our physicist, Laura T. Smith, MS, DABR at (586) 808-3058.

Sincerely,

Laura T. Smith, MS, DABR
Consulting Physicist



Location: Schoenherr Diag Family Practice Date of Inspection: 10/16/2015

Before Move at location 28351 Schoenherr, Warren, Mi 48088 #21-17974-01
SEALED SOURCE INVENTORY and LEAK TESTS

NUCLIDE	TYPE	LOCATION	CALIBRATION		MFGR	M/N	S/N
			ACTIVITY	DATE			
Cs-137	Vial	Hot Lab	206 uCi	10/26/1977	NEN	NES-356	319
Ba-133	Vial	Hot Lab	258 uCi	9/4/1992	CIS	EGAG50	4309182
Cs-137	Rod	Hot Lab	500 nCi	7/1/2001	IPL		800-4
Cs-137	Vial	Storage	1030 uCi	9/9/1976	NEN	NES-9017	231-210-2
Cs-137	Vial	Storage	199 uCi	12/2/1982	NEN	NES-356	3561282A-42
Cs-137	Vial	Storage	93.9 uCi		NEN		AS-75
Cs-137/ Mo99eq	Vial	Storage	99.3 uCi		NEN		AK-24

Performed by: Laura T. Smith, MS, DABR Date Tested: 10/16/2015

RSO: _____

Comments: The sources listed above were leak tested using an alcohol wipe and were found to have less than 0.005 uCi removable activity.

After Move at location RPS, #21-262-53-01 3839 Napier Road, Plymouth, MI 48312
added to current site INVENTORY and these source were SEALED SOURCE LEAK TESTED

NUCLIDE	TYPE	LOCATION	CALIBRATION		MFGR	M/N	S/N
			ACTIVITY	DATE			
Cs-137	Vial	Hot Lab	206 uCi	10/26/1977	NEN	NES-356	319
Ba-133	Vial	Hot Lab	258 uCi	9/4/1992	CIS	EGAG50	4309182
Cs-137	Rod	Hot Lab	500 nCi	7/1/2001	IPL		800-4
Cs-137	Vial	Storage	1030 uCi	9/9/1976	NEN	NES-9017	231-210-2
Cs-137	Vial	Storage	199 uCi	12/2/1982	NEN	NES-356	3561282A-42
Cs-137	Vial	Storage	93.9 uCi		NEN		AS-75
Cs-137/ Mo99eq	Vial	Storage	99.3 uCi		NEN		AK-24

RSO: _____

Performed by: Laura T. Smith, MS, DABR Date Tested: 10/16/2015

Comments: The sources listed above were leak tested using an alcohol wipe and were found to have less than 0.005 uCi removable activity.

Parker, Bryan

From: Laura T. Smith- Physics <lsphysics@att.net>
Sent: Wednesday, November 11, 2015 4:20 PM
To: Parker, Bryan
Subject: [External_Sender] Re: Re: Termination request - Tri-County Radiologists
Attachments: Schoenherr 2015Decom.pdf; TriCountySchoen NRC decom 2015 2.pdf

I am sorry I did not send this information until today, I got home so late, and then was exhausted, and even though I typed I yesterday, I did not want to send it, because I thought it should be reviewed first.
Thank you for your patience.

Laura Smith I prefer email communications

From: "Parker, Bryan" <Bryan.Parker@nrc.gov>
To: Laura T. Smith- Physics <lsphysics@att.net>
Sent: Tuesday, November 10, 2015 9:30 AM
Subject: RE: Re: Termination request - Tri-County Radiologists

Ok, sounds good. Thank you for the prompt reply.

Bryan

-----Original Message-----

From: Laura T. Smith- Physics [mailto:lsphysics@att.net]
Sent: Tuesday, November 10, 2015 9:29 AM
To: Parker, Bryan <Bryan.Parker@nrc.gov>
Subject: [External_Sender] Re: Termination request - Tri-County Radiologists

Mr. Parker,
I will do that tonight - late (as in after 9pm). I suppose faxing in leaves lots of missed pages, or perhaps I forgot to include them.
Either way, I have all that information you requested, and will send it out via email scanned in tonight.
Thanks,

Laura Smith
I prefer email communications

On Tue, 11/10/15, Parker, Bryan <Bryan.Parker@nrc.gov> wrote:

Subject: Termination request - Tri-County Radiologists
To: "Laura T. Smith- Physics" <lsphysics@att.net>
Date: Tuesday, November 10, 2015, 9:19 AM

Hey
Laura,

I have received the request to terminate Lic. No. 21-17974-01, Tri-County Radiologists and it looks ok; however, Item 4 of letter dated 10/16/15 indicates that leak tests and inventories of sealed sources transferred are attached, but they are not.

Could you please send me those
(or the last set after transfer will suffice)?

Also, could you provide the
license number of where the sources were transferred to, just for the record?

The letter indicates they were transferred to Radiological Physics Service, but no license number is given.

It is fine
to send by e-mail, but
please include a short
signed cover letter with it too.

If you have
any questions, please let me know.

Thanks.

Bryan

Bryan A. Parker
Health
Physicist
USNRC Region
III
bryan.parker@nrc.gov

678-828-7050

630-515-1078