



418 East College Drive
Cheyenne, WY 82007
Tel: 307.638.6648
Fax: 307.637.4315
frminuclear@aol.com

**Attention: Nuclear Materials Licensing Branch
U.S. Nuclear Regulatory Commission, Region IV
1600 E. Lamar Boulevard
Arlington, TX 76011-4511**

Subject: Reference Letter for the attached Amendment Application for Materials License.

This letter is supplied to offer further explanation of line items contained in NRC Form 313 (01-2016). For ease of reference I will number explanations in direct reference to NRC Form 313 (01-2016).

1. This is an amendment application for license number 49-29310-01.
2. We do not request an amendment to this section.
3. We do not request an amendment to this section.
4. We do not request an amendment to this section.
5. We do not request an amendment to this section.
6. We do not request an amendment to this section.

RECEIVED
JAN 25 2016

DNMS

PUBLIC

Immediate Release
 Normal Release

NON-PUBLIC

A.3 Sensitive-Security Related
 A.7 Sensitive Internal
 Other: _____

Reviewer: RLC Date: 2-3-16

h 5 8 9 9 9 9

Page 1 of 3
h 5 8 9 9 9 9

7. The individual(s) responsible for the Radiation Safety Program and their training experience remain the same as previously submitted.

However, we would like to add the following Authorized Users:

<u>Authorized User</u>	<u>Material and Use</u>
Eric B. Hoyer, M.D. <i>Currently listed on Front Range Nuclear Services U.S. NRC license #49-27531-01 for these materials. Please see the attached license.</i>	35.100; 35.200
William Byron Taylor III, M.D. <i>Currently listed on Memorial Hospital of Sheridan County U.S. NRC license #49-10982-02 for these materials. Please see the attached license.</i>	35.100; 35.200

We would like to remove the following Authorized Users:

Robert Jakob Kahn, M.D.	35.100; 35.200
David W. McNaul, M.D.	35.100; 35.200
Robert L. Stears, M.D.	35.100; 35.200
Thomas M. Cunningham III, M.D.	35.100; 35.200
Stephen Horn, M.D.	35.100; 35.200
Paul Peters, M.D.	35.100; 35.200
Geoffrey G. Smith, M.D.	35.100; 35.200
Daniel F. Sulser, M.D.	35.100; 35.200
M.L. Sloan, M.D.	35.100; 35.200
Thomas L. McCallum, M.D.	35.100; 35.200
James R. Taylor, M.D.	35.100; 35.200
Richard D. Clark, M.D.	35.100; 35.200
Mary F. Murphy, M.D.	35.100; 35.200
Jason E. Brown, M.D.	35.100; 35.200
James P. Edlin, M.D.	35.100; 35.200
Steven D. Smith, M.D.	35.100; 35.200

8. We do not request an amendment to this section.
9. We do not request an amendment to this section.
10. We do not request an amendment to this section.
11. We do not request an amendment to this section.

If you have any question, comments, or concerns, please contact me directly at (307) 631-8973 and/or john.mastin@frmiinc.com.

Sincerely,

A handwritten signature in blue ink, appearing to read "John Mastin".

John Mastin, CNMT
Radiation Safety Officer
Chief Operating Officer
Front Range Mobile Imaging
Mile High Mobile PET

Attachments: NRC Form 313 (01-2016)
Front Range Nuclear Services U.S. NRC license #49-27531-01
Memorial Hospital of Sheridan County U.S. NRC license #49-10982-02



APPLICATION FOR MATERIALS LICENSE

Estimated burden per response to comply with this mandatory collection request: 4.3 hours. Submittal of the application is necessary to determine that the applicant is qualified and that adequate procedures exist to protect the public health and safety. Send comments regarding burden estimate to the FOIA, Privacy, and Information Collections Branch (T-5 F53), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to Infocollects.Resource@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3150-0120), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

INSTRUCTIONS: SEE THE APPROPRIATE LICENSE APPLICATION GUIDE FOR DETAILED INSTRUCTIONS FOR COMPLETING APPLICATION. SEND TWO COPIES OF THE ENTIRE COMPLETED APPLICATION TO THE NRC OFFICE SPECIFIED BELOW. *AMENDMENTS/RENEWALS THAT INCREASE THE SCOPE OF THE EXISTING LICENSE TO A NEW OR HIGHER FEE CATEGORY WILL REQUIRE A FEE.

APPLICATION FOR DISTRIBUTION OF EXEMPT PRODUCTS FILE APPLICATIONS WITH:

MATERIALS SAFETY LICENSING BRANCH
 DIVISION OF MATERIAL SAFETY, STATE, TRIBAL AND RULEMAKING PROGRAMS
 OFFICE OF NUCLEAR MATERIALS SAFETY AND SAFEGUARDS
 U.S. NUCLEAR REGULATORY COMMISSION
 WASHINGTON, DC 20555-0001

ALL OTHER PERSONS FILE APPLICATIONS AS FOLLOWS:

IF YOU ARE LOCATED IN:

ALABAMA, CONNECTICUT, DELAWARE, DISTRICT OF COLUMBIA, FLORIDA, GEORGIA, KENTUCKY, MAINE, MARYLAND, MASSACHUSETTS, NEW HAMPSHIRE, NEW JERSEY, NEW YORK, NORTH CAROLINA, PENNSYLVANIA, PUERTO RICO, RHODE ISLAND, SOUTH CAROLINA, TENNESSEE, VERMONT, VIRGINIA, VIRGIN ISLANDS, OR WEST VIRGINIA,

SEND APPLICATIONS TO:

LICENSING ASSISTANCE TEAM
 DIVISION OF NUCLEAR MATERIALS SAFETY
 U.S. NUCLEAR REGULATORY COMMISSION, REGION I
 2100 RENAISSANCE BOULEVARD, SUITE 100
 KING OF PRUSSIA, PA 19406-2713

IF YOU ARE LOCATED IN:

ILLINOIS, INDIANA, IOWA, MICHIGAN, MINNESOTA, MISSOURI, OHIO, OR WISCONSIN,
SEND APPLICATIONS TO:

MATERIALS LICENSING BRANCH
 U.S. NUCLEAR REGULATORY COMMISSION, REGION III
 2443 WARRENVILLE ROAD, SUITE 210
 LISLE, IL 60532-4352

ALASKA, ARIZONA, ARKANSAS, CALIFORNIA, COLORADO, HAWAII, IDAHO, KANSAS, LOUISIANA, MISSISSIPPI, MONTANA, NEBRASKA, NEVADA, NEW MEXICO, NORTH DAKOTA, OKLAHOMA, OREGON, PACIFIC TRUST TERRITORIES, SOUTH DAKOTA, TEXAS, UTAH, WASHINGTON, OR WYOMING,

SEND APPLICATIONS TO:

NUCLEAR MATERIALS LICENSING BRANCH
 U.S. NUCLEAR REGULATORY COMMISSION, REGION IV
 1600 E. LAMAR BOULEVARD
 ARLINGTON, TX 76011-4511

PERSONS LOCATED IN AGREEMENT STATES SEND APPLICATIONS TO THE U.S. NUCLEAR REGULATORY COMMISSION ONLY IF THEY WISH TO POSSESS AND USE LICENSED MATERIAL IN STATES SUBJECT TO U.S. NUCLEAR REGULATORY COMMISSION JURISDICTIONS.

1. THIS IS AN APPLICATION FOR (Check appropriate item)

- A. NEW LICENSE
- B. AMENDMENT TO LICENSE NUMBER 49-29310-01
- C. RENEWAL OF LICENSE NUMBER _____

2. NAME AND MAILING ADDRESS OF APPLICANT (Include ZIP code)

Front Range Mobile Imaging, Inc.
 418 East College Drive
 Cheyenne, WY 82007

3. ADDRESS WHERE LICENSED MATERIAL WILL BE USED OR POSSESSED

Within our mobile units while at temporary job sites as listed on our current license and stored at locations currently listed on our license.

4. NAME OF PERSON TO BE CONTACTED ABOUT THIS APPLICATION

John Mastin

BUSINESS TELEPHONE NUMBER
 (307) 638-6648

BUSINESS CELLULAR TELEPHONE NUMBER
 (307) 631-8973

BUSINESS EMAIL ADDRESS
john.mastin@frmiinc.com

SUBMIT ITEMS 5 THROUGH 11 ON 8-1/2 X 11" PAPER. THE TYPE AND SCOPE OF INFORMATION TO BE PROVIDED IS DESCRIBED IN THE LICENSE APPLICATION GUIDE.

5. RADIOACTIVE MATERIAL
 a. Element and mass number; b. chemical and/or physical form; and c. maximum amount which will be possessed at any one time.

6. PURPOSE(S) FOR WHICH LICENSED MATERIAL WILL BE USED.

7. INDIVIDUAL(S) RESPONSIBLE FOR RADIATION SAFETY PROGRAM AND THEIR TRAINING EXPERIENCE.

8. TRAINING FOR INDIVIDUALS WORKING IN OR FREQUENTING RESTRICTED AREAS.

9. FACILITIES AND EQUIPMENT.

10. RADIATION SAFETY PROGRAM.

11. WASTE MANAGEMENT.

12. LICENSE FEES (Fees required only for new applications, with few exceptions*)
 (*See 10 CFR 170 and Section 170.31)

FEE CATEGORY AMOUNT ENCLOSED \$

13. CERTIFICATION. (Must be completed by applicant) THE APPLICANT UNDERSTANDS THAT ALL STATEMENTS AND REPRESENTATIONS MADE IN THIS APPLICATION ARE BINDING UPON THE APPLICANT.

THE APPLICANT AND ANY OFFICIAL EXECUTING THIS CERTIFICATION ON BEHALF OF THE APPLICANT, NAMED IN ITEM 2, CERTIFY THAT THIS APPLICATION IS PREPARED IN CONFORMITY WITH TITLE 10, CODE OF FEDERAL REGULATIONS, PARTS 30, 32, 33, 34, 35, 36, 37, 39, AND 40, AND THAT ALL INFORMATION CONTAINED HEREIN IS TRUE AND CORRECT TO THE BEST OF THEIR KNOWLEDGE AND BELIEF.
 WARNING: 18 U.S.C. SECTION 1001 ACT OF JUNE 25, 1948 62 STAT. 749 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

CERTIFYING OFFICER -- TYPED/PRINTED NAME AND TITLE

John L Mastin, COO & RSO

SIGNATURE

DATE

1/21/16

FOR NRC USE ONLY

TYPE OF FEE	FEE LOG	FEE CATEGORY	AMOUNT RECEIVED	CHECK NUMBER	COMMENTS
APPROVED BY			\$	DATE	

MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 37, 39, 40, 70 and 71, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

<p>Licensee</p> <p>1. Front Range Nuclear Services</p> <p>2. 213 South Avenue C-1 Cheyenne, Wyoming 82007</p>	<p>In accordance with letters dated February 23, 2015 and March 25, 2015</p> <p>3. License number 49-27531-01 is amended in its entirety to read as follows:</p> <p>4. Expiration date September 30, 2021</p> <p>5. Docket No. 030-03968 Reference No.</p>
---	--

<p>6. Byproduct, source, and/or special nuclear material</p> <p>A. Any byproduct material permitted by 10 CFR 35.100</p> <p>B. Any byproduct material permitted by 10 CFR 35.200</p> <p>C. Any byproduct material permitted by 10 CFR 35.65</p>	<p>7. Chemical and/or physical form</p> <p>A. Any</p> <p>B. Any</p> <p>C. Sealed sources</p>	<p>8. Maximum amount that licensee may possess at any one time under this license</p> <p>A. As needed</p> <p>B. As needed</p> <p>12 millicuries total</p>
---	--	---

<p>9. Authorized use:</p> <p>A. Any uptake, dilution and excretion study permitted by 10 CFR 35.100.</p> <p>B. Any imaging and localization study permitted by 10 CFR 35.200.</p> <p>C. To be used for calibration of licensee's gamma cameras in the licensee's repair shop in accordance with letter dated February 23, 2015.</p>

CONDITIONS

- 10. A. Licensed material identified in Items 6.A. and 6.B. may be used or stored at Cheyenne Medical Specialist (CMS), 5050 Powderhouse Road, Cheyenne, Wyoming.
- B. Licensed material identified in Items 6.A. and 6.B. may be used and stored at 2000 Campbell Drive, Torrington, Wyoming, as described in letters dated May 1, 2007 (ML071340168), June 19, 2007 (ML071710318), and March 25, 2015 (ML15090A785).

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License Number
49-27531-01

Docket or Reference Number
030-33968

Amendment No. 31

- C. Licensed material identified in Items 6.A. and 6.B. may be stored at 213 South Avenue C-1, Cheyenne, Wyoming.
- D. Licensed material identified in Items 6.A. and 6.B. may be stored at 6658 Road 43, Torrington, Wyoming.
- E. Licensed material identified in Items 6.A. and 6.B. may be used and stored at Platte County Memorial, 201 14th St. Wheatland, Wyoming.
- F. Licensed material identified in Item 6.C. may be used or stored at 4912 Ridge Road, Cheyenne, Wyoming.
- G. Licensed material identified in Items 6.A. and 6.B. (except gases), may be used at temporary job sites anywhere in the United States where the U.S. Nuclear Regulatory Commission maintains jurisdiction for regulating the use of licensed material, including areas of exclusive Federal jurisdiction within Agreement States.

If the jurisdiction status of a Federal facility within an Agreement State is unknown, the licensee should contact the federal agency controlling the job site in question to determine whether the proposed job site is an area of exclusive Federal jurisdiction. Authorization for use of radioactive materials at job sites in Agreement States not under exclusive Federal jurisdiction shall be obtained from the appropriate state regulatory agency.

11. The Radiation Safety Officer for this license is Valerie Johnson, CNMT, RT(R)(N).
12. Licensed material is only authorized for use by or under the supervision of:
- A. Individuals permitted to work as an authorized user, authorized nuclear pharmacist, and/or authorized medical physicist in accordance with 10 CFR 35.13 and 35.14.
- B. The following individuals are authorized users for the material and medical uses indicated:

<u>Authorized Users</u>	<u>Material and Use</u>
S. A. Hayden, M.D.	35.100; 35.200
Mark E. Howshar, M.D.	35.100; 35.200
Eric B. Hoyer, M.D.	35.100; 35.200
J. G. Hubbard, M.D.	35.100; 35.200
Robert Jakob Kahn, M.D.	35.100; 35.200
D. M. Kellam, MD.	35.100; 35.200
William D. Ketcham II, M.D.	35.100; 35.200
Daniel R. Possehn, D.O.	35.100; 35.200
J. W. Wright, M.D.	35.100; 35.200

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License Number

49-27531-01

Docket or Reference Number

030-33968

Amendment No. 31

<u>Authorized Users</u>	<u>Material and Use</u>
Wendy Austin, M.D.	35.200
William Baker, M.D.	35.200
Anthony H. Doing, M.D.	35.200
Tristan Dow, M.D.	35.200
C. Patrick Green, M.D.	35.200
Gerald I. Myers, M.D.	35.200
John B. Oldemeyer, M.D.	35.200
Matthew Purvis, M.D.	35.200
Chad L. Stoltz, M.D.	35.200
Stephen A. Treat, M.D.	35.200

13. In addition to the possession limits in Item 8, the licensee shall further restrict the possession of licensed material to quantities below the minimum limit specified in 10 CFR 30.35(d) for establishing financial assurance for decommissioning.
14. A. Sealed sources shall be tested for leakage and/or contamination at intervals not to exceed the intervals specified in the certificate of registration issued by U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or by an Agreement State.
- B. In the absence of a certificate from a transferor indicating that a leak test has been made within the intervals specified in the certificate of registration issued by U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or under equivalent regulations of an Agreement State, prior to the transfer, a sealed source received from another person shall not be put into use until tested and the test results received.
- C. Sealed sources need not be leak tested if they contain only hydrogen-3; or they contain only a radioactive gas; or the half-life of the isotope is 30 days or less; or they contain not more than 100 microcuries of beta and/or gamma emitting material or not more than 10 microcuries of alpha emitting material.
- D. Sealed sources need not be tested if they are in storage and are not being used; however, when they are removed from storage for use or transferred to another person, and have not been tested within the required leak test interval, they shall be tested before use or transfer. No sealed source shall be stored for a period of more than 10 years without being tested for leakage and/or contamination.

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License Number
49-27531-01

Docket or Reference Number
030-33968

Amendment No. 31

- E. The leak test shall be capable of detecting the presence of 0.005 microcuries (185 becquerels) of radioactive material on the test sample. If the test reveals the presence of 0.005 microcuries (185 becquerels) or more of removable contamination, a report shall be filed with the U.S. Nuclear Regulatory Commission in accordance with 10 CFR 30.50(c)(2), and the source shall be removed immediately from service and decontaminated, repaired, or disposed of in accordance with Commission regulations. The report shall be filed within 5 days of the date the leak test result is known with the U.S. Nuclear Regulatory Commission, Region IV, 1600 East Lamar Boulevard, Arlington, Texas 76011-4511, ATTN: Director, Division of Nuclear Materials Safety. The report shall specify the source involved, the test results, and the corrective action taken.
- F. Tests for leakage and/or contamination, including leak test sample collection and analysis, shall be performed by the licensee or by persons specifically licensed by the U.S. Nuclear Regulatory Commission or an Agreement State to perform such services.
- G. Records of leak tests results shall be kept in units of microcuries and shall be maintained for 3 years.
15. The licensee shall conduct a physical inventory every 6 months, or at other intervals approved by the U.S. Nuclear Regulatory Commission, to account for all sources and/or devices received and possessed under the license. Records of inventories shall be maintained for 5 years from the date of each inventory, and shall include the radionuclides, quantities, manufacturer's name and model numbers, and the date of the inventory.
16. Sealed sources or detector cells containing licensed material shall not be opened or sources removed or detached from source rods or gauges by the licensee.
17. The licensee is authorized to transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."

589999

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License Number
49-27531-01

Docket or Reference Number
030-33968

Amendment No. 31

18. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. This license condition applies only to those procedures that are required to be submitted in accordance with the regulations. Additionally, this license condition does not limit the licensee's ability to make changes to the radiation protection program as provided for in 10 CFR 35.26. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.

- | | |
|---|---------------|
| A. Application dated May 28, 2011 | (ML11172A337) |
| B. Letter dated August 24, 2011 | (ML11238A083) |
| C. Letter dated December 8, 2011 | (ML11362A388) |
| D. Letter dated March 16, 2012 | (ML12088A392) |
| E. Letter dated February 10, 2014 with enclosures | (ML14057A829) |
| F. Letter dated February 23, 2015 with enclosures | (ML15090A786) |
| G. Letter dated March 25, 2015 | (ML15090A786) |



FOR THE U.S. NUCLEAR REGULATORY COMMISSION

Date: April 21, 2015

By:


Roberto J. Torres, M.S., Senior Health Physicist
Nuclear Materials Safety Branch B
Region IV
Arlington, Texas 76011-4511


MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 37, 39, 40, 70 and 71, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

<p style="text-align: center;">Licensee</p> <p>1. Memorial Hospital of Sheridan County</p> <p>2. 1401 West 5th Street Sheridan, Wyoming 82801</p>	<p>In accordance with letters dated January 6, 2015 and February 8, 2015</p> <p>3. License number 49-10982-02 is amended in its entirety to read as follows:</p> <p>4. Expiration date December 31, 2021</p> <p>5. Docket No. 030-13772</p> <p style="text-align: center;">Reference No.</p>
---	--

<p>6. Byproduct, source, and/or special nuclear material</p> <p>A. Any byproduct material permitted by 10 CFR 35.100</p> <p>B. Any byproduct material permitted by 10 CFR 35.200</p> <p>C. Any byproduct material permitted by 10 CFR 35.300</p> <p>D. Any byproduct material permitted by 10 CFR 35.400</p> <p>E. Any byproduct material permitted by 10 CFR 31.11</p>	<p>7. Chemical and/or physical form</p> <p>A. Any</p> <p>B. Any</p> <p>C. Any</p> <p>D. P-sealed sources (M Health Physics Service, and Minnesota Mining and Manufacturing Model Series 6500, formerly Model 6D60)</p> <p>E. Prepackaged kits</p>	<p>8. Maximum amount that licensee may possess at any one time under this license</p> <p>A. As needed</p> <p>B. As needed</p> <p>C. 1,100 millicuries</p> <p>D. 500 millicuries</p> <p>E. 50 millicuries</p>
---	---	--

9. Authorized use:
- A. Any uptake, dilution and excretion studies permitted by 10 CFR 35.100.
 - B. Any imaging and localization studies permitted by 10 CFR 35.200.
 - C. Any use permitted by 10 CFR 35.300.
 - D. For storage only.
 - E. In vitro studies.



**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License Number
49-10982-02

Docket or Reference Number
030-13772

Amendment No. 32

CONDITIONS

10. Licensed material may be used or stored only at the licensee's facilities located at 1401 West 5th Street, Sheridan, Wyoming.
11. The Radiation Safety Officer for this license is William Byron Taylor III, M.D.
12. Licensed material is only authorized for use by, or under the supervision of:
- A. Individuals permitted to work as an authorized user, authorized nuclear pharmacist, and/or authorized medical physicist in accordance with 10 CFR 35.13 and 35.14.
- B. The following individuals are authorized users for the material and medical uses indicated:

<u>Authorized Users</u>	<u>Material and Use</u>
Robert L. Stears, M.D.	35.100; 35.200; Oral administration of sodium iodide I-131; 31.11
Jason M. White, M.D.	35.100; 35.200; Oral administration of sodium iodide I-131; 31.11
Daniel R. Alzheimer, M.D.	35.100; 35.200; Oral administration of sodium iodide I-131; 31.11
William Byron Taylor III, M.D.	35.100; 35.200; Oral administration of sodium iodide I-131; 31.11
Michael W. Brennan, M.D.	35.200
Joseph Garcia, M.D.	35.200
John P. Stamato, M.D.	35.300; 35.400

13. In addition to the possession limits in Item 8, the licensee shall further restrict the possession of licensed material to quantities below the minimum limit specified in 10 CFR 30.35(d) for establishing financial assurance for decommissioning.
14. The licensee shall conduct a physical inventory every 6 months, or at other intervals approved by the U.S. Nuclear Regulatory Commission, to account for all sources and/or devices received and possessed under the license. Records of inventories shall be maintained for 3 years from the date of each inventory and shall include the radionuclides, quantities, manufacturer's name and model numbers, and the date of the inventory.
15. The licensee is authorized to transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License Number
49-10982-02

Docket or Reference Number
030-13772

Amendment No. 32

16. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. This license condition applies only to those procedures that are required to be submitted in accordance with the regulations. Additionally, this license condition does not limit the licensee's ability to make changes to the radiation protection program as provided for in 10 CFR 35.26. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.

- A. Application dated August 26, 2011 (ML11244A032)
 B. Letter dated December 8, 2011 (ML113480266)
 C. Facsimile dated July 22, 2014 (ML14213A442)
 D. Letter dated February 8, 2015 with enclosure (ML15042A371)



FOR THE U.S. NUCLEAR REGULATORY COMMISSION

By

Roberto J. Torres
 Roberto J. Torres, M.S., Senior Health Physicist
 Nuclear Materials Safety Branch B
 Region IV
 Arlington, Texas 76011-4511

Date February 11, 2015

FedEx

66689


fedex.com 1.800.GoFedEx 1.800.463.3339

FedEx
K# 8746 3356 1751

TUE - 26 JAN AA
EXPRESS SAVER

76011
TX-US
DFW

OR FWHA



FID 5019892 21JAN16 CYSA 539C1/0E61/481B

FedEx US Airbill
Express

FedEx Tracking Number 8746 3356 1751

020

1 From

Date: 1/21/16 Sender's FedEx Account Number: 2021-127000

Sender's Name: John H. ... Phone: ...

Company: State Range Mobile Imaging

Address: 412 E College Dr. City: ... State: TX ZIP: 75007

2 Your Internal Billing Reference

3 To

Recipient's Name: U.S. Marine Logistics Group Phone: 917-360-1100

Company: U.S. Marine Logistics Group

Address: 1600 E Lamar Boulevard City: Austin State: TX ZIP: 78701

City: Arlington State: TX ZIP: 76011-4511



8746 3356 1751

4a Express Packag

01 FedEx Priority Overnight
03 FedEx 2Day
20 FedEx Express Saver
4b Express Freight Service
70 FedEx 1Day Freight
80 FedEx 2Day Freight
5 Packaging
06 FedEx Envelope
02 FedEx Pak
6 Special
03 SATUR
No Signature
Does this shipper
7 Payment Bill to:
1 Sender
2 Recipient
3 Third Party
4 Credit Card
5 Cash/Check

RT 219
16:30
A
1751

fedex.com 1.800.GoFedEx 1.800.463.3339

Insert shipping document here.

606



DATE
01/25/2016

NAME AND ADDRESS OF APPLICANT AND/OR LICENSEE

Mr. John L. Mastin
Radiation Safety Officer
Front Range Mobile Imaging, Inc.
418 East College Drive
Cheyenne, WY 82007

LICENSE NUMBER

49-29310-01

MAIL CONTROL NUMBER

589999

LICENSING AND/OR TECHNICAL REVIEWER

CH

This is to acknowledge the receipt of your:

LETTER and/or APPLICATION DATED: 01/21/2016

The initial processing, which included an administrative review, has been performed.

AMENDMENT TERMINATION NEW LICENSE RENEWAL

- There were no administrative omissions identified during our initial review.
- This is to acknowledge receipt of your application for renewal of the material(s) license identified above. Your application is deemed timely filed, and accordingly, the license will not expire until final action has been taken by this office.
- Your application for a new NRC license did not include your taxpayer identification number. Please fill out NRC Form 531, located at the following link:

<http://www.nrc.gov/reading-rm/doc-collections/forms/nrc531.pdf>

Send the completed NRC Form 531, by facsimile, to the following number: (301) 415-5387

A copy of your action has been emailed to our License Fee and Accounts Receivable Branch, in our Headquarters office in Rockville, MD. You will be contacted separately if there is a fee issue involved.

Your application has been assigned the above listed **MAIL CONTROL NUMBER**. When calling to inquire about this action, please refer to this control number. Your application has been forwarded to a technical reviewer. Please note that the technical review, which is normally completed within 180 days for a renewal application (90 days for all other requests), may identify additional omissions or require additional information. If you have any questions concerning the processing of your application, our contact information is listed below:

Region IV
U. S. Nuclear Regulatory Commission
DNMS/NMSB - B
1600 E. Lamar Boulevard
Arlington, TX 76011-4511
(817) 200-1140

✓ 1/25/16

BETWEEN:

Accounts Receivable/Payable
and
Regional Licensing Branches

[FOR ARPB USE]
INFORMATION FROM WBL

Program Code: 02220
Status Code: Pending Amendment
Fee Category: 7C
Exp. Date: 01/31/2019
Fee Comments:
Decom Fin Assur Req: N

License Fee Worksheet - License Fee Transmittal

A. REGION

1. APPLICATION ATTACHED

Applicant/Licensee: Front Range Mobile Imaging, Inc.
Received Date: 01/25/2016
Docket Number: 3037854
Mail Control Number: 589999
License Number: 49-29310-01
Action Type: Amendment

2. FEE ATTACHED

Amount: _____

Check No.: _____

3. COMMENTS

Signed: _____

Carol A. Heie

Date: _____

1/25/16

B. LICENSE FEE MANAGEMENT BRANCH (Check when milestone 03 is entered / /)

1. Fee Category and Amount: _____

2. Correct Fee Paid. Application may be processed for:

Amendment: _____

Renewal: _____

License: _____

3. OTHER _____

Signed: _____

Date: _____
