

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

<p>1. LICENSEE/LOCATION INSPECTED:</p> <p>Nuclear Cardiac Imaging, LLC 3303 Dobson Drive Bay City, MI 48706</p> <p>REPORT NUMBER(S) 12-01</p>	<p>2. NRC/REGIONAL OFFICE</p> <p>Region III U. S. Nuclear Regulatory Commission 2443 Warrenton Road, Suite 210 Lisle, IL 60532-4352</p>
-------------------------------------------------------------------------------------------------------------------------------------------------------	-----------------------------------------------------------------------------------------------------------------------------------------------------

<p>3. DOCKET NUMBER(S)</p> <p>030-36938</p>	<p>4. LICENSE NUMBER(S)</p> <p>21-32577-01</p>	<p>5. DATE(S) OF INSPECTION</p> <p>11/6/12</p>
---------------------------------------------	------------------------------------------------	------------------------------------------------

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

1. Based on the inspection findings, no violations were identified.
2. Previous violation(s) closed.
3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):




✓ 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

10 CFR 71.5(a) requires that a licensee who transports licensed material outside of the site of usage, as specified in the NRC license, or where transport is on public highways, or who delivers licensed material to a carrier for transport, comply with the applicable requirements of the regulations appropriate to the mode of transport.

49 CFR 177.842 requires, in part, that packages of radioactive materials be so blocked and braced that they cannot change position during conditions normally incident to transportation.

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE	Laura M. Lura, RSO		12-7-12
NRC INSPECTOR	Michael G. Herr, CHP		12-4-12
BRANCH CHIEF	Tamara E. Bloomer		12/4/12

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Nuclear Cardiac Imaging, LLC
3303 Dobson Drive
Bay City, MI 48706

REPORT NUMBER(S) 12-01

2. NRC/REGIONAL OFFICE

Region III
U. S. Nuclear Regulatory Commission
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

3. DOCKET NUMBER(S)

030-36938

4. LICENSE NUMBER(S)

21-32577-01

5. DATE(S) OF INSPECTION

11/6/12

(Continued)

Contrary to the above, from October 2009 to November 6, 2012, on numerous occasions the licensee transported a package containing license materials, outside the site of usage, as specified on the NRC license, or on a public highway, and the package was not blocked and braced such that it could not change position during conditions normally incident to transportation. Specifically, packages containing various calibration source were stored in the cabinet loose and unused doses of Tc-99m were stored on the floor of the mobile nuclear medicine vehicle without strapping them down.

The licensee did secure the calibration sources in the cabinet and will properly strap packages containing unused doses of Tc-99m on the floor on the vehicle in the future. No spills or package damaged occurred during the October 2009 to November 6, 2012 time frame.