

GREENWAY ENGINEERING

151 WINDY HILL LANE

Winchester, VA. 226042

NUCLEAR GAUGE PHYSICAL INVENTORY

GREENWAY ENGINEERING

NAVITUS ENGINEERING

Gauge Number	Date Checked		Gauge Number	Date Checked
7 ✓	1-12-16		4 ✓	1-13-16
14 ✓	1-12-16		6 ✓	1-13-16
21 ✓	1-12-16		12 ✓	1-13-16
65 ✓	1-12-16		15 ✓	1-13-16
			32 ✓	1-13-16
			45 ✓	1-13-16
			46	bulk TEST 1/26/16
43*	At NETS for Repair			
Remarks: NETS - North East Technical Services				

→ sending in for
Repair
No bulk TEST
performed

Revised: 09/24/2014
Reviewed By: [Signature]
Date: JAN 13, 2016

40

#4



North East Technical Services, Inc.
75 Aileron Ct., Suite 4
Westminster, MD 21157
Ph: 410.751.5090 Fax: 410.751.5091

Greenway Engineering, Inc.
151 Windy Hill Lane
Winchester, VA 22602
ATTN: Lloyd Winters

Shipping Address: 151 Windy Hill Lane
Winchester, VA 22602

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	3411	Gauge S/N	13033	Leak Test Date	1/13/2016
Source	Reading in microCuries				
50-1410	0.00005927				
47-8343	0.00000				

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: Don de Silva

Date: JAN 19 2016



North East Technical Services, Inc.
 75 Aileron Ct., Suite 4
 Westminster, MD 21157
 Ph: 410.751.5090 Fax: 410.751.5091

Greenway Engineering, Inc.
 151 Windy Hill Lane
 Winchester, VA 22602
 ATTN: Lloyd Winters

*Shipping Address: 151 Windy Hill Lane
 Winchester, VA 22602*

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	3411	Gauge S/N	14223	Leak Test Date	1/26/2016
Source		Reading in microCuries			
	50-2834		0.00007321		
	47-9587		0.00000412		

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: *Douglas Sims*

Date: JAN 28 2016

#12



North East Technical Services, Inc.
75 Aileron Ct., Suite 4
Westminster, MD 21157
Ph: 410.751.5090 Fax: 410.751.5091

Greenway Engineering, Inc.
151 Windy Hill Lane
Winchester, VA 22602
ATTN: Lloyd Winters

Shipping Address: 151 Windy Hill Lane
Winchester, VA 22602

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	3411	Gauge S/N	9115	Leak Test Date	1/13/2016
Source		Reading in microCuries			
	40-6429		0.00005927		
	47-5584		0.00000		

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: Douglas Sims

Date: JAN 19 2016

#15



North East Technical Services, Inc.
75 Aileron Ct., Suite 4
Westminster, MD 21157
Ph: 410.751.5090 Fax: 410.751.5091

Greenway Engineering, Inc.
151 Windy Hill Lane
Winchester, VA 22602
ATTN: Lloyd Winters

Shipping Address: 151 Windy Hill Lane
Winchester, VA 22602

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	3411	Gauge S/N	8311	Leak Test Date	1/13/2016
Source		Reading in microCuries			
	40-5550		0.00005927		
	47-4673		0.00000		

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: Douglas Sims

Date: JAN 19 2016

#32



North East Technical Services, Inc.
75 Aileron Ct., Suite 4
Westminster, MD 21157
Ph: 410.751.5090 Fax: 410.751.5091

Greenway Engineering, Inc.
151 Windy Hill Lane
Winchester, VA 22602
ATTN: Lloyd Winters

Shipping Address: 151 Windy Hill Lane
Winchester, VA 22602

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	3411	Gauge S/N	5515	Leak Test Date	1/13/2016
Source		Reading in microCuries			
	40-2572		0.00008030		
	47-1621		0.00000		

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: Douglas Sims

Date: JAN 19 2016

Thursday, January 21, 2016

#45



North East Technical Services, Inc.
75 Aileron Ct., Suite 4
Westminster, MD 21157
Ph: 410.751.5090 Fax: 410.751.5091

Greenway Engineering, Inc.
151 Windy Hill Lane
Winchester, VA 22602
ATTN: Lloyd Winters

*Shipping Address: 151 Windy Hill Lane
Winchester, VA 22602*

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	3411	Gauge S/N	18576	Leak Test Date	1/13/2016
Source			Reading in microCuries		
50-8128			0.00005927		
47-14034			0.00000		

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: *Douglas Sims*

Date: JAN 19 2016

Thursday, January 21, 2016