

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

INFINITY PRIMARY CARE, PLLC
28455 Haggerty Rd
SUITE 102
Novi, MI 48377

2. NRC/REGIONAL OFFICE

Region III
U. S. Nuclear Regulatory Commission
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

REPORT NUMBER(S) 2016001

3. DOCKET NUMBER(S)

030-36689

4. LICENSE NUMBER(S)

21-32542-01

5. DATE(S) OF INSPECTION

3/2/16

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	E.L. Kulzer		3/2/16
BRANCH CHIEF	David T. McCann		3/9/16

Docket File Information

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Infinity Primary Care, PLLC 28455 Haggerty Rd Suite 102 Novi, MI 48377 REPORT NUMBER(S) 2016001	2. NRC/REGIONAL OFFICE Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352
--	---

3. DOCKET NUMBER(S) 030-36689	4. LICENSE NUMBER(S) 21-32542-01	5. DATE(S) OF INSPECTION 3/2/2016
----------------------------------	-------------------------------------	--------------------------------------

6. INSPECTION PROCEDURES USED IP 87130	7. INSPECTION FOCUS AREAS 03.01 through 03.07
---	--

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 02201	2. PRIORITY 5	3. LICENSEE CONTACT James Botti, M.S	4. TELEPHONE NUMBER (248) 533-0050
-----------------------------	------------------	---	---------------------------------------

Main Office Inspection Next Inspection Date: 3/2/2021

Field Office Inspection _____

Temporary Job Site Inspection _____

PROGRAM SCOPE

This was a routine unannounced inspection of a cardiac clinic authorized by the license to use any byproduct material as needed for any imaging and localization study permitted by 10 CFR 35.200 at the location specified on the license. The nuclear medicine department was staffed with one nuclear medicine technologist (NMT) who routinely conducted an average of 2-5 patient studies, normally both rest and stress tests, using Myoview, Monday, Wednesday, and Friday each week. The licensee received doses as ordered from a local pharmacy. All waste was held for decay-in-storage or returned to the nuclear pharmacy as a limited quantity shipment.

Performance Observations

During the inspection, the licensee's NMT demonstrated/discussed: (1) package check-in procedures and wipe test counting; (2) dosimetry; (3) dose calibrator tests; (4) unit dose handling procedures; (5) security of licensed material; (6) radiation safety program audits; (7) rad waste procedures; (8) sealed source inventory and leak tests; and (9) any spill response.

No violations of NRC regulatory requirements were identified during this inspection.