

**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

<b>1. LICENSEE/LOCATION INSPECTED:</b> Evanston Hospital Corporation dba Evanston Regional Hospital 190 Arrowhead Drive Evanston, WY REPORT NO: 030-33981/2015-001		<b>2. NRC/REGIONAL OFFICE</b>  U.S. Nuclear Regulatory Commission Region IV, 1600 East Lamar Blvd Arlington, Texas 76011-4511	
<b>3. DOCKET NUMBER</b>  030-33981	<b>4. LICENSE NUMBER</b>  49-27532-01	<b>5. DATE OF INSPECTION</b>  November 16, 2015	

**LICENSEE:**  
 The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy to exercise discretion, were satisfied.

Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):  
 Title 10 of the Code of Federal Regulations Part 20.1501 (a)(2) requires, in part, that each licensee make surveys and evaluate concentrations or quantities of residual radioactivity.  
 Contrary to the above, from June 29, 2015 to August 3, 2015, the licensee did not make surveys and evaluate quantities of residual radioactivity. Specifically, weekly wipe tests were not performed in the hot lab where radioactive materials were stored and used.

- The licensee has committed to performing weekly wipe tests in the future.
- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

**Licensee's Statement of Corrective Actions for Item 4, above.**

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE	Jon V Sigurdson	<i>Jon V Sigurdson</i>	11/16/2015
NRC INSPECTOR	Jason Dykert	<i>Jason Dykert</i>	11/16/2015
BRANCH CHIEF	Bruce G. Smith	<i>Bruce G. Smith</i>	11/30/15

Non-Public   
  Sensitive - Security-Related   
  Public   
  Non-Sensitive