

Seaman Nuclear Corporation

7315 South First Street Oak Creek, WI 53154 USA

Tel 414-762-5100 Fax 414-762-5106

info@seamannuclear.com

Leak Test Certificate

April 15, 2016

A leak test has been performed on meter, serial number **L640**, a model C-200, containing the radionuclide Ra 226

Owned By: Patriot Eng. & Environmental Inc.
6330 E. 75th St., Suite 216
Indianapolis, IN 46250

Date Sample Collected: 4/14/16
Collected By: B. King
Date Sample Analyzed: 4/15/16
Analyzed By: K. Krysiak

Most regulatory agencies consider a source to be leaking if a leak test reveals the presence of more than 0.005 microcurie of removable contamination.


Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by Wisconsin license 079-1257-01.

LEAK TEST DUE: 10/14/16



Scott C. Seaman
Radiation Safety Officer
Seaman Nuclear Corporation

S/N: L640
Leak Tested: 4/14/16
Due: 10/14/16
 Seaman Nuclear Corporation
Oak Creek, WI (414) 762-5100