

**U.S. NUCLEAR REGULATORY COMMISSION**

Date: May 27, 2016

**TELEPHONE CONVERSATION RECORD**

|                         |                     |                    |             |                   |           |
|-------------------------|---------------------|--------------------|-------------|-------------------|-----------|
| <b>Mail Control No.</b> | Not assigned<br>yet | <b>License No.</b> | 06-06697-02 | <b>Docket No.</b> | 030-01265 |
|-------------------------|---------------------|--------------------|-------------|-------------------|-----------|

**Name of Licensee:** Stamford Hospital

**Name of Participant(s):** Sarah Bull, RSO  
Shawn Seeley, Health Physicist, RI DNMS

**Telephone No.** 203-276-4036

**Subject:** Questions to Amendment Application

(NOTE: This will be used as the Documents Title in ADAMS)

Ms. Bull called and left a message on the division mailbox with questions regarding an amendment letter that she recently submitted for changes to the HDR program. I returned her call and found out they were changing the room where the HDR was to be used and had run into a construction delay.

She asked whether it was possible to decommission the HDR until such time as the renovations to the room were completed. I told her that was an acceptable solution. I reminded her that in any case, the HDR unit should be secured from unauthorized access while not in use during renovations to the area/room.

I told her to call back if she had any further questions.

**Document Availability:****Publicly Available****Non-Publicly Available****Non-Sensitive****Non-Sensitive Copyright****Sensitive- Proprietary** **Sensitive – Privacy Act/ PII** **Sensitive – Internal** **Sensitive – Security-Related** **Immediate Release****Normal Release Date:****Delayed Release Date:****SUNSI Review Completed By:** SSeeley / RA /**Document Accession #:** \_\_\_\_\_