

June 13, 2016

Mr. Dennis Lawyer
Division of Nuclear Materials Safety
US Nuclear Regulatory Commission, Region 1
2100 Renaissance Boulevard, Suite 100
King of Prussia, PA 19406-2713


RE License: 44-35148-01

Ladies/Gentlemen;

As part of our request for amendment to our license to remove our facility located at 186 S. Main Street, White River Jct., Vermont, we have included the most recent leak test results for the gauges that were stored there.

Please contact me should you have any questions or require additional information.

Thank you.



Alan I. Brown
Construction Services Manager, RSO

Cc: Paul Kohler, Corporate RSO
Chad Michaud

Attachments: Leak tests 1-22-16

Send all kits and requests to:

SUNTRAC Services, Inc.
1818 East Main Street
League City, TX 77573
(281) 338-2133

ATTN: SIT-KIT

CAUTION: Conduct a survey on the outside of each package placed in the U.S. Mails. Any reading over 0.5 mR/hr at contact with the envelope or package shall not be mailed.

LEAK TEST INVENTORY/REPORT FORM

Company Name: S. W. Cole Engineering, Inc. (11331)

Address: 10 Centre Road City: Somersworth State: NH Zip Code: 03878

Isotope: Cs 137 / Am 241 Be Activity: 10 / 40 mCi

Source Serial No: 4634CN/NJ5971 Leak Test Date: 01/22/16

Device Serial No: 608

Manufacturer: Instrotek Model No.: Xplorer 3500

Smear Taken By: Alan Brown

----- **DO NOT WRITE BELOW THIS LINE** -----

This is to certify that the above described smear/swab has been assayed at our facilities for indication of source leakage.

Our findings show the leakage to be:

ALPHA	BETA-GAMMA
<.0001	<.0001

µCi (Wet)

Certified By: 

Date: 02/09/16

SUNTRAC Services, Inc. Representative
(Texas Radioactive License No. L03062)

3358316

Send all kits and requests to:

SUNTRAC Services, Inc.
1818 East Main Street
League City, TX 77573
(281) 338-2133

ATTN: SIT-KIT

CAUTION: Conduct a survey on the outside of each package placed in the U.S. Mails. Any reading over 0.5 mR/hr at contact with the envelope or package shall not be mailed.

LEAK TEST INVENTORY/REPORT FORM

Company Name: S. W. Cole Engineering, Inc. (11331)

Address: 10 Centre Road City: Somersworth State: NH Zip Code: 03878

Isotope: Cs 137 / Am 241 Be Activity: 10 / 40 mCi

Source Serial No: CZ4418/167/13 Leak Test Date: 01/22/16

Device Serial No: 1730

Manufacturer: Instrotek Model No.: Xplorer 3500

Smear Taken By: Alan Brown

----- **DO NOT WRITE BELOW THIS LINE** -----

This is to certify that the above described smear/swab has been assayed at our facilities for indication of source leakage.

Our findings show the leakage to be:

ALPHA	BETA-GAMMA
<.0001	<.0001

µCi (Wet)

Certified By: 

Date: 02/09/16

SUNTRAC Services, Inc. Representative
(Texas Radioactive License No. L03062)

3358316

Send all kits and requests to:

SUNTRAC Services, Inc.
1818 East Main Street
League City, TX 77573
(281) 338-2133

ATTN: SIT-KIT

CAUTION: Conduct a survey on the outside of each package placed in the U.S. Mails. Any reading over 0.5 mR/hr at contact with the envelope or package shall not be mailed.

LEAK TEST INVENTORY/REPORT FORM

Company Name: S. W. Cole Engineering, Inc. (11331)

Address: 10 Centre Road City: Somersworth State: NH Zip Code: 03878

Isotope: Cs 137 / Am 241 Be Activity: 10 / 40.0 mCi

Source Serial No: 6487CM/NJ05906 Leak Test Date: 01/22/16

Device Serial No: 5500

Manufacturer: Humboldt Model No.: 5001EZ

Smear Taken By: Alan Brown

----- **DO NOT WRITE BELOW THIS LINE** -----

This is to certify that the above described smear/swab has been assayed at our facilities for indication of source leakage.

Our findings show the leakage to be:

ALPHA	BETA-GAMMA
<.0001	<.0001

µCi (Wet)

Certified By: 

Date: 02/09/16

SUNTRAC Services, Inc. Representative
(Texas Radioactive License No. L03062)

3358316

Send all kits and requests to:

SUNTRAC Services, Inc.
1818 East Main Street
League City, TX 77573
(281) 338-2133

ATTN: SIT-KIT

CAUTION: Conduct a survey on the outside of each package placed in the U.S. Mails. Any reading over 0.5 mR/hr at contact with the envelope or package shall not be mailed.

LEAK TEST INVENTORY/REPORT FORM

Company Name: S. W. Cole Engineering, Inc. (11331)

Address: 10 Centre Road City: Somersworth State: NH Zip Code: 03878

Isotope: Cs 137 / Am 241 Be Activity: 8 / 40 mCi

Source Serial No: 47-2173/705-9113 Leak Test Date: 01/22/16

Device Serial No: 33541

Manufacturer: Troxler Model No.: 3430

Smear Taken By: Alan Brown

----- **DO NOT WRITE BELOW THIS LINE** -----

This is to certify that the above described smear/swab has been assayed at our facilities for indication of source leakage.

Our findings show the leakage to be:

ALPHA	BETA-GAMMA
<.0001	<.0001

µCi (Wet)

Certified By: 

Date: 02/09/16

SUNTRAC Services, Inc. Representative
(Texas Radioactive License No. L03062)

3358316