

**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

<p>1. LICENSEE/LOCATION INSPECTED:</p> <p>Gibson General Hospital 1808 Sherman Drive Princeton, Indiana 47670</p> <p>REPORT NUMBER(S) 2016001</p>	<p>2. NRC/REGIONAL OFFICE</p> <p>Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352</p>
---	---

<p>3. DOCKET NUMBER(S)</p> <p>030-13826</p>	<p>4. LICENSE NUMBER(S)</p> <p>13-18524-01</p>	<p>5. DATE(S) OF INSPECTION</p> <p>June 14, 2016</p>
---	--	--

**LICENSEE:**

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

\_\_\_\_\_ Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.  
(Violations and Corrective Actions)

**Statement of Corrective Actions**

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Luis Nieves Folch		6/14/16
BRANCH CHIEF	Aaron T. McCraw		6/21/16

**Docket File Information**  
**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE/LOCATION INSPECTED:  Gibson General Hospital 1808 Sherman Drive Princeton, Indiana 47670  REPORT NUMBER(S) 2016001	2. NRC/REGIONAL OFFICE  Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352
--	---

3. DOCKET NUMBER(S)  030-13826	4. LICENSE NUMBER(S)  13-18524-01	5. DATE(S) OF INSPECTION  June 14, 2016
--------------------------------------	---	---

6. INSPECTION PROCEDURES USED  87130	7. INSPECTION FOCUS AREAS  03.01-03.07
--	--

**SUPPLEMENTAL INSPECTION INFORMATION**

1. PROGRAM CODE(S)  02121	2. PRIORITY  5	3. LICENSEE CONTACT  Thomas E. Schultheis - RSO	4. TELEPHONE NUMBER  (812) 385-3401
---------------------------------	----------------------	---	---

Main Office Inspection      Next Inspection Date: June 14, 2021

Field Office Inspection \_\_\_\_\_

Temporary Job Site Inspection \_\_\_\_\_

**PROGRAM SCOPE**

This was a routine, unannounced inspection of a 30-bed hospital located in Princeton, Indiana, with authorization to use byproduct materials in Sections 35.100 and 35.200. Licensed activities were conducted only at the location indicated on the license. The nuclear medicine department was staffed with two technologists who worked primarily at Good Samaritan Hospital in Vincennes, Indiana, but staffed this office as needed. The licensee's nuclear medicine staff typically administered 15 diagnostic doses monthly. The diagnostic procedures were predominately technetium-99m cardiac, hepatobiliary, and bone imaging, as well as iodine-123 for thyroid scans. The department received unit doses as needed from a licensed nuclear pharmacy or prepared doses from bulk technetium obtained from the nuclear pharmacy. All waste was either held for decay-in-storage (DIS) or returned to the nuclear pharmacy.

**Performance Observations**

No administrations of licensed material were performed during the inspection. Licensee personnel demonstrated dose calibrator constancy, package receipt and return surveys, administration of licensed material, and daily and weekly contamination surveys. The inspector noted no concerns with these activities. Interviews with licensee personnel indicated adequate knowledge of radiation safety concepts and procedures. The inspector performed independent and confirmatory radiation measurements which indicated results consistent with licensee survey records and postings.

No violations of NRC requirements were identified as a result of this inspection.