



6765 Langley Drive
Baton Rouge, Louisiana 70809
Telephone: 225-751-5893

Date: June 16, 2016

Buckeye Pipeline
2000 East SR 28
Muncie, IN 47303

Reference: 2016-123

This is to advise that the Radioactive Material as detailed below has been received by QSA Global, Inc. as of 6-15-16 and we have taken possession of this source:

Manufacturer	Model	Source S/N	Isotope	Activity
Kay-Ray	7063P	S98D2303	Cs-137	500 mci

Please retain this record for your files. Should you require further assistance, please contact us at QSA Global, Inc.

Regards,

Rusty Barrett
Technical Service Manager

GLTS
GL-727633

Send all kits and requests to:

SUNTRAC Services, Inc.
1818 East Main Street
League City, TX 77573
(281) 338-2133

ATTN: SIT-KIT

CAUTION: Conduct a survey on the outside of each package placed in the U.S. Mails. Any reading over 0.5 mR/hr at contact with the envelope or package shall not be mailed.

LEAK TEST INVENTORY/REPORT FORM

Company Name: Buckeye Pipeline - Muncie, IN BETMN

Address: 2000 E. SR 28 City: Muncie State: IN Zip Code: 47303

Isotope: Cs 137 Activity: 500 mCi

Line/Area: Muncie Unit: Muncie

Location: _____

Source Serial No: S98D2303 Leak Test Date: 6/13/2016

Device Serial No: N/A

Manufacturer: Kay Ray Model No.: 7063P

Smear Taken By: Josh Jones

----- **DO NOT WRITE BELOW THIS LINE** -----

This is to certify that the above described smear/swab has been assayed at our facilities for indication of source leakage.

Our findings show the leakage to be:

ALPHA	BETA-GAMMA
-----	<.0001

µCi (Wet)

Certified By: _____



Date: 6/16/2016

Kevin McMahon

Register Number: 3471516