



GENERAL MILLS

July 11, 2016

Director, Office of Nuclear Material Safety and Safeguards
ATTN: GLTS
U.S. Nuclear Regulatory Commission
Washington, DC 20555-001

Re: Transfer of Ownership (Disposal Transfer)
GL-705013-19
(NRC Materials License 13-1883-01)
General Mills
707 Pillsbury Lane
New Albany, IN 47150

Dear Sir/ Madam:

This letter serves as notification that, General Mills (license holder), has disposed of fifteen (15) nuclear gauges listed below to RAM Services, Inc (Two Rivers WI) on 6/27/16.

Device Serial#	mCi	Current Leak Test	Manufacturer	Model
18838-A	5	2/4/16	Kay-Ray	7062BP
18838-B	5	2/4/16	Kay-Ray	7062BP
18838-F	5	2/4/16	Kay-Ray	7062BP
18838-I	5	2/4/16	Kay-Ray	7062BP
18838-P	5	2/4/16	Kay-Ray	7062BP
20397	5	2/4/16	Kay-Ray	7062BP
26709A	5	2/4/16	Kay-Ray	7062BP
26709B	5	2/4/16	Kay-Ray	7062BP
26709C	5	2/4/16	Kay-Ray	7062BP
26709D	5	2/4/16	Kay-Ray	7062BP
26709E	5	2/4/16	Kay-Ray	7062BP
26710	10	2/4/16	Kay-Ray	7062BP
26711	5	2/4/16	Kay-Ray	7062BP
27570	10	6/28/16	Kay-Ray	7062BP
S94C3102	5	2/4/16	Kay-Ray	7062BP

The licensee contact person/ site RSO is Marie Puckett at the contact information below.

We are decommissioning this facility over the course of the next three (3) months. The plan is as follows:


- As equipment is taken out of service, any identified K-Ray associated with this equipment will be removed and placed into storage. Our specific device license allows us to do this. We will survey the device and equipment area before and after its removal from the equipment.
- The storage cabinet has always been surveyed on a regular basis and will continue to do so. Once the radiation levels for the stored devices reaches close but not over 5mR/hr at the closed cabinet door, we will call RAM Services, Inc to arrange for pick up.
- We conduct leak tests and surveys right before they are picked up by RAM to ensure there are no leaks.

July 11, 2016

- After all devices are removed I will alert your office and the Region III office of the termination of the license.
- I have attached all the relevant paperwork to this letter (RAM's receipt of materials letter; shipping documents and leak test certificates.)

Please contact me if you have any questions regarding this information.

Sincerely,

A handwritten signature in cursive script that reads "Marie Puckett".

Marie Puckett, CSP
Safety Manager
Radiation Safety Officer

FORM 540		RAM SERVICES, INC		5. SHIPPER - NAME AND FACILITY		SHIPPER I.D. NUMBER		7. NRC FORM 540 AND 540A PAGE 1 OF		8. MANIFEST NUMBER									
UNIFORM LOW-LEVEL RADIOACTIVE WASTE MANIFEST SHIPPING PAPER 1. EMERGENCY TELEPHONE NUMBER (Include Area Code) 800-424-9300				RAM Services, Inc. FOR General Mills 707 Pillsbury Lane New Albany IN 47150		47150		1 PAGE(S) 1 PAGE(S) 1 PAGE(S) 1 PAGE(S)		(Use this number on all continuation pages) 2016-0098									
				USER PERMIT NUMBER 2016-0098		SHIPMENT NUMBER 2016-0098		COLLECTOR PROCESSOR <input checked="" type="checkbox"/> GENERATOR TYPE IND		9. CONSIGNEE - Name and Facility Address RAM Services, Inc. 510 COUNTY HIGHWAY V TWO RIVERS WI 54241		CONTACT JERRY WIZA, RPT TELEPHONE NUMBER (Include Area Code) 1-920-686-3889							
ORGANIZATION CHEMTREC [for RAM SERVICES CONTRACT 210964]				CONTACT Marie Puckett or Jerry Wiza TELEPHONE NUMBER 812 941-4293		EPA I.D. NUMBER NA		SIGNATURE - Authorized consignee acknowledging waste receipt <i>Jerry Wiza</i>		DATE 7 JULY 2016									
2. IS THIS AN "EXCLUSIVE USE" SHIPMENT? <input type="checkbox"/> YES <input checked="" type="checkbox"/> NO		3. TOTAL NUMBER OF PACKAGES IDENTIFIED ON THIS MANIFEST 1		6. CARRIER - Name and Address RAM Services, Inc. 510 COUNTY HIGHWAY V TWO RIVERS WI 54241		SHIPPING DATE		10. CERTIFICATION This is to certify that the herein-named materials are properly classified, described, packaged, marked, and labeled and are in proper condition for transportation according to the applicable regulations of the Department of Transportation. This also certifies that the materials are classified, packaged, marked, and labeled and are in proper condition for transportation and disposal as described in accordance with the applicable requirements of 10 CFR Parts 20 and 61, or equivalent state regulations.											
4. DOES EPA REGULATED WASTE REQUIRING A MANIFEST ACCOMPANY THIS SHIPMENT? <input type="checkbox"/> YES <input checked="" type="checkbox"/> NO		EPA MANIFEST NUMBER NOT APPLICABLE		CONTACT JERRY WIZA, RPT		TELEPHONE NUMBER (Include Area Code) 920-686-3889		DATE 6/27/16		AUTHORIZED SIGNATURE <i>Jerry Wiza</i>									
11. U.S. DEPARTMENT OF TRANSPORTATION DESCRIPTION (Including UN ID number, proper shipping name, hazard class, and any additional information)				12. DOT LABEL "RADIOACTIVE"		13. TRANSPORT INDEX		14. PHYSICAL AND CHEMICAL FORM		15. INDIVIDUAL RADIONUCLIDES		16. TOTAL PACKAGE ACTIVITY IN SI UNITS		17. LSA/SCO CLASS		18. TOTAL WEIGHT OR VOLUME (Use appropriate units)		19. IDENTIFICATION NUMBER OF PACKAGE	
UN2915, Radioactive material, type A package, 7				Yellow II		0.1		SOLID OXIDE		Cs-137		mCi MBq 3.90E+01 1.44E+03		NONE		260.0 Kg 575.0 Lbs 7.5 FT3		RPK-1208	
FOR CONSIGNEE USE ONLY																			
_____ Record Waste Description Inadequate _____ Contamination or Leakage Detected _____ Unexpected Exposure Rates Detected _____ Labels, Markings, etc. Inadequate _____ Container Integrity Inadequate _____ Other _____ No Violations Detected on this Shipment																			

FORM 542		RAM SERVICES, INC.		1. WASTE COLLECTOR/PROCESSOR						2. MANIFEST NUMBER							
UNIFORM LOW-LEVEL RADIOACTIVE WASTE MANIFEST MANIFEST INDEX AND REGION COMPACT TABULATION List all original "PROCESSED WASTE" generators (if any) before "COLLECTED WASTE" generators.				Name		SHIPPER USE ONLY				2016-0098							
				RAM Services, Inc.													
				IDENTIFICATION NUMBER													
54241		7 Shipping Date				Page 1 of 1											
4. GENERATOR IDENTIFICATION NUMBER	5. GENERATOR NAME AND TELEPHONE NUMBER	6. GENERATOR FACILITY ADDRESS	6A. WASTE DESCRIPTION (NOMENCLATURE)	7. PREPROCESSED WASTE (OR MATERIAL) VOLUME		8. MANIFEST NUMBER(S) UNDER WHICH WASTE (OR MATERIAL) RECEIVED AND DATE OF RECEIPT	9. WASTE CODE P= PROCESSED C = COLLECTED	10. ORIGINATING COMPACT REGION OR STATE	11. AS PROCESSED/COLLECTED TOTAL								
				(m ³)	(lb)				A. SOURCE MATERIAL (kg) (lb)		B. SNM (g)	C. ACTIVITY (MBq) (mCi)		D. VOLUME (m ³) (ft ³)		E. WEIGHT (pounds)	F. MAXIMUM PACKAGE RADIATION LEVEL (mrem/hr)
47150	General Mills 812 941-4293	707 Pillsbury Lane New Albany, IN 47150	Sealed Source/Device	0.212376	7.50		C	IN	0	0	0	1.44E+03	3.90E+01	0.212	7.50	575.00	
TOTALS OF ALL PAGES(FORMS 542 AND 542A)									0.00E+00	0.00E+00	0.00E+00	1.44E+03	3.90E+01	2.12E-01	7.50E+00	5.75E+02	



510 County Highway V
 Two Rivers, WI 54241
 Voice +1-920-686-3889
 Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135848	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	<u>18838-A</u>	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	10 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-11.3684 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	-1E-05 µCi -0.21053 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST	Virginia Parlato	<i>Virginia Parlato</i>	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: CFL GLUTEN HI LEVEL, B2			



510 County Highway V
 Two Rivers, WI 54241
 Voice +1-920-686-3889
 Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135849	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON	MARIE PUCKETT	PHONE 812-941-4293	
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	<u>18838-B</u>	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	16 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-5.81053 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi -0.10526 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST	Virginia Parlato	<i>Virginia Parlato</i>	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: PHL GLUTEN HI LEVEL, A2			



510 County Highway V
 Two Rivers, WI 54241
 Voice +1-920-686-3889
 Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135853	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	18838-F	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	20 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-2.04211 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi -0.03509 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST	Virginia Parlato	<i>Virginia Parlato</i>	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: PIZZA GLUTEN LO LEVEL, A1			



510 County Highway V
 Two Rivers, WI 54241
 Voice +1-920-686-3889
 Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135856	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	<u>18838-1</u>	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	9 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-11.7684 dpm	MINIMUM DETECTABLE ACTIVITY	.118E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	-1E-05 µCi -0.22807 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST	Virginia Parlato	<i>Virginia Parlato</i>	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: PHL GLUTEN HI LEVEL, A1			



510 County Highway V
 Two Rivers, WI 54241
 Voice +1-920-686-3889
 Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135863	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	18838-P	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	13 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-8.62105 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi -0.15789 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST	Virginia Parlato	<i>Virginia Parlato</i>	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: CFL GLUTEN HI LEVEL, B1			



510 County Highway V
 Two Rivers, WI 54241
 Voice +1-920-686-3889
 Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135865	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON	MARIE PUCKETT	PHONE 812-941-4293	
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.	NA	DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	20397	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	11 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-10.0737 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	-1E-05 µCi -0.19298 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST	Virginia Parlato	<i>Virginia Parlato</i>	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: Pizza Gluten One Batch Level - A1			



510 County Highway V
 Two Rivers, WI 54241
 Voice +1-920-686-3889
 Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135872	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	26709A	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	16 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-5.81053 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi -0.10526 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST	Virginia Parlato	<i>Virginia Parlato</i>	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: Bread Line Sapp Bin Low Level			



510 County Highway V
 Two Rivers, WI 54241
 Voice +1-920-686-3889
 Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135873	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON	MARIE PUCKETT	PHONE 812-941-4293	
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	<u>26709B</u>	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	7 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-13.2632 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	-1E-05 µCi -0.26316 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST	Virginia Parlato	<i>Virginia Parlato</i>	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: Bread Line Soda Bin Low Level			



510 County Highway V.
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135874	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	26709C	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	14 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-8.08421 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi -0.14035 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST	Virginia Parlato	<i>Virginia Parlato</i>	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: Bread Line Salp Bin Low Level			



510 County Highway V
 Two Rivers, WI 54241
 Voice +1-920-686-3889
 Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135875	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON	MARIE PUCKETT	PHONE 812-941-4293	
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	<u>26709D</u>	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	10 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-10.8632 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	-1E-05 µCi -0.21053 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST	Virginia Parlato	<i>Virginia Parlato</i>	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: Bread Line B-1 Preblend One Batch Level			



510 County Highway V
 Two Rivers, WI 54241
 Voice +1-920-686-3889
 Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135876	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	26709E	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	6 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-13.9789 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	-1E-05 µCi -0.2807 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST	Virginia Parlato	<i>Virginia Parlato</i>	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: Bread Line B-2 Preblend One Batch Level			



510 County Highway V
 Two Rivers, WI 54241
 Voice +1-920-686-3889
 Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135877	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	<u>26710</u>	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	10 mCi 0.37 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	11 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-10.0737 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	-1E-05 µCi -0.19298 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST	Virginia Parlato	<i>Virginia Parlato</i>	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: Bulk Batching Unenriched Flour Bin High Level			



510 County Highway V
 Two Rivers, WI 54241
 Voice +1-920-686-3889
 Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135878	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	<u>26711</u>	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	31 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	23 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	0.957895 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi 0.01754 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST	Virginia Parlato	<i>Virginia Parlato</i>	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: Bulk Batching Unenriched Flour Bin High Level			



510 County Highway V
 Two Rivers, WI 54241
 Voice +1-920-686-3889
 Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		140204	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 UNITED STATES			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	<u>27570</u>	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	1 mCi 0.037 GBq
DATE SOURCE WIPED	26 Jun 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	28 Jun 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	11 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	21 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-8.73684 dpm	MINIMUM DETECTABLE ACTIVITY	1.15E-05 µCi 0.43 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi -0.17544 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST	Virginia Parlato	<i>Virginia Parlato</i>	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION:			



510 County Highway V
 Two Rivers, WI 54241
 Voice +1-920-686-3889
 Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135894	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	<u>S94C3102</u>	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	23 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	0.957895 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi 0.01754 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST	Virginia Parlato	<i>Virginia Parlato</i>	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: Bread Corn Sugar Low Level			