



July 19, 2016

Mr. Robert L. Gallagher
Health Physicist
U.S. NRC, Region I
2100 Renaissance Blvd.
King of Prussia, PA 19406

RE: Request for Additional Information - Control No. 590794

License No. 52-21325-01

Docket No. 03020510

Control No. 590794

Dear Mr. Gallagher:

1. We confirm that our Radiation Safety Officer will attend the training in our procedures. The training is to all individuals involved in Y-90 microsphere use, commensurate with the individual's duties to be performed and that this training would be provided to all individuals preparing, measuring, performing dosimetry calculations, administering Y-90 microspheres or part of the radiation safety personnel.
2. The Atomlab 500 dose has been purchase and we are waiting delivery of it. The calibrator will be used in our current Hot Lab in addition to measure the Y-90 SIR-Spheres.
3. We will be using the SIRTEX Medical Limited SIR Spheres microspheres.
4. We confirm that we will follow the manufacturer's procedures for determining shunting to non-treatment sites prior to treatments.

Cordially,

A handwritten signature in black ink, appearing to read 'Claudia V. Guzmán-Rivera'.

Claudia V. Guzmán-Rivera, JD, MHSA
Executive Director