



**UNITED STATES  
NUCLEAR REGULATORY COMMISSION**

REGION IV  
1600 E. LAMAR BLVD.  
ARLINGTON, TX 76011-4511

August 9, 2016

V. Troy Curnutt  
Radiation Safety Officer  
Diagnostic Imaging Service of Idaho  
1951 Bench Road, Suite F  
Pocatello, ID 83201

**SUBJECT: RESPONSE TO NRC INSPECTION REPORT 030-36948/2015-001 AND  
NOTICE OF VIOLATION**

Dear Mr. Curnutt:

On June 7, 2016, the U.S. Nuclear Regulatory Commission (NRC) cited Diagnostic Imaging Service of Idaho for four violations. The violations involved the failure to: (1) secure licensed materials from unauthorized removal or access; (2) calibrate instrumentation in accordance with nationally recognized standards or the manufacturer's instructions; (3) obtain a letter from each client that permits the use of byproduct material and clearly delineates the authority and responsibility of the licensee and the client; and (4) block and brace packages containing Class 7 (radioactive) materials during transport. The Notice of Violation, enclosed with NRC Inspection Report 030-36948/2015-001 (ML16158A440), required you to respond in writing within 30 days. You responded by letter received on June 30, 2016 (ML16202A087). Based on discussions with the inspector on July 11, 2016, you provided additional information by letter received on July 29, 2016 (ML16215A230).

After reviewing the information provided in your responses to the Notice of Violation, we have concluded that the completed and proposed corrective actions appear to adequately address the regulatory noncompliance described in the Notice of Violation. We will review the implementation of these corrective actions during a future inspection to determine whether full compliance has been restored and whether compliance will be maintained.

In accordance with 10 CFR 2.390 of the NRC's "Agency Rules of Practice and Procedure," a copy of this letter will be made available electronically for public inspection in the NRC Public

Document Room or from the NRC's Agencywide Documents Access Management System (ADAMS), accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Sincerely,

*/RA/*

Ray L. Kellar, P.E., Chief  
Nuclear Materials Safety Branch A

Docket No. 030-36948  
License No. 11-34221-01

cc: State of Idaho Radiation Control Program Director

SUBJECT: RESPONSE TO NRC INSPECTION REPORT 030-36948/2015-001 AND  
NOTICE OF VIOLATION

Dear Mr. Curnutt:

On June 7, 2016, the U.S. Nuclear Regulatory Commission (NRC) cited Diagnostic Imaging Service of Idaho for four violations. The violations involved the failure to: (1) secure licensed materials from unauthorized removal or access; (2) calibrate instrumentation in accordance with nationally recognized standards or the manufacturer's instructions; (3) obtain a letter from each client that permits the use of byproduct material and clearly delineates the authority and responsibility of the licensee and the client; and (4) block and brace packages containing Class 7 (radioactive) materials during transport. The Notice of Violation, enclosed with NRC Inspection Report 030-36948/2015-001 (ML16158A440), required you to respond in writing within 30 days. You responded by letter received on June 30, 2016 (ML16202A087). Based on discussions with the inspector on July 11, 2016, you provided additional information by letter received on July 29, 2016 (ML16215A230).

After reviewing the information provided in your responses to the Notice of Violation, we have concluded that the completed and proposed corrective actions appear to adequately address the regulatory noncompliance described in the Notice of Violation. We will review the implementation of these corrective actions during a future inspection to determine whether full compliance has been restored and whether compliance will be maintained.

In accordance with 10 CFR 2.390 of the NRC's "Agency Rules of Practice and Procedure," a copy of this letter will be made available electronically for public inspection in the NRC Public

Document Room or from the NRC's Agencywide Documents Access Management System (ADAMS), accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Sincerely,

*/RA/*

Ray L. Kellar, P.E., Chief  
Nuclear Materials Safety Branch A

Docket No. 030-36948  
License No. 11-34221-01

cc: State of Idaho Radiation Control Program Director

**DISTRIBUTION:**

S. Morris, DRA  
M. Shaffer, D:DNMS  
L. Howell, DD:DNMS

R4DNMS\_MS-A  
R4DNMS\_MS-B

<input checked="" type="checkbox"/> SUNSI Review By: <b>JFK</b>		ADAMS <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	<input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available	<input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Sensitive	Keyword:
OFFICE	DNMS/NMSB-A	C:NMSB-A			
NAME	JFKatanic	RLKellar			
SIGNATURE	<i>/RA/</i>	<i>/RA/</i>			
DATE	08/09/16	08/09/16			

OFFICIAL RECORD COPY