

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Wayne State University
Office of Environmental Health and Safety
5425 Woodward Avenue
Detroit, MI 48202

REPORT NUMBER(S) 2016-001

2. NRC/REGIONAL OFFICE

Region III
U. S. Nuclear Regulatory Commission
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

3. DOCKET NUMBER(S)

030-33995

4. LICENSE NUMBER(S)

21-00741-14

5. DATE(S) OF INSPECTION

7/28-30/16 with continued in-office review through 8/5/16

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

_____ Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Deborah A. Piskura, Senior Health Physicist	<i>Deborah A. Piskura</i>	8/11/16
BRANCH CHIEF	Aaron T. McCraw, Chief, MIB	<i>AJM for ATM</i>	8/12/16

Docket File Information

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Wayne State University Office of Environmental Health and Safety 5425 Woodward Avenue Detroit, MI 48202 REPORT NUMBER(S) 2016-001	2. NRC/REGIONAL OFFICE Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352
--	---

3. DOCKET NUMBER(S) 030-33995	4. LICENSE NUMBER(S) 21-00741-14	5. DATE(S) OF INSPECTION 7/28-30/16 with continued in-office review through 8/5/16
--------------------------------------	---	---

6. INSPECTION PROCEDURES USED 87126	7. INSPECTION FOCUS AREAS 03.01 - 03.07
--	--

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 03800	2. PRIORITY 3	3. LICENSEE CONTACT M. Srinivassan, M.S., RSO	4. TELEPHONE NUMBER (313) 577-0019
---------------------------------	----------------------	--	---

Main Office Inspection Next Inspection Date: 7/30/2019

Field Office Inspection _____

Temporary Job Site Inspection _____

PROGRAM SCOPE

This was a special inspection conducted to review the removal and decommissioning of three devices used by the University for research purposes. The devices were used occasionally by 5 principal investigators. In 2015, the University placed its material "in storage" and notified the NRC of its decision to permanently cease use. The licensee renewed its -14 license "for storage only" pending the removal and disposal of the material (Amendment No. 07, issued on 2/23/2016).

This inspection consisted of interviews with selected licensee personnel; review of selected records, including sealed source leak tests (last July 20, 2016); a tour of each room where licensed material was used and stored; and independent measurements. This inspection included in-office review through August 5, 2016, to confirm the receipt of the licensee's material by the disposal site and to discuss the license termination process.

No violations of NRC requirements were identified during this inspection.

*Gw
for ATM*