



**Docket File Information**

**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE/LOCATION INSPECTED:  Norlivo Internal Medicine, PLLC 20311 Farmington Rd Livonia, MI 48152-1411  REPORT NUMBER(S) 2016001	2. NRC/REGIONAL OFFICE  Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352
---	---

3. DOCKET NUMBER(S)  030-34560	4. LICENSE NUMBER(S)  21-32033-01	5. DATE(S) OF INSPECTION  7/14/2016
--------------------------------------	---	---

6. INSPECTION PROCEDURES USED  IP 87130	7. INSPECTION FOCUS AREAS  03.01 through 03.07
---	--

**SUPPLEMENTAL INSPECTION INFORMATION**

1. PROGRAM CODE(S)  02201	2. PRIORITY  5	3. LICENSEE CONTACT  Christie Adair	4. TELEPHONE NUMBER  (248) 442-1400
---------------------------------	----------------------	---	---

Main Office Inspection                      Next Inspection Date: 7/14/2021

Field Office Inspection \_\_\_\_\_

Temporary Job Site Inspection \_\_\_\_\_

**PROGRAM SCOPE**

This was a routine unannounced inspection of a cardiology group authorized to use byproduct material for diagnostic administrations of radiopharmaceuticals at its facilities in Livonia, Michigan. The licensee conducted studies on Mondays, Tuesdays, Thursdays, and Fridays. A nuclear medicine technologist performed all patient procedures. The licensee received licensed material as unit doses from a nuclear pharmacy.

**Performance Observations**

The licensee had not performed any cardiac stress studies with technetium-99m since June 2016. Interviews of available staff revealed an adequate level of understanding of emergency and material handling procedures and techniques. Dose calibrator constancy checks, package receipt, daily surveys, waste handling and disposal procedures were implemented successfully. Licensed material was secured as required. A survey meter was calibrated, operational, and it performed well in side-by-side comparison with an NRC instrument. Independent measurements indicated acceptable exposure rates in restricted and unrestricted areas. Personal whole body and extremity dosimetry badge records were reviewed during the inspection, and they did not indicate doses in excess of 10 CFR Part 20 limits.

No violations of NRC regulatory requirements were identified during this inspection.

GW  
for ATM