

From: [Sepulveda, Lymari X](#)
To: debra.bensen@elekta.com
Cc: [Herrera, Tomas](#)
Subject: Request for Additional Information
Date: Monday, August 29, 2016 1:01:00 PM

Good Morning Debbie,

This is in response to your request to amend registration certificate NR-0469-D-114-S to include updated labels for the microSelectron device. Please provide the dimension of the labels, and indicate if the materials of construction and method of attachment has changed. In addition, please confirm that the label that is attached to the treatment head when a new source is loaded has not changed.

Please submit the requested information within 30 calendar days of the date of this e-mail. The information should be submitted in an official company letter as an attachment to an e-mail or via facsimile at 301-415-5955. If we do not receive complete information within 30 days, we will consider your request as having been abandoned by you. This is without prejudice to the submission of a complete new request.

Thank you,

Lymari Sepulveda

Mechanical Engineer

Materials Safety Licensing Branch

Division of Material Safety, State, Tribal, and Regulatory Programs

Office of Nuclear Material Safety and Safeguards