



**UNITED STATES
NUCLEAR REGULATORY COMMISSION**
REGION IV
1600 E. LAMAR BLVD.
ARLINGTON, TX 76011-4511

September 21, 2016

Mr. Steve Dobos
Radiation Safety Officer
Energy Laboratories, Inc.
P.O. Box 247
Casper, WY 82602

**SUBJECT: RESPONSE TO NRC INSPECTION REPORT 030-29502/2016-001 AND
NOTICE OF VIOLATION**

Dear Mr. Dobos:

On August 9, 2016, the U.S. Nuclear Regulatory Commission (NRC) cited Energy Laboratories, Inc. for a violation involving the failure to follow licensed procedures for creating records which could demonstrate compliance with the discharge limits of 10 *CFR* 20.2003, "Disposal by Release into Sanitary Sewerage." Inspection Report 030-29502/2016-001 (ML16222A762), required you to respond in writing within 30 days. You responded by letter dated August 31, 2016 (ML16253A009), and supplement letter dated September 14, 2016 (ML16264A429).

After reviewing the information you provided in your response to the Notice of Violation, we have concluded that your completed and proposed actions appear to adequately address the regulatory noncompliance described in the Notice of Violation. We will review your implementation of these corrective actions during a future inspection to determine whether full compliance has been restored and whether compliance will be maintained.

In accordance with 10 *CFR* 2.390 of the NRC's "Agency Rules of Practice and Procedure," a copy of this letter will be made available electronically for public inspection in the NRC Public Document Room or from the NRC's Agencywide Documents Access Management System (ADAMS), accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Sincerely,

/RA/

Ray L. Kellar, P.E., Chief
Nuclear Materials Safety Branch A
Division of Nuclear Materials Safety

Docket No. 030-29502
License No. 49-26846-01

cc: Wyoming Radiation Control Program Director

SUBJECT: RESPONSE TO NRC INSPECTION REPORT 030-29502/2016-001 AND NOTICE OF VIOLATION

Dear Mr. Dobos:

On August 9, 2016, the U.S. Nuclear Regulatory Commission (NRC) cited Energy Laboratories, Inc. for a violation involving the failure to follow licensed procedures for creating records which could demonstrate compliance with the discharge limits of 10 CFR 20.2003, "Disposal by Release into Sanitary Sewerage." Inspection Report 030-29502/2016-001 (ML16222A762), required you to respond in writing within 30 days. You responded by letter dated August 31, 2016 (ML16253A009), and supplement letter dated September 14, 2016 (ML16264A429).

After reviewing the information you provided in your response to the Notice of Violation, we have concluded that your completed and proposed actions appear to adequately address the regulatory noncompliance described in the Notice of Violation. We will review your implementation of these corrective actions during a future inspection to determine whether full compliance has been restored and whether compliance will be maintained.

In accordance with 10 CFR 2.390 of the NRC's "Agency Rules of Practice and Procedure," a copy of this letter will be made available electronically for public inspection in the NRC Public Document Room or from the NRC's Agencywide Documents Access Management System (ADAMS), accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Sincerely,

/RA/

Ray L. Kellar, P.E., Chief
Nuclear Materials Safety Branch A
Division of Nuclear Materials Safety

Docket No. 030-29502
License No. 49-26846-01

cc: Wyoming Radiation Control Program Director

DISTRIBUTION:

S. Morris, DRA:RIV
R4DNMS_MS-A

M. Shaffer, D:DNMS
R4DNMS_MS-B

L. Howell, DD:DNMS

ADAMS ACCESSION NUMBER:ML16257A595

X SUNSI Review By: JCD		ADAMS X Yes <input type="checkbox"/> No	X Publicly Available <input type="checkbox"/> Non-Publicly Available	X Non-Sensitive <input type="checkbox"/> Sensitive	Keyword: n/a
OFFICE	NMSB-A	C:NMSB-A			
NAME	JCDykert	RLKellar			
SIGNATURE	<i>/RA/</i>	<i>/RA/</i>			
DATE	09/16/16	09/21/16			

OFFICIAL RECORD COPY