

September 19, 2016

Jennifer L. Bishop  
Health Physicist – Licensing  
NRC – Region III  
Re.: Amendment of NRC License No. 21-20137-01. Control No. 591831


Dear Ms. Bishop,

Below is additional clarification regarding the License Amendment for Port Huron Hospital  
Equipment for Emergency procedures related to Prostate permanent implants using sources listed  
under 10 CFR 35.400:

1. Survey meter: Fluke Biomedical ASM-992-PTO Advanced Survey Meter with internal GM detector.
2. Wipe tests: Fluke Biomedical 489-200, Pancake Scintillation Probe, 2x2x0.50 NaI(Tl)
3. Annual calibration: Fluke Electronics Global Calibration Laboratory
4. Ancillary equipment and supplies including long nose tweezers, plastic vial, lead pouch and disposable pads and gloves. See attached picture.



Mark Yudelev, Ph.D. DABMP  
Chief Physicist: McLaren-Macomb



## **Bishop, Jennifer**

---

**From:** Yudelev, Mark <Mark.Yudelev@mclaren.org>  
**Sent:** Monday, September 19, 2016 8:18 AM  
**To:** Bishop, Jennifer  
**Subject:** [External\_Sender] RE: Emergency equipment  
**Attachments:** NRC for MPH Emergency Equipment 09-19-2016.pdf

Attached.

Thank you,  
Mark.

**From:** Bishop, Jennifer [mailto:Jennifer.Dalzell-Bishop@nrc.gov]  
**Sent:** Monday, September 19, 2016 7:54 AM  
**To:** Yudelev, Mark <Mark.Yudelev@mclaren.org>  
**Subject:** RE: Emergency equipment

Dear Mr. Yudelev,

Could you please put your response in a signed cover letter?

Thank you,

Jennifer L. Bishop  
Health Physicist – Licensing  
NRC-Region III  
[Jennifer.Dalzell-Bishop@nrc.gov](mailto:Jennifer.Dalzell-Bishop@nrc.gov)  
630-829-9607

**From:** Yudelev, Mark [mailto:Mark.Yudelev@mclaren.org]  
**Sent:** Friday, September 16, 2016 5:22 PM  
**To:** Bishop, Jennifer <[Jennifer.Dalzell-Bishop@nrc.gov](mailto:Jennifer.Dalzell-Bishop@nrc.gov)>  
**Subject:** [External\_Sender] Emergency equipment

Jennifer,

Attached is a list of equipment to use in emergencies related to seed implants at Port Huron hospital.  
Please let me know if this is sufficient.

Thanks,  
Mark.

**Mark Yudelev, Ph.D. DABMP FAAPM**

Chief Physics:  
McLaren-Macomb  
1080 Harrington Blvd.

*Mt. Clemens, MI 48043*

*Phone: 586 493-7542*

*Fax: 586 493-7511*

*E-mail: [mark.yudelev@mclaren.org](mailto:mark.yudelev@mclaren.org)*

Confidentiality statement: "The information contained in this communication, including attachments, is confidential, may be privileged, and is intended only for the use of the named recipient(s). Unauthorized use, disclosure, forwarding or copying is strictly prohibited and may be unlawful. If you have received this communication in error, please notify me IMMEDIATELY at the phone number or pager listed above."

Confidentiality statement: "The information contained in this communication, including attachments, is confidential, may be privileged, and is intended only for the use of the named recipient(s). Unauthorized use, disclosure, forwarding or copying is strictly prohibited and may be unlawful. If you have received this communication in error, please notify me IMMEDIATELY at the phone number or pager listed above."

Confidentiality statement: "The information contained in this communication, including attachments, is confidential, may be privileged, and is intended only for the use of the named recipient(s). Unauthorized use, disclosure, forwarding or copying is strictly prohibited and may be unlawful. If you have received this communication in error, please notify me IMMEDIATELY at the phone number or pager listed above."