



IDAHO UROLOGIC INSTITUTE

2855 E. Magic View Dr.
Meridian, ID 83642
Main: (208) 639-4900
Fax: (208) 639-4901

Confidential

Fax

To: MICHELLE SIMMONS From: KEN NOTE

Fax: 817-200-1210 Pages: 1

Phone: _____ Date: 9-20-17

Re: _____ cc: _____

Urgent For Review Please Comment Please Reply Please Recycle

● Comments:

LICENSE # 11-35313-01

IC
 Immediate Release
 Normal Release

NON-PUBLIC

A.3 Sensitive-Security Related
 A.7 Sensitive Internal
 Other: _____

Reviewer: MS Date: 9/20/16

Confidentiality Notice: The information in this communication, including any attachments, is for the sole use of the intended recipient(s) and may contain confidential and privileged information. Any unauthorized review, use, disclosure or distribution is strictly prohibited and may be subject to legal restriction and sanction. If you have received this communication in error, please notify the sender immediately at 208-639-4900. Thank you for your cooperation.

September 16, 2016

Michelle Simmons
Health Physicist
US NRC
1600 East Lamar Blvd.
Arlington, TX 76011

RE: License No. 11-35313-01


Dear Ms. Simmons:

Thank you for your letter received by email on September 15, 2016, requesting additional information related to the Idaho Urologic Institute radioactive materials license application. I am providing the requested information below.

- We request the possession limit for Ra-223 to be 100 mCi.
- We commit to the following as part of our application: We will determine our radium-223 dose calibrator dial setting based on a unit dosage of radium-223 activity received from a commercial radiopharmacy. This dial setting will be recorded and used to make all subsequent dose calibrator radium-223 activity measurements.
- We commit to the following as part of our application: We have developed and will implement and maintain written waste disposal procedures for licensed material in accordance with 10 CFR 20.1101, that also meet the requirements of the applicable section of 10 CFR Part 20, Subpart K, and of 10 CFR 35.92.

Please advise if you require any additional information.

Sincerely,



Kenneth M. Nute
Radiation Safety Officer
Idaho Urologic Institute