

**From:** Davidson, Barry S CIV USARMY 20 CBRNE (US)  
**To:** [Lawyer, Dennis](#)  
**Subject:** [External\_Sender] Question on Amendment 8 of our License  
**Date:** Monday, October 17, 2016 3:21:23 PM

---

19-31127-01  
03037133

Mr. Lawyer,

I our renewal application 10 MAR 16, Section 10.7, 20th CBRNE Command stated "Leak tests are not required if the sources contain 100 microcuries or less of beta-gamma emitting material or 10 uCi or less of alpha-emitting material." I believe that this was consistent with NUREG 1556, Vol. 7 guidance.

Amendment 8 does not include this provision.

If this statement is not in the license what is the threshold for performing a leak test? If the criterion for a leaking source is 0.005 microcuries, is that the minimum quantity needing leak testing?

Sincerely,

Scott

B. Scott Davidson, CHP, CSP, Health Physicist  
20th CBRNE Command Risk Management Team  
2400 21st Street, Aberdeen Proving Ground, MD 21010  
Desk: 410-436-9643  
DSN: 584-9643  
Cell: 412-414-6673  
Alt. e-mail [bsdnuke@gmail.com](mailto:bsdnuke@gmail.com)

REC'D IN LAT 10/17/2016

590461  
NRC/RCNI MATERIALS-002