



**UNITED STATES  
NUCLEAR REGULATORY COMMISSION**

REGION III  
2443 WARRENVILLE RD. SUITE 210  
LISLE, IL 60532-4352

August 19, 2016

Laura T. Smith, M.S., DABR  
Radiation Safety Officer  
Bronson Methodist Hospital  
601 John Street  
Kalamazoo, MI 49007

Dear Ms. Smith:

In the letter dated August 2, 2016, you requested to: (1) add a Paw Paw, Michigan location of use to your NRC License No. 21-13125-01; (2) add 5 Authorized Users (AUs) currently listed on NRC License No. 21-26716-01; and (3) remove AU Robert Davis, M.D. from your license. Please clarify any errors in this understanding of your request.

Regarding the added location of use, in your letter, you indicated that the location being added is currently listed on the radioactive materials license for Bronson Lakeview Hospital, NRC License No. 21-26716-01. Please note that the location of use listed on the referenced license does not include a P.O. Box; we cannot include the P.O. Box in a location of use authorization. Please also note that we cannot add a location of use listed on another NRC license prior to receipt of all final materials disposition information for that other license. To date, such information has been requested from Bronson Lakeview Hospital, NRC License No. 21-26716-01, but has not yet been received.

**Concerning the location of use to be listed on your NRC License No. 21-13125-01, additional information is needed to complete our review of that location as follows:**


- 1. Please indicate the authorized uses (e.g. 10 CFR 35.100 and 35.200) to be listed on the license as authorized for use at the added Paw Paw, Michigan location of use, and whether any PET radionuclides will be used at that location.**
- 2. Please submit a diagram of the requested Paw Paw, Michigan location of use, drawn to the scale of 8 ½ by 11 inches. The diagram should show details for the nuclear medicine suite, including locations of any sinks, waste areas, and exhaust hoods. The hot lab should be clearly indicated on the diagram, and be free of superfluous architectural markings, unrelated to radioactive materials use concerns. It also should show what is above, below, and adjacent to the area(s) of use.**

L. Smith

- 2 -

Please provide a written response to this letter by September 2, 2016. Your response should be dated and signed by authorized personnel. You may submit your response via facsimile to my attention at (630) 515-1078. Include the reference control number 591664 with your response. We will resume our review once we receive your response. If you have any questions, please do not hesitate to contact me at 630-829-9892 or [sara.forster@nrc.gov](mailto:sara.forster@nrc.gov).

Sincerely,

A handwritten signature in cursive script that reads "Sara A. Forster".

Sara A. Forster, M.S.  
Health Physicist  
Materials Licensing Branch

License No. 21-13125-01  
Docket No. 030-02146

## Forster, Sara

---

**From:** Forster, Sara  
**Sent:** Friday, August 19, 2016 8:54 AM  
**To:** Smith, Laura T. Smith (Laura.Smith2@stjohn.org); 'lsphysics@att.net'  
**Subject:** Additional Information request for Bronson Methodist Hospital, NRC License No. 21-13125-01, CN591664  
**Attachments:** 02240.591664.21-13125-01 Bronson Methodist RFAI telecon.docx

Dear Ms. Smith:

Please see the attached file for additional information needed to complete the review of the recent amendment request concerning the above referenced applicant. Note that the attached letter requests additional information on or before close of business on September 2, 2016. Note that the additional information must be submitted via a dated letter signed a duly authorized management official. Additional guidance may be found in NUREG 1556, Vol. 9, Rev. 2, "Program Program-Specific Guidance About Medical Use Licenses;" which may be found at:

<http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v9/r2/>

Submission of your response as a pdf file attached to an email or via facsimile will allow for the quickest processing. Do not hesitate to call me with any questions you may have. Please also send a quick email when you receive this message, to confirm receipt.

Sincerely yours,

Sara A. Forster, Health Physicist Licensing Reviewer  
U.S. Nuclear Regulatory Commission - Region III  
Division of Nuclear Materials Safety  
2443 Warrenville Rd. - Ste. 210  
Lisle, IL 60532-4352  
[sara.forster@nrc.gov](mailto:sara.forster@nrc.gov)  
Direct: (630) 829-9892  
Facsimile: (630) 515-1078

