

U. S. Nuclear Regulatory Commission  
Region III  
Attn: Colleen Carol Casey  
Materials Licensing Section  
2443 Warrenville Road, Suite 210  
Lisle, Illinois 60532-4351

**License Control # 591836**

**Re: Response to EMAIL received; 11/14/2016**

Ms. Casey:

The following itemized response is directed toward the questions and concerns brought forward in your email received 11-14-16.

1. A) In the absence of our AMP Yinghui Zang, Ph.D., We are requesting that license item 6. (H) 10 CFR 35.600, Iridium - 192 for use in a High Dose Remote Afterloader unit, be removed from our license. The unit was taken out of service in March of 2016 and placed in lock down.

With the removal of 10 CFR 35.600, Iridium - 192, Mr. Beanblossom is qualified for all other uses named in our current license.

- B). (1) Mr. Beanblossom is the RSO at one other NRC license:

Fitzgibbon Hospital, Lic #; 24-32668-01

and is the RSO of three Agreement State Licenses:

Jennie Stuart Medical Center Lic #: 202-038-26

M.S. Community Health Lic #: 202-299-24


OH Muhlenburg Community Lic #: 202-132-24

- (2) Mr. Beanblossom, will be on campus during the second month of each calendar quarter. He is physically present for approximately three to four hours during each of those visits. He is also available 24/7 via phone, cell phone, Fax or E-mail. His office is approximately three hours from Good Samaritan Hospital.

- (3) Currently Mr. Beanblossom will be on our campus quarterly to review all records and to address radiation safety concerns pertaining to our licensed activities. Additionally, He is also available 24/7 via phone, cell phone, Fax or E-mail. His office is approximately three hours from Good Samaritan Hospital.
- (4) The typical response time would be, 24/7 via phone, cell phone, Fax or email. His office is approximately three hours from Good Samaritan Hospital should he be required to be on the campus.
- C.) As stated previously, in the absence of our AMP, Yinghui Zang, Ph.D., We are requesting that license Item 6. (H) 10 CFR 35.600, Iridium - 192 for use in a High Dose Remote Afterloader unit, be removed from our license. The unit was taken out of service in March of 2016 and placed in lock down. Dr. Zhang will remain on our license until the request for removal of 35:600 materials from our license is approved.

If you have further questions, please feel free to contact Mark Beanblossom at :  
636-798-2250 Direct  
636-399-1682 Cell  
636-798-2201 Fax

Sincerely,

  
\_\_\_\_\_  
Title President/CEO  
\_\_\_\_\_  
Date 11/28/16

## Casey, Colleen

---

**From:** Mark Schafer <mschafer@gshvin.org>  
**Sent:** Monday, November 28, 2016 4:06 PM  
**To:** Casey, Colleen  
**Cc:** Dan Kane (D.Kane@ampmedicalphysics.com); Mark Beanblossom (m.beanblossom@ampmedicalphysics.com); Mike Dixon  
**Subject:** [External\_Sender] License Control 591836 Good Samaritan Hospital Vincennes Indiana Response to Email.pdf  
**Attachments:** License Control 591836 Good Samaritan Hospital Vincennes Indiana Response to Email.pdf  
**Importance:** High


Good Samaritan Hospital  
Nuclear Medicine Department  
520 South Seventh Street  
Vincennes, Indiana 47591

**RE: RESPONSE TO LICENSE AMENDMENT  
LICENSE CONTROL #591836**

Ms. Casey,  
Per your request, attached is our response to your questions and concerns in your email received November 14, 2016  
Please contact Mr. Beanblossom by any of the means listed below should you have further questions or concerns:  
Direct: 636-798-2250  
Cell: 636-399-1682  
Fax: 636-798-2201

Respectfully Submitted,

**Mark Schafer, MHA, R.T., (R), (ARRT)**  
Administrative Director, Radiology Services



520 S. Seventh St. | Vincennes, Indiana | 47591  
Hospital: 812.882.5220 | Direct: 812.885.3628  
Cell: 765.277.1741  
[mschafer@gshvin.org](mailto:mschafer@gshvin.org)