



151 Windy Hill Lane
Winchester, Virginia 22602
Telephone: (888) 662-4185

Mr. Dennis Lawyer
Licensing Assistance Team
US Nuclear Regulatory Commission Region I
2100 Renaissance Boulevard, Suite 100
King of Prussia, PA 19406-2713

December 29, 2016

Mr. Lawyer,

Thank you for your reply. Each item in question is addressed below:


1. As the Management representative, I have reviewed the amendment application and concur with the statements and representations contained therein.
2. The increase from 8 Gauges to 12 Gauges would calculate to the following:
 $12 \times 8.3 \text{ mCi CS-137} = 99.6 \text{ mCi}$
 $12 \times 40 \text{ mCi Am-241} = 480 \text{ mCi}$
3. We wish to be licensed for the Troxler Electronic Laboratories 3400 Series Gauges.
4. Our amendment is only for the increase from 8 Gauges to 12 Gauges. And we wish to keep your licensing support documents the same as stated in the application dated October 14, 2015.

Per your email dated December 29, 2016, Mail Control No. 592482, we have address the first four items as requested.

Docket No. 03038868
NRC License No.: 45-35268-01

Should you need further information, please do not hesitate to contact us.

Thank you



Mark D. Smith
President

Lloyd C. Winters, CET/RSO
Cell: 540-539-0691
Office: 888-662-4185
lwinters@navituseng.com