



L-1

State of Vermont

Highway Division

Construction & Materials Bureau
Material Testing, Certification & Geotechnical Engineering
2178 Airport Road, Unit B
Berlin, Vermont 05641-8628
<http://vtrans.vermont.gov>

[phone] 802-828-2561
[fax] 802-828-2792
[ttd] 800-253-0191

Agency of Transportation

License No.: 44-14149-02
Docket No.: 03032893
Control No.: 592406

Scott Wilson, Health Physicist
NRC Region I

Subject: Request for Additional Information: Amendment # 11 to the State of Vermont, Agency of Transportation, (VTrans), Materials License # 44-14149-02.

Please reply with a message acknowledging your receipt of the additional information.

Mr. Wilson,

I have attached the leak tests for the nuclear gauges, SN: 33642 & 33645, that have been stored at the following storage sites for VTrans. District #3, 122 State Place, Rutland, Vermont, and 235 Joshua Place, Rutland, Vermont. These gauges at these two storage facilities have only sealed sources and the results of the leak tests were negative for contamination.

John E. Domey, RSO, 12/29/2016

Sincerely,
John Domey, RSO
Construction & Materials Bureau
Materials Testing & Certifications
Independent Assurance Unit
Office: 802-828-6914
Work Mobile: 802-498-8258
New Email address: john.domey@vermont.gov



592406
NM09/RGN1 MATERIALS-002

Applied Health Physics, LLC

2986 Industrial Blvd., Bethel Park, Pa 15102 Phone: (412) 835-9555 Fax: (412) 835-9559

E-Mail address: kmobley.ahp@comcast.net - Web Page address: WWW.APPLIEDHEALTHPHYSICS.COM

Mark V Leak Test Certificate

This certificate shall not be reproduced except in full, without the written approval of Applied Health Physics.

Device Information

Device Serial Number:	33642	Source Serial Number:	750-9310 / 47-3676		
Device Manufacturer:	Troxler	Source Manufacturer:	Troxler		
Device Model:	3430	Source Model:	3430		
Device Type:	Gauge	Make:	Moisture Density	Location:	V trans Berlin Mats. Lab
Radioisotope:	Cs-137/Am-241Be	Activity:	8 mCi / 40 mCi		

Leak Test Information

Leak Test Performed by:	John Domey	Leak Test Performed on:	10/4/2016
Leak Test Specimen Number:	A-042210	Leak Test Analysis Date:	10/11/2016
Leak Test Analysis Performed by:	L. McAllister		
Leak Test Results Indicate an Activity of:	< 0.005 uCi		

Notification

Analysis of this leak test sample was provided by Applied Health Physics, LLC. Pennsylvania Department of Environmental Protection License Number PA-0228A. In the event this specimen indicates leakage and or contamination in excess of 0.005 microcuries, Applied Health Physics, LLC will notify you promptly by telephone; at such time you will need to take the appropriate actions required by the governing regulatory agency. Applied Health Physics, LLC utilizes the following for leak test (wet-swab) analysis: Model: MARK V Leak Test Kit. Applied Health Physics, LLC utilizes the following instruments for all MARK V Leak Test analysis: Protean Instruments - ASC-950 ; Perkin Elmer - Tri-Carb 2800TR ; Ludlum Measurements - 2200 and Ludlum Measurements - 2929.

Client Information

State of Vermont Agency of Transportation	Contact:	John Domey
2178 Airport Rd Unit B	Phone:	(802) 828-6914
Materials & Research Section	Contact:	john.domey@state.vt.us
	Phone:	(802) 828-2792
Berlin	VT	05641

This certificate is an essential record and should be maintained for inspection by the regulatory agency.

Next LT Due 4/4/2017

Applied Health Physics, LLC

2986 Industrial Blvd., Bethel Park, Pa 15102 Phone: (412) 835-9555 Fax: (412) 835-9559
E-Mail address: kmobley.ahp@comcast.net - Web Page address: WWW.APPLIEDHEALTHPHYSICS.COM

Mark V Leak Test Certificate

This certificate shall not be reproduced except in full, without the written approval of Applied Health Physics.

Device Information

Device Serial Number:	33642	Source Serial Number:	750-9310 / 47-3676		
Device Manufacturer:	Troxler	Source Manufacturer:	Troxler		
Device Model:	3430	Source Model:	3430		
Device Type:	Gauge	Make:	Moisture Density	Location:	V trans Berlin Mats. Lab
Radioisotope:	Cs-137/Am-241Be	Activity:	8 mCi / 40 mCi		

Leak Test Information

Leak Test Performed by:	John Domey	Leak Test Performed on:	3/25/2016
Leak Test Specimen Number:	A-038964	Leak Test Analysis Date:	4/4/2016
Leak Test Analysis Performed by:	L. McAllister		
Leak Test Results Indicate an Activity of:	< 0.005 μ Ci		

Notification

Analysis of this leak test sample was provided by Applied Health Physics, LLC. Pennsylvania Department of Environmental Protection License Number PA-0228A. In the event this specimen indicates leakage and or contamination in excess of 0.005 microcuries, Applied Health Physics, LLC will notify you promptly by telephone; at such time you will need to take the appropriate actions required by the governing regulatory agency. Applied Health Physics, LLC utilizes the following for leak test (wet-swab) analysis: Model: MARK V Leak Test Kit. Applied Health Physics, LLC utilizes the following instruments for all MARK V Leak Test analysis: Protean Instruments - ASC-950 - Serial#: 01216140; Perkin Elmer - Tri-Carb 2800TR - Serial#: 073408; Ludlum Measurements - 2200 - Serial#: 25404.

Client Information

State of Vermont Agency of Transportation	Contact:	John Domey
2178 Airport Rd. Unit B	Phone:	(802) 828-6914
Materials & Research Section	Contact:	john.domey@state.vt.us
	Phone:	(802) 828-2792
Berlin	VT	05641

This certificate is an essential record and should be maintained for inspection by the regulatory agency.

Next LT Due 9/25/2016

Applied Health Physics, LLC

2986 Industrial Blvd., Bethel Park, Pa 15102 Phone: (412) 835-9555 Fax: (412) 835-9559
E-Mail address: kmobley.ahp@comcast.net - Web Page address: WWW.APPLIEDHEALTHPHYSICS.COM

Mark V Leak Test Certificate

This certificate shall not be reproduced except in full, without the written approval of Applied Health Physics.

Device Information

Device Serial Number:	33642	Source Serial Number:	750-9310 / 47-3676		
Device Manufacturer:	Troxler	Source Manufacturer:	Troxler		
Device Model:	3430	Source Model:	3430		
Device Type:	Gauge	Make:	Moisture Density	Location:	Vtrans Berlin Mats. Lab
Radioisotope:	Cs-137/Am-241Be	Activity:	8 mCi / 40 mCi		

Leak Test Information

Leak Test Performed by:	John Domey	Leak Test Performed on:	10/2/2015
Leak Test Specimen Number:	A-044129	Leak Test Analysis Date:	10/8/2015
Leak Test Analysis Performed by:	L. McAllister		
Leak Test Results Indicate an Activity of:	< 0.005 uCi		

Notification

Analysis of this leak test sample was provided by Applied Health Physics, LLC. Pennsylvania Department of Environmental Protection License Number PA-0228A. In the event this specimen indicates leakage and or contamination in excess of 0.005 microcuries, Applied Health Physics, LLC will notify you promptly by telephone; at such time you will need to take the appropriate actions required by the governing regulatory agency. Applied Health Physics, LLC utilizes the following for leak test (wet-swab) analysis: Model: MARK V Leak Test Kit. Applied Health Physics, LLC utilizes the following instruments for all MARK V Leak Test analysis: Protean Instruments - ASC-950 - Serial#: 01216140; Perkin Elmer - Tri-Carb 2800TR - Serial#: 073408; Ludlum Measurements - 2200 - Serial#: 25404.

Client Information

State of Vermont Agency of Transportation	Contact:	John Domey
2178 Airport Rd. Unit B	Phone:	(802) 828-6914
Materials & Research Section	Contact:	john.domey@state.vt.us
	Phone:	(802) 828-2792
Berlin	VT	05641

This certificate is an essential record and should be maintained for inspection by the regulatory agency.

Next LT Due 4/2/2016

Applied Health Physics, LLC

2986 Industrial Blvd., Bethel Park, Pa 15102 Phone: (412) 835-9555 Fax: (412) 835-9559
E-Mail address: knobley.ahp@comcast.net - Web Page address: WWW.APPLIEDHEALTHPHYSICS.COM

Mark V Leak Test Certificate

This certificate shall not be reproduced except in full, without the written approval of Applied Health Physics.

Device Information

Device Serial Number:	33642	Source Serial Number:	750-9310 / 47-3676		
Device Manufacturer:	Troxler	Source Manufacturer:	Troxler		
Device Model:	3430	Source Model:	3430		
Device Type:	Gauge	Make:	Moisture Density	Location:	Stall 3 Central Garage
Radioisotope:	Cs-137/Am-241Be	Activity:	8 mCi / 40 mCi		

Leak Test Information

Leak Test Performed by:	John Domey	Leak Test Performed on:	2/4/2015
Leak Test Specimen Number:	A-038767	Leak Test Analysis Date:	2/11/2015
Leak Test Analysis Performed by:	L. McAllister		
Leak Test Results Indicate an Activity of:	<0.005 uCi		

Notification

The analysis of this leak test sample was provided by Applied Health Physics, Pennsylvania Department of Environmental Protection License Number PA-0228A. In the event this specimen indicates leakage and or contamination in excess of 0.005 microcuries, Applied Health Physics will notify you promptly by telephone; at such time you will need to take the appropriate actions required by the governing regulatory agency. Applied Health Physics utilized the following for leak test analysis: Sample model: MARK V Leak Test Kit for wet swab analysis. Leak test analysis instrument: Manufacturer - Protean Instruments Model ASC950 serial number 01216140.

Client Information

State of Vermont Agency of Transportation	Contact:	John Domey
2178 Airport Rd Unit B	Phone:	(802) 828-6914
Materials & Research Section	Contact:	john.domey@state.vt.us
	Phone:	(802) 828-2792
Berlin	VT	05641

This certificate is an essential record and should be maintained for inspection by the regulatory agency.

Next LT Due 8/4/2015

Applied Health Physics

2986 Industrial Blvd., Bethel Park, Pa 15102 Phone: (412) 835-9555 Fax: (412) 835-9559
E-Mail address: ahp.inc@comcast.net - Web Page address: WWW.APPLIEDHEALTHPHYSICS.COM

RECEIVED
AUG 28 2014
PA DEPARTMENT OF
ENVIRONMENTAL PROTECTION

Mark V Leak Test Certificate

This certificate shall not be reproduced except in full, without the written approval of Applied Health Physics.

Device Information

Device Serial Number:	33642	Source Serial Number:	750-9310 / 47-3676		
Device Manufacturer:	Troxler	Source Manufacturer:	Troxler		
Device Model:	3430	Source Model:	3430		
Device Type:	Gauge	Make:	Moisture Density	Location:	SW Regional Lab
Radioisotope:	Cs-137/Am-241Be	Activity:	8 mCi / 40 mCi		

Leak Test Information

Leak Test Performed by:	John Dorney	Leak Test Performed on:	8/5/2014
Leak Test Specimen Number:	A-038714	Leak Test Analysis Date:	8/13/2014
Leak Test Analysis Performed by:	L. McAllister		
Leak Test Results Indicate an Activity of:	< 0.005 uCi		

Notification

The analysis of this leak test sample was provided by Applied Health Physics. Pennsylvania Department of Environmental Protection License Number PA-0228A. In the event this specimen indicates leakage and or contamination in excess of 0.005 microcuries, Applied Health Physics will notify you promptly by telephone; at such time you will need to take the appropriate actions required by the governing regulatory agency. Applied Health Physics utilized the following for leak test analysis: Sample model: MARK V Leak Test Kit for wet swab analysis. Leak test analysis instrument: Manufacturer - Protean Instruments Model ASC950 serial number 01216140.

Client Information

State of Vermont Agency of Transportation	Contact:	John Dorney
2178 Airport Rd Unit B	Phone:	(802) 828-6914
Materials & Research Section	Contact:	john.dorney@state.vt.us
	Phone:	(802) 828-2792
Berlin	VT	05641

This certificate is an essential record and should be maintained for inspection by the regulatory agency.

Next LT Due 2/5/2015

Applied Health Physics

2986 Industrial Blvd., Bethel Park, Pa 15102 Phone: (412) 835-9555 Fax: (412) 835-9559
E-Mail address: ahp.inc@comcast.net - Web Page address: WWW.APPLIEDHEALTHPHYSICS.COM

Mark V Leak Test Certificate

This certificate shall not be reproduced except in full, without the written approval of Applied Health Physics.

Device Information

Device Serial Number:	33642	Source Serial Number:	750-9310 / 47-3676		
Device Manufacturer:	Troxler	Source Manufacturer:	Troxler		
Device Model:	3430	Source Model:	3430		
Device Type:	Gauge	Make:	Moisture Density	Location:	Central Garage Stall 3
Radioisotope:	Cs-137/Am-241Be	Activity:	8 mCi / 40 mCi		

Leak Test Information

Leak Test Performed by:	John Domey	Leak Test Performed on:	2/4/2014
Leak Test Specimen Number:	A-037691	Leak Test Analysis Date:	2/10/2014
Leak Test Analysis Performed by:	L. McAllister		
Leak Test Results Indicate an Activity of:	< 0.005 uCi		

Notification

The analysis of this leak test sample was provided by Applied Health Physics. Pennsylvania Department of Environmental Protection License Number PA-0228A. In the event this specimen indicates leakage and or contamination in excess of 0.005 microcuries, Applied Health Physics will notify you promptly by telephone; at such time you will need to take the appropriate actions required by the governing regulatory agency. Applied Health Physics utilized the following for leak test analysis: Sample model: MARK V Leak Test Kit for wet swab analysis. Leak test analysis instrument: Manufacturer - Protean Instruments Model ASC950 serial number 01216140.

Client Information

State of Vermont Agency of Transportation	Contact:	John Domey
2178 Airport Rd Unit B	Phone:	(802) 828-6914
Materials & Research Section	Contact:	john.domey@state.vt.us
Berlin	Phone:	(802) 828-2792
VT		05641

This certificate is an essential record and should be maintained for inspection by the regulatory agency.

Next LT Due 8/4/2014

Applied Health Physics, Inc.

2986 Industrial Blvd., Bethel Park, Pa 15102 Phone: (412) 835-9555 Fax: (412) 835-9559
E-Mail address: ahp.inc@comcast.net - Web Page address: WWW.APPLIEDHEALTHPHYSICS.COM

Mark V Leak Test Certificate

This certificate shall not be reproduced except in full, without the written approval of Applied Health Physics, Inc.

Device Information

Device Serial Number:	33642	Source Serial Number:	750-9310 / 47-3676		
Device Manufacturer:	Troxler	Source Manufacturer:	Troxler		
Device Model:	3430	Source Model:	3430		
Device Type:	Gauge	Make:	Moisture Density	Location:	SW Regional Lab
Radioisotope:	Cs-137/Am-241Be	Activity:	8 mCi / 40 mCi		

Leak Test Information

Leak Test Performed by:	Christopher Rea	Leak Test Performed on:	8/27/2013
Leak Test Specimen Number:	A-035250	Leak Test Analysis Date:	9/4/2013
Leak Test Analysis Performed by:	L. McAllister		
Leak Test Results Indicate an Activity of:	< 0.005 uCi		

Notification

The analysis of this leak test sample was provided by Applied Health Physics, Inc. Pennsylvania Department of Environmental Protection License Number PA-0228A. In the event this specimen indicates leakage and or contamination in excess of 0.005 microcuries, Applied Health Physics, Inc. will notify you promptly by telephone; at such time you will need to take the appropriate actions required by the governing regulatory agency. Applied Health Physics, Inc. utilized the following for leak test analysis: Sample model: MARK V Leak Test Kit for wet swab analysis. Leak test analysis instrument: Manufacturer - Protean Instruments Model ASC950 serial number 01216140.

Client Information

State of Vermont Agency of Transportation	Contact:	Christopher Rea
	Phone:	(802) 828-6923
1716 US Route 302	Contact:	FAX
	Phone:	(802) 828-2792
Berlin	VT	05602

This certificate is an essential record and should be maintained for inspection by the regulatory agency.

RECEIVED

SEP 12 2013

MATERIALS and
RESEARCH SECTION

Next LT Due 2/27/2014

Applied Health Physics, Inc.

2986 Industrial Blvd., Bethel Park, Pa 15102 Phone: (412) 835-9555 Fax: (412) 835-9559
E-Mail address: ahp.inc@comcast.net - Web Page address: WWW.APPLIEDHEALTHPHYSICS.COM

Mark V Leak Test Certificate

This certificate shall not be reproduced except in full, without the written approval of Applied Health Physics, Inc.

Device Information

Device Serial Number:	33642	Source Serial Number:	750-9310 / 47-3676		
Device Manufacturer:	Troxler	Source Manufacturer:	Troxler		
Device Model:	3430	Source Model:	3430		
Device Type:	Gauge	Make:	Moisture Density	Location:	
Radioisotope:	Cs-137/Am-241Be	Activity:	8 mCi / 40 mCi		

Leak Test Information

Leak Test Performed by:	Christopher Rea	Leak Test Performed on:	2/7/2013
Leak Test Specimen Number:	A-032658	Leak Test Analysis Date:	2/18/2013
Leak Test Analysis Performed by:	L. McAllister		
Leak Test Results Indicate an Activity of:	< 0.005 uCi		

Notification

The analysis of this leak test sample was provided by Applied Health Physics, Inc. Pennsylvania Department of Environmental Protection License Number PA-0228A. In the event this specimen indicates leakage and or contamination in excess of 0.005 microcuries, Applied Health Physics, Inc. will notify you promptly by telephone; at such time you will need to take the appropriate actions required by the governing regulatory agency. Applied Health Physics, Inc. utilized the following for leak test analysis: Sample model: MARK V Leak Test Kit for wet swab analysis. Leak test analysis instrument: Manufacturer - Protean Instruments Model ASC950 serial number 01216140.

Client Information

State of Vermont Agency of Transportation	Contact:	Christopher Rea
Materials & Research Section	Phone:	(802) 828-6923
1 National Life Drive	Contact:	FAX
	Phone:	(802) 828-2792
Montpelier	VT	05633

This certificate is an essential record and should be maintained for inspection by the regulatory agency.

Next LT Due 8/7/2013

Applied Health Physics, Inc.

2986 Industrial Blvd., Bethel Park, Pa 15102 Phone: (412) 835-9555 Fax: (412) 835-9559
E-Mail address: ahp.inc@comcast.net - Web Page address: WWW.APPLIEDHEALTHPHYSICS.COM

RECEIVED

JUN 27 2012

MATERIALS and
RESEARCH SECTION

Mark V Leak Test Certificate

This certificate shall not be reproduced except in full, without the written approval of Applied Health Physics, Inc.

Device Information

Device Serial Number:	33642	Source Serial Number:	750-9310 / 47-3676		
Device Manufacturer:	Troxler	Source Manufacturer:	Troxler		
Device Model:	3430	Source Model:	3430		
Device Type:	Gauge	Make:	Moisture Density	Location:	Central Lab
Radioisotope:	Cs-137/Am-241Be	Activity:	8 mCi / 40 mCi		

Leak Test Information

Leak Test Performed by:	Chris Rea	Leak Test Performed on:	5/7/2012
Leak Test Specimen Number:	A-033822	Leak Test Analysis Date:	6/21/2012
Leak Test Analysis Performed by:	M. Lazzari		
Leak Test Results Indicate an Activity of:	< 0.005 uCi		

Notification

The analysis of this leak test sample was provided by Applied Health Physics, Inc. Pennsylvania Department of Environmental Protection License Number PA-0228A. In the event this specimen indicates leakage and or contamination in excess of 0.005 microcuries, Applied Health Physics, Inc. will notify you promptly by telephone; at such time you will need to take the appropriate actions required by the governing regulatory agency.

Client Information

State of Vermont Agency of Transportation	Contact:	Christopher Rea
Materials & Research Section	Phone:	(802) 828-6923
1 National Life Drive	Contact:	FAX
	Phone:	(802) 828-2792
Montpelier	VT	05633

This certificate is an essential record and should be maintained for inspection by the regulatory agency.

Next LT Due 11/7/2012

Applied Health Physics, LLC

2986 Industrial Blvd., Bethel Park, Pa 15102 Phone: (412)-835-9555 Fax: (412) 835-9559
E-Mail address: kmoble.ahp@comcast.net - Web Page address: WWW.APPLIEDHEALTHPHYSICS.COM

Mark V Leak Test Certificate

This certificate shall not be reproduced except in full, without the written approval of Applied Health Physics.

Device Information

Device Serial Number:	33645	Source Serial Number:	750-9313 / 47-4761		
Device Manufacturer:	Troxler	Source Manufacturer:	Troxler		
Device Model:	3430	Source Model:	3430		
Device Type:	Gauge	Make:	Moisture Density	Location:	SW Construction Mats. Lab
Radioisotope:	Cs-137 / Am-241Be	Activity:	8 mCi / 40 mCi		

Leak Test Information

Leak Test Performed by:	John Domey	Leak Test Performed on:	9/30/2016
Leak Test Specimen Number:	A-042262	Leak Test Analysis Date:	10/11/2016
Leak Test Analysis Performed by:	L. McAllister		
Leak Test Results Indicate an Activity of:	< 0.005 uCi		

Notification

Analysis of this leak test sample was provided by Applied Health Physics, LLC. Pennsylvania Department of Environmental Protection License Number PA-0228A. In the event this specimen indicates leakage and or contamination in excess of 0.005 microcuries, Applied Health Physics, LLC will notify you promptly by telephone; at such time you will need to take the appropriate actions required by the governing regulatory agency. Applied Health Physics, LLC utilizes the following for leak test (wet-swab) analysis: Model: MARK V Leak Test Kit. Applied Health Physics, LLC utilizes the following instruments for all MARK V Leak Test analysis: Protean Instruments - ASC-950 ; Perkin Elmer - Tri-Carb 2800TR ; Ludlum Measurements - 2200 and Ludlum Measurements - 2929.

Client Information

State of Vermont Agency of Transportation	Contact:	John Domey
2178 Airport Rd Unit B	Phone:	(802) 828-6914
Materials & Research Section	Contact:	john.domey@state.vt.us
	Phone:	(802) 828-2792
Berlin	VT	05641

This certificate is an essential record and should be maintained for inspection by the regulatory agency.

Next LT Due 3/30/2017

LEAK TEST CERTIFICATE

Wipe Date 3/29/16

DESCRIPTION OF DEVICE/SOURCE

Model 3430 Ser 33645
 SOURCE 1 SOURCE 2
 Material Cs 137 Am 241
 Source Ser 750-9313 47-4761
 RSO Name J. Domey
 Telephone ()

Do Not Write In This Space
 For HSI Use Only

REMOVABLE ACTIVITY
 Be/Gam Alpha
0 uCi 0 uCi
 Humboldt Scientific
 By [Signature]
 Date 3/29/16

Instructions

- 1 Use wipe procedures as described in the device manual.
- 2 Enter all information under description.
- 3 Print or type return address in the space provided.
- 4 Wipe source (s) and put filter paper in the plastic bag.
- 5 Keep the middle copy and mail this form and bag to Humboldt.
- 6 Regulations require that sources with removable activity greater than 0.005 uCi be removed from service for repair and decontamination or disposal. Authorities must be notified.

5200174
 Return
 Address
 Label
 Please
 Type or
 Print
 Clearly

- VT DOT
 - John Domey
 - 2178 Airport Rd Unit B
 - Berlin, VT 05641

HUMBOLDT SCIENTIFIC INC. 2525 Atlantic Ave., Raleigh, NC 27604, (919) 832-3190

Applied Health Physics, LLC

2986 Industrial Blvd., Bethel Park, Pa 15102 Phone: (412) 835-9555 Fax: (412) 835-9559
E-Mail address: knobley.ahp@comcast.net - Web Page address: WWW.APPLIEDHEALTHPHYSICS.COM

Mark V Leak Test Certificate

This certificate shall not be reproduced except in full, without the written approval of Applied Health Physics.

Device Information

Device Serial Number:	33645	Source Serial Number:	750-9313 / 47-4761		
Device Manufacturer:	Troxler	Source Manufacturer:	Troxler		
Device Model:	3430	Source Model:	3430		
Device Type:	Gauge	Make:	Moisture Density	Location:	SW Construction Mats. Lab
Radioisotope:	Cs-137 / Am-241Be	Activity:	8 mCi / 40 mCi		

Leak Test Information

Leak Test Performed by:	John Domey	Leak Test Performed on:	9/28/2015
Leak Test Specimen Number:	A-039750	Leak Test Analysis Date:	10/8/2015
Leak Test Analysis Performed by:	L. McAllister		
Leak Test Results Indicate an Activity of:	< 0.005 uCi		

Notification

Analysis of this leak test sample was provided by Applied Health Physics, LLC. Pennsylvania Department of Environmental Protection License Number PA-0228A. In the event this specimen indicates leakage and or contamination in excess of 0.005 microcuries, Applied Health Physics, LLC will notify you promptly by telephone; at such time you will need to take the appropriate actions required by the governing regulatory agency. Applied Health Physics, LLC utilizes the following for leak test (wet-swab) analysis: Model: MARK V Leak Test Kit. Applied Health Physics, LLC utilizes the following instruments for all MARK V Leak Test analysis: Protean Instruments - ASC-950 - Serial#: 01216140; Perkin Elmer - Tri-Carb 2800TR - Serial#: 073408; Ludlum Measurements - 2200 - Serial#: 25404.

Client Information

State of Vermont Agency of Transportation	Contact:	John Domey
2178 Airport Rd. Unit B	Phone:	(802) 828-6914
Materials & Research Section	Contact:	john.domey@state.vt.us
	Phone:	(802) 828-2792
Berlin	VT	05641

This certificate is an essential record and should be maintained for inspection by the regulatory agency.

Next LT Due 3/28/2016

Applied Health Physics, LLC

2986 Industrial Blvd., Bethel Park, Pa 15102 Phone: (412) 835-9555 Fax: (412) 835-9559
E-Mail address: kinobley.alp@comcast.net - Web Page address: WWW.APPLIEDHEALTHPHYSICS.COM

Mark V Leak Test Certificate

This certificate shall not be reproduced except in full, without the written approval of Applied Health Physics.

Device Information

Device Serial Number:	33645	Source Serial Number:	750-9313 / 47-4761		
Device Manufacturer:	Troxler	Source Manufacturer:	Troxler		
Device Model:	3430	Source Model:	3430		
Device Type:	Gauge	Make:	Moisture Density	Location:	Stall 3 Central Garage
Radioisotope:	Cs-137 / Am-241Be	Activity:	8 mCi / 40 mCi		

Leak Test Information

Leak Test Performed by:	John Domey	Leak Test Performed on:	2/4/2015
Leak Test Specimen Number:	A-039754	Leak Test Analysis Date:	2/11/2015
Leak Test Analysis Performed by:	L. McAllister		
Leak Test Results Indicate an Activity of:	< 0.005 uCi		

Notification

The analysis of this leak test sample was provided by Applied Health Physics, Pennsylvania Department of Environmental Protection License Number PA-0228A. In the event this specimen indicates leakage and or contamination in excess of 0.005 microcuries, Applied Health Physics will notify you promptly by telephone; at such time you will need to take the appropriate actions required by the governing regulatory agency. Applied Health Physics utilized the following for leak test analysis: Sample model: MARK V Leak Test Kit for wet swab analysis. Leak test analysis instrument; Manufacturer - Protean Instruments Model ASC950 serial number 01216140.

Client Information

State of Vermont Agency of Transportation	Contact:	John Domey
2178 Airport Rd Unit B	Phone:	(802) 828-6914
Materials & Research Section	Contact:	john.domey@state.vt.us
	Phone:	(802) 828-2792
Berlin	VT	05641

This certificate is an essential record and should be maintained for inspection by the regulatory agency.

Next LT Due 8/4/2015

Applied Health Physics

2986 Industrial Blvd., Bethel Park, Pa 15102 Phone: (412) 835-9555 Fax: (412) 835-9559
E-Mail address: ahp.inc@comcast.net - Web Page address: WWW.APPLIEDHEALTHPHYSICS.COM

RECEIVED

AUG 28 2014

Mark V Leak Test Certificate

This certificate shall not be reproduced except in full, without the written approval of Applied Health Physics.

Device Information

Device Serial Number:	33645	Source Serial Number:	750-9313 / 47-4761		
Device Manufacturer:	Troxler	Source Manufacturer:	Troxler		
Device Model:	3430	Source Model:	3430		
Device Type:	Gauge	Make:	Moisture Density	Location:	SW Regional Lab
Radioisotope:	Cs-137 / Am-241Be	Activity:	8 mCi / 40 mCi		

Leak Test Information

Leak Test Performed by:	John Domey	Leak Test Performed on:	8/8/2014
Leak Test Specimen Number:	A-037630	Leak Test Analysis Date:	8/13/2014
Leak Test Analysis Performed by:	L. McAllister		
Leak Test Results Indicate an Activity of:	< 0.005 uCi		

Notification

The analysis of this leak test sample was provided by Applied Health Physics, Pennsylvania Department of Environmental Protection License Number PA-0228A. In the event this specimen indicates leakage and or contamination in excess of 0.005 microcuries, Applied Health Physics will notify you promptly by telephone; at such time you will need to take the appropriate actions required by the governing regulatory agency. Applied Health Physics utilized the following for leak test analysis: Sample model: MARK V Leak Test Kit for wet swab analysis. Leak test analysis instrument: Manufacturer - Protean Instruments Model ASC950 serial number 01216140.

Client Information

State of Vermont Agency of Transportation	Contact:	John Domey
2178 Airport Rd Unit B	Phone:	(802) 828-6914
Materials & Research Section	Contact:	john.domey@state.vt.us
Berlin VT 05641	Phone:	(802) 828-2792

This certificate is an essential record and should be maintained for inspection by the regulatory agency.

Next LT Due 2/8/2015

Applied Health Physics

2986 Industrial Blvd., Bethel Park, Pa 15102 Phone: (412) 835-9555 Fax: (412) 835-9559
E-Mail address: ahp.inc@comcast.net - Web Page address: WWW.APPLIEDHEALTHPHYSICS.COM

Mark V Leak Test Certificate

This certificate shall not be reproduced except in full, without the written approval of Applied Health Physics.

Device Information

Device Serial Number:	33645	Source Serial Number:	750-9313 / 47-4761		
Device Manufacturer:	Troxler	Source Manufacturer:	Troxler		
Device Model:	3430	Source Model:	3430		
Device Type:	Gauge	Make:	Moisture Density	Location:	Central Garage Stall 3
Radioisotope:	Cs-137 / Am-241Be	Activity:	8 mCi / 40 mCi		

Leak Test Information

Leak Test Performed by:	John Domey	Leak Test Performed on:	2/4/2014
Leak Test Specimen Number:	A-037651	Leak Test Analysis Date:	2/10/2014
Leak Test Analysis Performed by:	L. McAllister		
Leak Test Results Indicate an Activity of:	< 0.005 uCi		

Notification

The analysis of this leak test sample was provided by Applied Health Physics. Pennsylvania Department of Environmental Protection License Number PA-0228A. In the event this specimen indicates leakage and or contamination in excess of 0.005 microcuries, Applied Health Physics will notify you promptly by telephone; at such time you will need to take the appropriate actions required by the governing regulatory agency. Applied Health Physics utilized the following for leak test analysis: Sample model: MARK V Leak Test Kit for wet swab analysis. Leak test analysis instrument: Manufacturer - Protean Instruments Model ASC950 serial number 01216140.

Client Information

State of Vermont Agency of Transportation	Contact:	John Domey
2178 Airport Rd Unit B	Phone:	(802) 828-6914
Materials & Research Section	Contact:	john.domey@state.vt.us
	Phone:	(802) 828-2792
Berlin	VT	05641

This certificate is an essential record and should be maintained for inspection by the regulatory agency.

Next LT Due 8/4/2014

Applied Health Physics, Inc.

2986 Industrial Blvd., Bethel Park, Pa 15102 Phone: (412) 835-9555 Fax: (412) 835-9559
E-Mail address: ahp.inc@comcast.net - Web Page address: WWW.APPLIEDHEALTHPHYSICS.COM

Mark V Leak Test Certificate

This certificate shall not be reproduced except in full, without the written approval of Applied Health Physics, Inc.

Device Information

Device Serial Number:	33645	Source Serial Number:	750-9313 / 47-4761		
Device Manufacturer:	Troxler	Source Manufacturer:	Troxler		
Device Model:	3430	Source Model:	3430		
Device Type:	Gauge	Make:	Moisture Density	Location:	Central Storage
Radioisotope:	Cs-137 / Am-241Be	Activity:	8 mCi / 40 mCi		

Leak Test Information

Leak Test Performed by:	Christopher Rea	Leak Test Performed on:	8/22/2013
Leak Test Specimen Number:	A-035102	Leak Test Analysis Date:	9/4/2013
Leak Test Analysis Performed by:	L. McAllister		
Leak Test Results Indicate an Activity of:	< 0.005 uCi		

Notification

The analysis of this leak test sample was provided by Applied Health Physics, Inc. Pennsylvania Department of Environmental Protection License Number PA-0228A. In the event this specimen indicates leakage and or contamination in excess of 0.005 microcuries, Applied Health Physics, Inc. will notify you promptly by telephone; at such time you will need to take the appropriate actions required by the governing regulatory agency. Applied Health Physics, Inc. utilized the following for leak test analysis: Sample model: MARK V Leak Test Kit for wet swab analysis. Leak test analysis instrument: Manufacturer - Protean Instruments Model ASC950 serial number 01216140.

Client Information

State of Vermont Agency of Transportation	Contact:	Christopher Rea
	Phone:	(802) 828-6923
1716 US Route 302	Contact:	FAX
	Phone:	(802) 828-2792
Berlin	VT	05602

This certificate is an essential record and should be maintained for inspection by the regulatory agency.

Next LT Due 2/22/2014

RECEIVED

SEP 12 2013

MATERIALS and
RESEARCH SECTION

Applied Health Physics, Inc.

2986 Industrial Blvd., Bethel Park, Pa 15102 Phone: (412) 835-9555 Fax: (412) 835-9559
E-Mail address: ahp.inc@comcast.net - Web Page address: WWW.APPLIEDHEALTHPHYSICS.COM

Mark V Leak Test Certificate

This certificate shall not be reproduced except in full, without the written approval of Applied Health Physics, Inc.

Device Information

Device Serial Number:	33645	Source Serial Number:	750-9313 / 47-4761		
Device Manufacturer:	Troxler	Source Manufacturer:	Troxler		
Device Model:	3430	Source Model:	3430		
Device Type:	Gauge	Make:	Moisture Density	Location:	Central Storage
Radioisotope:	Cs-137 / Am-241Be	Activity:	8 mCi / 40 mCi		

Leak Test Information

Leak Test Performed by:	Christopher Rea	Leak Test Performed on:	12/20/2012
Leak Test Specimen Number:	A-032606	Leak Test Analysis Date:	2/18/2013
Leak Test Analysis Performed by:	L. McAllister		
Leak Test Results Indicate an Activity of:	< 0.005 uCi		

Notification

The analysis of this leak test sample was provided by Applied Health Physics, Inc. Pennsylvania Department of Environmental Protection License Number PA-0228A. In the event this specimen indicates leakage and or contamination in excess of 0.005 microcuries, Applied Health Physics, Inc. will notify you promptly by telephone; at such time you will need to take the appropriate actions required by the governing regulatory agency. Applied Health Physics, Inc. utilized the following for leak test analysis: Sample model: MARK V Leak Test Kit for wet swab analysis. Leak test analysis instrument: Manufacturer - Protean Instruments Model ASC950 serial number 01216140.

Client Information

State of Vermont Agency of Transportation	Contact:	Christopher Rea
Materials & Research Section	Phone:	(802) 828-6923
1 National Life Drive	Contact:	FAX
	Phone:	(802) 828-2792
Montpelier	VT	05633

This certificate is an essential record and should be maintained for inspection by the regulatory agency.

Next LT Due 6/20/2013

Applied Health Physics, Inc.

2986 Industrial Blvd., Bethel Park, Pa 15102 Phone: (412) 835-9555 Fax: (412) 835-9559
E-Mail address: ahp.inc@comcast.net - Web Page address: WWW.APPLIEDHEALTHPHYSICS.COM

RECEIVED

JUN 27 2012

MATERIALS and
RESEARCH SECTION

Mark V Leak Test Certificate

This certificate shall not be reproduced except in full, without the written approval of Applied Health Physics, Inc.

Device Information

Device Serial Number:	33645	Source Serial Number:	750-9313 / 47-4761		
Device Manufacturer:	Troxler	Source Manufacturer:	Troxler		
Device Model:	3430	Source Model:	3430		
Device Type:	Gauge	Make:	Moisture Density	Location:	Central Lab
Radioisotope:	Cs-137 / Am-241Be	Activity:	8 mCi / 40 mCi		

Leak Test Information

Leak Test Performed by:	Chris Rea	Leak Test Performed on:	5/7/2012
Leak Test Specimen Number:	A-033910	Leak Test Analysis Date:	6/21/2012
Leak Test Analysis Performed by:	M. Lazzari		
Leak Test Results Indicate an Activity of:	< 0.005 uCi		

Notification

The analysis of this leak test sample was provided by Applied Health Physics, Inc. Pennsylvania Department of Environmental Protection License Number PA-0228A. In the event this specimen indicates leakage and or contamination in excess of 0.005 microcuries, Applied Health Physics, Inc. will notify you promptly by telephone; at such time you will need to take the appropriate actions required by the governing regulatory agency.

Client Information

State of Vermont Agency of Transportation	Contact:	Christopher Rea
Materials & Research Section	Phone:	(802) 828-6923
1 National Life Drive	Contact:	FAX
	Phone:	(802) 828-2792
Montpelier	VT	05633

This certificate is an essential record and should be maintained for inspection by the regulatory agency.

Next LT Due 11/7/2012