

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

RHN Scott Memorial Hospital, LLC
1451 North Gardner Street
Scottsburg, IN 47170

REPORT NUMBER(S) 2016001

2. NRC/REGIONAL OFFICE

Region III
U. S. Nuclear Regulatory Commission
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

3. DOCKET NUMBER(S)

030-33192

4. LICENSE NUMBER(S)

13-26500-01

5. DATE(S) OF INSPECTION

November 14, 2016 to
December 30, 2016

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

_____ Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Luis Nieves Folch		12/30/16
BRANCH CHIEF	Aaron T. McCraw		12/30/16

Docket File Information

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: RHN Scott Memorial Hospital, LLC 1451 North Gardner Street Scottsburg, IN 47170 REPORT NUMBER(S) 2016001	2. NRC/REGIONAL OFFICE Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352
--	---

3. DOCKET NUMBER(S) 030-33192	4. LICENSE NUMBER(S) 13-26500-01	5. DATE(S) OF INSPECTION November 14, 2016 to December 30, 2016
----------------------------------	-------------------------------------	---

6. INSPECTION PROCEDURES USED 87131	7. INSPECTION FOCUS AREAS 03.01-03.07
--	--

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 02120	2. PRIORITY 5	3. LICENSEE CONTACT Jodi Daves, RSO	4. TELEPHONE NUMBER (812) 752-8506
-----------------------------	------------------	--	---------------------------------------

Main Office Inspection Next Inspection Date: November 14, 2021
 Field Office Inspection _____
 Temporary Job Site Inspection _____

PROGRAM SCOPE

This was an unannounced routine inspection of a 100-bed hospital authorized by its NRC license to use unsealed byproduct material for diagnostic procedures under 10 CFR 35.100 and 35.200 at its facility in Scottsburg, IN. A local radiopharmacy delivered unit doses to the hospital for a variety of diagnostic procedures. The clinic employed one full-time nuclear medicine technologist who staffed the department Monday to Friday. The licensee performed approximately three diagnostic studies per day. The licensee retained the services of a medical physics consultant to perform instrument calibrations, leak tests, and quarterly audits of the radiation safety program.

PERFORMANCE OBSERVATIONS

The inspector toured the nuclear medicine laboratory and observed the nuclear medicine technologist demonstrate package receipt surveys and instrument quality control checks. The inspector performed independent surveys of the hot lab and other areas of the nuclear medicine department and found no contamination or exposures to members of the public distinguishable from background. The nuclear medicine technologist demonstrated adequate knowledge of radiation safety principles and practices through interviews. The inspector reviewed quarterly audit reports, spill reports, and documentation of package receipt, area surveys, instrument quality control, waste disposal, and employee training. The inspector also reviewed monthly dosimetry reports, which indicated annual whole-body and extremity doses below regulatory limits.

No violations of NRC requirements were identified as a result of this inspection.

Nieves Folch, Luis

From: Nieves Folch, Luis
Sent: Wednesday, January 11, 2017 12:25 PM
To: 'amy.hinds@smh1.org'
Subject: NRC clear 591 report
Attachments: image2017-01-11-125851.pdf

Mrs. Hinds

Here is the clear 591 report for the inspection conducted on November 14, 2016. No further actions are needed on your behalf.

In accordance with Title 10 of the Code of Federal Regulations 2.390 of the NRC's "Rules of Practice," a copy of this letter will be available electronically for public inspection in the NRC Public Document Room or from the NRC's Agencywide Documents Access and Management System (ADAMS), accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Please feel free to contact me if you have any questions, or if there is anything else we at the NRC can do to assist you.

Luis Nieves
Health Physicist
US Nuclear Regulatory Commission
Materials Inspection Branch, Region III
(630) 829-9571