



January 11, 2017

US Nuclear Regulatory Commission, Region III  
2443 Warrenville Road  
Suite 210  
Lisle, IL 60532-4352

Dear Sir or Madam:

Please accept this letter of notification of the closing of the Community Westview Hospital facility at 3630 Guion Road, Indianapolis, IN 46222, **License # 13-18543-01**.

On December 8, 2016 all radioactive sources were removed from the facility and the facility was surveyed. One dose calibrator source was transferred to Community Health Network and the remaining were disposed via Bionomics, Inc. Receipt acknowledgement is attached. All short-lived waste had been decayed in storage and was disposed of after surveys indicated background levels.

Following are the long-lived sources removed from the facility:

Radionuclide	mCi	Type	Assay Date	Mod/Ser#	Final disposition
Cs-137	0.000104	Rod	12/1975	076005-0589	Bionomics, Inc.
Co-57	0.000126	Rod	9/30/2011	BM080511271107	Bionomics, Inc.
Cs-137	0.188	Vial	7/25/1979	NES 356	Bionomics, Inc.
Co-57	5.9	Vial	10/10/2011	BM06057E11278150	Bionomics, Inc.
Cs-137	0.01	Disk	10/13/2014	CS1371014282113	Bionomics, Inc.
Co-57	10.0	Flood	4/27/2016	BM01L1016088304	Bionomics, Inc.
Ba-133	0.287	Vial	9/2/2011	BM06133E11256116	Community Health Network License # 13-18543-01

The facility was surveyed with a GM and the exposure levels in all areas were at background. Wipes were taken and all were below contamination limits. Please see attached map and wipe results.

All dosimetry records have been transferred to Community Health Network at 1500 North Ritter Avenue, Indianapolis, IN 46219. Employees will be notified of final dosimetry readings upon receipt of final dosimetry report.

RECEIVED JAN 18 2017



No radioactive materials remain at this facility and it is requested that this license be terminated.

If you have further questions or concerns, please contact Erin Bell, RSO of Community Health Network, who will be retaining all records from Community Westview Hospital.

Sincerely,

A handwritten signature in black ink that reads "Erin Bell". The signature is written in a cursive, flowing style.

Erin Bell, MHP  
Community Health Network  
1500 North Ritter Avenue  
Indianapolis, IN 46219  
(o) 317.355.5528  
(c) 317.313.0406



P.O. Box 817 – Kingston, TN 37763 – (865) 220-8501

January 2, 2017

**Receipt Acknowledgment**

Erin Bell  
Community Westview Hospital  
3630 Guion Road  
Indianapolis, IN 46222

Dear Erin Bell:

As required by 10 CFR Part 20 (Appendix G), this letter is notification that EnergySolutions (formerly Duratek) has received the shipment recently picked up at your facility on **December 8, 2016**.

Attached you will find a copy of your NRC Form 540, the only change from the original is in Item No. 9 "signature" which identifies that EnergySolutions is acknowledging receipt of waste from your facility.

Please keep this with your original, as well as future disposal certifications.

If you have any questions, please feel free to contact me at (865) 220-8501.

Sincerely,


*D. Agana*

Denise Agana  
Shipping/Transportation

Cc: File: 2016-23  
Manifest# 1216CWA



FORM 540 EnergySolutions, Bear Creek Processing Operations UNIFORM LOW-LEVEL RADIOACTIVE WASTE MANIFEST SHIPPING PAPER		5. SHIPPER - NAME AND FACILITY Community Westview Hospital 3630 Guion Road Indianapolis, IN 46222			SHIPMENT ID NUMBER 1216CWhA	7. FORM 540 AND 540A FORM 541 AND 541A FORM 542 AND 542A ADDITIONAL INFORMATION		PAGE 1 OF 1 PAGE(S) 1 PAGE(S) NONE PAGE(S) NONE PAGE(S)	8. MANIFEST NUMBER (Use this number on all continuation pages) 1216CWhA
1. EMERGENCY TELEPHONE NUMBER (Include Area Code) (865) 220-8520		USER PERMIT NUMBER		SHIPMENT NUMBER 1216CWhA	X GENERATOR TYPE (Specify) M	9. CONSIGNEE - Name and Facility Address EnergySolutions, Bear Creek Processing Operations Operated By EnergySolutions 1560 Bear Creek Road Oak Ridge, TN 37830		CONTACT Fred Schulz TELEPHONE NUMBER (Include Area Code) (865) 481-0222	
ORGANIZATION Bionomics, Inc. Attn: Emergency Duty Officer		CONTACT Erin Bell		TELEPHONE NUMBER (Include Area Code) (317) 355-5528		SIGNATURE - Authorized consignee acknowledging waste receipt <i>[Signature]</i>		DATE 12/19/16	
2. IS THIS AN "EXCLUSIVE USE" SHIPMENT? [X] YES [ ] NO		3. TOTAL NUMBER OF PACKAGES IDENTIFIED ON THIS MANIFEST 1		6. CARRIER - Name and Address Bionomics, Inc. 1550 Bear Creek Road Oak Ridge, TN 37830		Truck #: Trailer #:		EPA I.D. NUMBER TND982116493	
4. DOES EPA REGULATED WASTE REQUIRING A MANIFEST ACCOMPANY THIS SHIPMENT? [ ] YES [X] NO If "Yes", provide Manifest Number →		EPA MANIFEST NUMBER N/A		CONTACT John McCormick		SHIPPING DATE 12/08/2016		TELEPHONE NUMBER (Include Area Code) 865-220-8501	
11. U.S. DEPARTMENT OF TRANSPORTATION DESCRIPTION (Including proper shipping name, hazard class, UN ID number, and any additional information)		12. DOT LABEL "RADIOACTIVE"		13. TRANSPORT INDEX		14. PHYSICAL AND CHEMICAL FORM		15. INDIVIDUAL RADIONUCLIDES	
UN3321, Radioactive material, low specific activity (LSA-II), 7 PLASTIC 1 - FIBER BOX		NA		NA		SOLID/OXIDES		CO-57	
								16. TOTAL PACKAGE ACTIVITY MBq : 2.2940E+02 mCi : (6.2000E+00)	
								17. LSA/SCO CLASS LSA II	
								18. TOTAL WEIGHT OR VOLUME (Use appropriate units) 3.88 ft <sup>3</sup> 8.00 lb	
								19. IDENTIFICATION NUMBER OF PACKAGE CWH-2 (16-004407)	
FOR CONSIGNEE USE ONLY				20. Generator Certification Statement					
Tennessee "License For Delivery" No. _____				A) Radioactive Materials. Certification is hereby made that this shipment of low-level radioactive waste has been prepared in accordance with a radioactive waste management program which has been approved by the Nuclear Regulatory Commission or an Agreement State regulatory agency and with the current revision of the site Material Acceptance Criteria.					
South Carolina Transport Permit No. _____				B) Hazardous Materials. Generator hereby certifies that this material _____ does (or) Z does not contain a hazardous waste as defined in 40 CFR 261.					
US Ecology Generator No. _____				C) Data. Generator hereby represents and warrants that all data set forth in this (UNIFORM LOW-LEVEL RADIOACTIVE WASTE MANIFEST) are true and correct in all respects and in accordance with all applicable governmental laws, rules, regulations and site Radioactive Material Licenses.					
US Ecology Permit No. _____				D) INFECTIOUS SUBSTANCE: Generator hereby certifies that this material does not contain an infectious substance as defined in 49CFR 173.134					
				ERIN PELL Print Name		Erin Bell Signature		12/8/16 Date	





# Community Westview

3630 Guion Rd  
Indianapolis, IN 46222

## Wipe Test Detail Report

Printed on: 12/8/2016 11:04:38 AM

Report Options: Today, For Area: Weekly Wipes

Thursday, December 08, 2016

Area: Weekly Wipes

Type: Restricted (50%)

Location : **1 (Pass)**

Time Stamp : 12/8/2016 10:50 AM

Count Time : 30 (secs)

Performed By : Jodi Hartzler

Detector : Well - 241 (S/N: 241)

Isotope	Background	MDA	Net CPM	Net Activity	Trigger	Status
Wide Window	959 DPM	353 DPM	-6.0	-17 DPM	20,000 DPM	Pass
I-123 (Well)	81 DPM	9 DPM	-12.0	-19 DPM	2,000 DPM	Pass
I-131	141 DPM	19 DPM	-2.0	-7 DPM	2,000 DPM	Pass
Tc-99m	42 DPM	7 DPM	-16.0	-20 DPM	2,000 DPM	Pass

Location : **10 (Pass)**

Time Stamp : 12/8/2016 11:00 AM

Count Time : 30 (secs)

Performed By : Jodi Hartzler

Detector : Well - 241 (S/N: 241)

Isotope	Background	MDA	Net CPM	Net Activity	Trigger	Status
Wide Window	959 DPM	353 DPM	58.0	160 DPM	20,000 DPM	Pass
I-123 (Well)	81 DPM	9 DPM	12.0	19 DPM	2,000 DPM	Pass
I-131	141 DPM	19 DPM	-8.0	-28 DPM	2,000 DPM	Pass
Tc-99m	42 DPM	7 DPM	-4.0	-5 DPM	2,000 DPM	Pass

Location : **11 (Pass)**

Time Stamp : 12/8/2016 11:01 AM

Count Time : 30 (secs)

Performed By : Jodi Hartzler

Detector : Well - 241 (S/N: 241)

Isotope	Background	MDA	Net CPM	Net Activity	Trigger	Status
Wide Window	959 DPM	353 DPM	-24.0	-66 DPM	20,000 DPM	Pass
I-123 (Well)	81 DPM	9 DPM	-14.0	-23 DPM	2,000 DPM	Pass
I-131	141 DPM	19 DPM	-8.0	-28 DPM	2,000 DPM	Pass
Tc-99m	42 DPM	7 DPM	-2.0	-2 DPM	2,000 DPM	Pass

Location : **12 (Pass)**

Time Stamp : 12/8/2016 11:02 AM

Count Time : 30 (secs)


Performed By : Jodi Hartzler

Detector : Well - 241 (S/N: 241)


Isotope	Background	MDA	Net CPM	Net Activity	Trigger	Status
Wide Window	959 DPM	353 DPM	-30.0	-83 DPM	20,000 DPM	Pass
I-123 (Well)	81 DPM	9 DPM	4.0	6 DPM	2,000 DPM	Pass
I-131	141 DPM	19 DPM	-16.0	-57 DPM	2,000 DPM	Pass
Tc-99m	42 DPM	7 DPM	-2.0	-2 DPM	2,000 DPM	Pass

Comments :

Signature :



Date Signed :



# Community Westview

3630 Guion Rd  
Indianapolis, IN 46222

Location : **13 (Pass)**

Time Stamp : 12/8/2016 11:03 AM

Count Time : 30 (secs)

Performed By : Jodi Hartzler

Detector : Well - 241 (S/N: 241)

Isotope	Background	MDA	Net CPM	Net Activity	Trigger	Status
Wide Window	959 DPM	353 DPM	6.0	17 DPM	20,000 DPM	Pass
I-123 (Well)	81 DPM	9 DPM	-12.0	-19 DPM	2,000 DPM	Pass
I-131	141 DPM	19 DPM	2.0	7 DPM	2,000 DPM	Pass
Tc-99m	42 DPM	7 DPM	2.0	2 DPM	2,000 DPM	Pass

Location : **2 (Pass)**

Time Stamp : 12/8/2016 10:52 AM

Count Time : 30 (secs)

Performed By : Jodi Hartzler

Detector : Well - 241 (S/N: 241)

Isotope	Background	MDA	Net CPM	Net Activity	Trigger	Status
Wide Window	959 DPM	353 DPM	48.0	132 DPM	20,000 DPM	Pass
I-123 (Well)	81 DPM	9 DPM	2.0	3 DPM	2,000 DPM	Pass
I-131	141 DPM	19 DPM	2.0	7 DPM	2,000 DPM	Pass
Tc-99m	42 DPM	7 DPM	-10.0	-12 DPM	2,000 DPM	Pass

Location : **3 (Pass)**

Time Stamp : 12/8/2016 10:53 AM

Count Time : 30 (secs)

Performed By : Jodi Hartzler

Detector : Well - 241 (S/N: 241)

Isotope	Background	MDA	Net CPM	Net Activity	Trigger	Status
Wide Window	959 DPM	353 DPM	62.0	171 DPM	20,000 DPM	Pass
I-123 (Well)	81 DPM	9 DPM	-12.0	-19 DPM	2,000 DPM	Pass
I-131	141 DPM	19 DPM	-6.0	-21 DPM	2,000 DPM	Pass
Tc-99m	42 DPM	7 DPM	-8.0	-10 DPM	2,000 DPM	Pass

Location : **4 (Pass)**

Time Stamp : 12/8/2016 10:54 AM

Count Time : 30 (secs)

Performed By : Jodi Hartzler

Detector : Well - 241 (S/N: 241)

Isotope	Background	MDA	Net CPM	Net Activity	Trigger	Status
Wide Window	959 DPM	353 DPM	82.0	226 DPM	20,000 DPM	Pass
I-123 (Well)	81 DPM	9 DPM	-6.0	-10 DPM	2,000 DPM	Pass
I-131	141 DPM	19 DPM	-2.0	-7 DPM	2,000 DPM	Pass
Tc-99m	42 DPM	7 DPM	-10.0	-12 DPM	2,000 DPM	Pass

Location : **5 (Pass)**

Time Stamp : 12/8/2016 10:55 AM

Count Time : 30 (secs)


Performed By : Jodi Hartzler

Detector : Well - 241 (S/N: 241)

Isotope	Background	MDA	Net CPM	Net Activity	Trigger	Status
Wide Window	959 DPM	353 DPM	-20.0	-55 DPM	20,000 DPM	Pass
I-123 (Well)	81 DPM	9 DPM	-22.0	-35 DPM	2,000 DPM	Pass
I-131	141 DPM	19 DPM	-4.0	-14 DPM	2,000 DPM	Pass
Tc-99m	42 DPM	7 DPM	-24.0	-30 DPM	2,000 DPM	Pass

Comments :

Signature :



Date Signed :

12/8/16



# Community Westview

3630 Guion Rd  
Indianapolis, IN 46222

Location : **6 (Pass)**

Time Stamp : 12/8/2016 10:56 AM

Count Time : 30 (secs)

Performed By : Jodi Hartzler

Detector : Well - 241 (S/N: 241)

Isotope	Background	MDA	Net CPM	Net Activity	Trigger	Status
Wide Window	959 DPM	353 DPM	28.0	77 DPM	20,000 DPM	Pass
I-123 (Well)	81 DPM	9 DPM	-2.0	-3 DPM	2,000 DPM	Pass
I-131	141 DPM	19 DPM	8.0	28 DPM	2,000 DPM	Pass
Tc-99m	42 DPM	7 DPM	-2.0	-2 DPM	2,000 DPM	Pass

Location : **7 (Pass)**

Time Stamp : 12/8/2016 10:57 AM

Count Time : 30 (secs)

Performed By : Jodi Hartzler

Detector : Well - 241 (S/N: 241)

Isotope	Background	MDA	Net CPM	Net Activity	Trigger	Status
Wide Window	959 DPM	353 DPM	-14.0	-39 DPM	20,000 DPM	Pass
I-123 (Well)	81 DPM	9 DPM	0.0	0 DPM	2,000 DPM	Pass
I-131	141 DPM	19 DPM	-14.0	-50 DPM	2,000 DPM	Pass
Tc-99m	42 DPM	7 DPM	4.0	5 DPM	2,000 DPM	Pass

Location : **8 (Pass)**

Time Stamp : 12/8/2016 10:58 AM

Count Time : 30 (secs)

Performed By : Jodi Hartzler

Detector : Well - 241 (S/N: 241)

Isotope	Background	MDA	Net CPM	Net Activity	Trigger	Status
Wide Window	959 DPM	353 DPM	68.0	187 DPM	20,000 DPM	Pass
I-123 (Well)	81 DPM	9 DPM	0.0	0 DPM	2,000 DPM	Pass
I-131	141 DPM	19 DPM	4.0	14 DPM	2,000 DPM	Pass
Tc-99m	42 DPM	7 DPM	0.0	0 DPM	2,000 DPM	Pass

Location : **9 (Pass)**

Time Stamp : 12/8/2016 10:59 AM

Count Time : 30 (secs)

Performed By : Jodi Hartzler

Detector : Well - 241 (S/N: 241)

Isotope	Background	MDA	Net CPM	Net Activity	Trigger	Status
Wide Window	959 DPM	353 DPM	40.0	110 DPM	20,000 DPM	Pass
I-123 (Well)	81 DPM	9 DPM	0.0	0 DPM	2,000 DPM	Pass
I-131	141 DPM	19 DPM	-4.0	-14 DPM	2,000 DPM	Pass
Tc-99m	42 DPM	7 DPM	-12.0	-15 DPM	2,000 DPM	Pass

Location : **Hot Lab (1-5) (Pass)**

Time Stamp : 12/8/2016 10:44 AM

Count Time : 30 (secs)


Performed By : Jodi Hartzler

Detector : Well - 241 (S/N: 241)


Isotope	Background	MDA	Net CPM	Net Activity	Trigger	Status
Wide Window	959 DPM	353 DPM	10.0	28 DPM	20,000 DPM	Pass
I-123 (Well)	81 DPM	9 DPM	2.0	3 DPM	2,000 DPM	Pass
I-131	141 DPM	19 DPM	-12.0	-42 DPM	2,000 DPM	Pass
Tc-99m	42 DPM	7 DPM	2.0	2 DPM	2,000 DPM	Pass

Comments :

Signature :



Date Signed :








**Community**  
Health Network

*BRW BEN WOOD*  
Community Hospital East  
1500 North Ritter Avenue  
Indianapolis, Indiana 46219-3095

RETURN  
SERVICE  
REQUESTED

neopost<sup>®</sup> FIRST-CLASS MAIL  
01/11/2017 AUTO  
US POSTAGE \$000.64<sup>8</sup>  
 ZIP 46219  
041M12250930

US Nuclear Regulatory Commission  
2443 Warrenville Road  
Suite 210  
Lisle, IL 60532-4352

RECEIVED JAN 18 2017