

**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE/LOCATION INSPECTED:

Greater Flint Imaging  
G-1071 North Ballenger Highway  
Flint, MI 48504

REPORT NUMBER(S) 2017001

2. NRC/REGIONAL OFFICE

Region III  
U. S. Nuclear Regulatory Commission  
2443 Warrenville Road, Suite 210  
Lisle, IL 60532-4352

3. DOCKET NUMBER(S)

030-18247

4. LICENSE NUMBER(S)

21-20320-01

5. DATE(S) OF INSPECTION

January 11, 2017

**LICENSEE:**

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.  
(Violations and Corrective Actions)

**Statement of Corrective Actions**

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Edward F. Harvey		1/11/17
BRANCH CHIEF	Aaron T. McCraw		1/26/17

**Docket File Information**

**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE/LOCATION INSPECTED:  Greater Flint Imaging G-1071 North Ballenger Highway Flint, MI 48504  REPORT NUMBER(S) 2017001	2. NRC/REGIONAL OFFICE  Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352
---	---

3. DOCKET NUMBER(S)  030-18247	4. LICENSE NUMBER(S)  21-20320-01	5. DATE(S) OF INSPECTION  January 11, 2017
--------------------------------------	---	--

6. INSPECTION PROCEDURES USED  87130	7. INSPECTION FOCUS AREAS  All
--	--------------------------------------

**SUPPLEMENTAL INSPECTION INFORMATION**

1. PROGRAM CODE(S)  02201	2. PRIORITY  5	3. LICENSEE CONTACT  Gurkan Ege, MD, RSO	4. TELEPHONE NUMBER  (810) 235-0668
---------------------------------	----------------------	--	---

Main Office Inspection      Next Inspection Date: January 11, 2022

Field Office Inspection \_\_\_\_\_

Temporary Job Site Inspection \_\_\_\_\_

**PROGRAM SCOPE**

This was a routine, unannounced inspection of a private medical clinic authorized to use byproduct materials for diagnostic medical procedures under 10 CFR 35.100 and 35.200. The licensee performed approximately 8-12 patient studies per month with one part-time nuclear medicine technologist. The licensee performed procedures as needed. Therefore, there were no fixed days of the week that were designated for administrations. The licensee received unit doses from a licensed nuclear pharmacy.

The inspector observed nuclear medicine staff demonstrate package receipt and surveying procedures, daily dose calibrator constancy checks, daily surveys, and waste disposal procedures. All patient procedures scheduled for the day had been completed prior to the start of the inspection; however the nuclear medicine technologist demonstrated adequate knowledge of radiation protection principles and emergency procedures in the event of a spill when explaining the administration process to the inspector.

The inspector reviewed a selection of licensee records, including program audits, dose calibrator linearity records, dose calibrator accuracy records, source inventories, survey meter calibration records, package receipt logs, and dosimetry with no issues noted. In addition, the inspector performed independent surveys, which revealed no readings that would indicate residual contamination or exposures to members of the public in excess of regulatory limits.

No violations were identified during this inspection.