

**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE/LOCATION INSPECTED: United Hospital Center (UHC) 327 Medical Park Drive Bridgeport, West Virginia 26330		2. NRC/REGIONAL OFFICE  U.S. Nuclear Regulatory Commission Region I, 2100 Renaissance Blvd, Suite 100 King of Prussia, Pennsylvania 19406-2713	
REPORT NUMBER(S) 2017001			
3. DOCKET NUMBER(S) 030-03375	4. LICENSE NUMBER(S) 47-01458-01	5. DATE(S) OF INSPECTION February 8, 16, 2017	

**LICENSEE:**

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s) and corrective action(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with the NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11. (Violations and Corrective Actions)



10 CFR 20.1802 requires, in part, that the licensee shall control and maintain constant surveillance of licensed material that is in a controlled or unrestricted area and that is not in storage. Contrary to the above, United Hospital Center did not control and maintain constant surveillance of licensed material that was in a controlled or unrestricted area. Specifically, on February 8, 2017, 18 millicuries of labeled Tc-99m was left unattended in an injection area, a controlled area, for a short period of time while the technologist left to retrieve a patient from the waiting room.

This is a Severity Level IV violation. (Section 6.7)

Normally UHC has one technologist control the dosage and one technologist retrieve the patient. If an occasion arises again where a single technologist must do both, UHC agreed to modify their procedure to bring the patient to the injection area first and then retrieve the dosage from the hot lab for injection so that it would not be left unattended.

**Statement of Corrective Actions**

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE	John Fernandez		2/16/17
NRC INSPECTOR	Robin Elliott		2/16/2017
BRANCH CHIEF	James P. Dwyer		2/19/2017

SUNSI Review Completed By:  / RA / Robin Elliott

Public  Non-Sensitive