

WEEKLY WIPES

Month: *February 2017*
 Equipment: Well Counter Captus 3000

Week	Bkgd	#1 Hot Lab	#2 Lab Door	#3 Vertex	#4 CVL	#5 Discovery	#6 Inj Table & Chair	#7 Unrestricted Area	Tech Initials	Date
	Trigger Level	1000 dpm	1000 dpm	1000 dpm	1000 dpm	1000 dpm	1000 dpm	1000 dpm		
1	0	0	0	34	44	18	0	0	NK	2-1-17
2	0	306	318	166	0	0	49	44	NK	2-8-17
3	0	29	32	0	0	18	0	20	CKC	2-11-17
4	0	0	29	0	0	140	57	40	MSO	2-13-17
5										

Daily Area Survey

Month: April

Year: 2017

Instrument: GM Survey Meter

Serial # 326566 Check source value: 19 mR/hr

p.3

Day	Bkgd	#1 Hot Lab	#2 Lab Door	#3 Camera 1	#4 CVL	#5 Camera 2	#6 Inj Table & Chair	#7 Unrestricted Area	#8 Dose Calibrator	#9 Hot Storage	#10 Hot Trash	Tech Initials
	Level (mR/hr)	1	1	1	0.1	1	1	0.1	0.1	0.1	0.1	
1	0.04	0.03	0.04		0.05	0.05	0.04	0.04	0.04	0.05	0.04	RJK
2												
3	0.03	0.03	0.03		0.03	0.03	0.03	0.03	0.03	0.03	0.03	NK
4	0.02	0.02	0.02		0.02	0.02	0.02	0.02	0.02	0.02	0.02	NK
5	0.02	0.02	0.02		0.02	0.02	0.02	0.02	0.02	0.02	0.02	NK
6	0.02	0.02	0.02		0.02	0.02	0.02	0.02	0.02	0.02	0.02	NK
7	0.02	0.02	0.02		0.02	0.02	0.02	0.02	0.02	0.02	0.02	NK
8												
9												
10	0.03	0.03	0.03		0.03	0.03	0.03	0.03	0.03	0.03	0.03	NK
11	0.03	0.03	0.03		0.03	0.03	0.03	0.03	0.03	0.03	0.03	NK
12	0.02	0.02	0.02		0.02	0.02	0.02	0.02	0.02	0.02	0.02	NK
13	0.02	0.02	0.02		0.02	0.02	0.02	0.02	0.02	0.02	0.02	NK
14	0.02	0.02	0.02		0.02	0.02	0.02	0.02	0.02	0.02	0.02	NK
15	0.02	0.02	0.02		0.02	0.02	0.02	0.02	0.02	0.02	0.02	NK
16												
17	0.03	0.03	0.03		0.03	0.03	0.03	0.03	0.03	0.05	0.03	NK
18	0.03	0.03	0.03		0.03	0.03	0.03	0.03	0.03	0.03	0.03	NK
19	0.03	0.03	0.03		0.03	0.03	0.03	0.03	0.03	0.03	0.03	NK
20	0.03	0.03	0.03		0.03	0.03	0.03	0.03	0.03	0.03	0.03	NK
21												
22												
23												
24												
25												
26												
27												
28												
29												
30												
31												

Daily Area Survey

Month: March

Year: 2017

Instrument: GM Survey Meter #1

Serial # 326566

Check source value: 19 mR/hr

P. 4

Day	Bkgd	#1 Hot Lab	#2 Lab Door	#3 Vertex	#4 CVL	#5 Discovery	#6 Inj Table & Chair	#7 Unrestricted Area	#8 Dose Calibrator	#9 Hot Storage	#10 Hot Trash	Tech Initials
	Trigger Level (mR/hr)	1	1	1	0.1	1	1	0.1	0.1	0.1	0.1	
1	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	JK
2	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.04	JK
3	0.05	0.05	0.06	0.05	0.05	0.05	0.06	0.05	0.08	0.07	0.05	JK
4	0.05	0.05	0.05	0.05	0.04	0.05	0.06	0.05	0.06	0.05	0.05	JK
5												
6	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03	JK
7	0.05	0.07	0.07	0.05	0.05	0.05	0.05	0.05	0.08	0.07	0.05	JK
8	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.08	0.07	0.05	JK
9	0.05	0.05	0.04	0.04	0.05	0.05	0.05	0.05	0.05	0.05	0.05	JK
10	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.04	JK
11												
12				Vertex removed								
13	0.04	0.04	0.04	NA	0.04	0.04	0.04	0.04	0.04	0.04	0.04	JK
14	0.04	0.04	0.04		0.04	0.04	0.04	0.04	0.04	0.04	0.04	JK
15	0.03	0.05	0.03		0.03	0.03	0.03	0.03	0.04	0.04	0.04	JK
16	0.03	0.03	0.03		0.03	0.03	0.03	0.03	0.03	0.03	0.03	JK
17	0.03	0.03	0.03		0.03	0.03	0.03	0.03	0.03	0.03	0.03	JK
18	0.03	0.03	0.03		0.03	0.03	0.03	0.03	0.05	0.05	0.03	JK
19												
20	0.05	0.06	0.05		0.05	0.05	0.05	0.05	0.05	0.09	0.05	JK
21	0.02	0.02	0.02		0.02	0.02	0.02	0.02	0.02	0.02	0.02	JK
22	0.02	0.03	0.02		0.02	0.02	0.02	0.02	0.02	0.02	0.02	JK
23	0.03	0.03	0.03		0.03	0.03	0.03	0.03	0.03	0.03	0.03	JK
24	0.02	0.02	0.02		0.02	0.02	0.02	0.02	0.02	0.02	0.02	JK
25												
26												
27	0.03	0.03	0.03		0.03	0.03	0.03	0.03	0.03	0.03	0.02	JK
28	0.03	0.03	0.03		0.03	0.03	0.03	0.03	0.03	0.03	0.03	JK
29	0.03	0.03	0.03		0.03	0.03	0.03	0.03	0.03	0.03	0.03	JK
30	0.03	0.03	0.03		0.03	0.03	0.03	0.03	0.03	0.03	0.03	JK
31	0.03	0.03	0.03	✓	0.03	0.03	0.03	0.03	0.02	0.03	0.03	JK

HP LASERJET FAX

APR 21 2017 9:19AM



Nuclear Medicine
Phone: 765-453-8597
Fax: 765-864-8729

FACSIMILE TRANSMISSION COVER SHEET

Date: 4-21-17 Fax #: 630-515-1078
To: Magdalena
From: Watill

Number of pages (including this page): 3

If there are any transmission problems, please call (765) 453-8597

CONFIDENTIALITY NOTICE

The cover sheet and the materials enclosed with this transmission are the private confidential property of Community Howard Regional Health, and the materials are privileged communications intended solely for the receipt, use, benefit and information of the intended recipient. If you are not the intended recipient, you are hereby notified that any review, disclosure, copying, distribution, or the taking of any other action in reliance on the contents of this transmission is strictly prohibited, and may result in legal liability on your part. If you have received this transmission in error, please notify us immediately at the telephone number above and arrange for return of this transmission to us. Thank You

Comments: