



P.O. Box 606
Kingwood, WV 26537
304.329.2129

mountaineercontractors.com

Unites States Nuclear Regulatory Commission
Region 1
King of Prussia, PA 19406-1415

April 19, 2017

Re: Amend NRC License No. 47-17517-02

Dear Sirs,

I have reviewed the requested amendment and agree with the request to remove the storage location at 1127 Short Creek Road, Wheeling, WV from our license. Please feel free to contact me at (304) 329-2129 if any additional information is needed.

Sincerely,

A handwritten signature in black ink that reads 'Robert E. Leigh'.

Robert E. Leigh, P.E.
Vice President



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
Research Triangle Park, NC 27709
Tel: (877) 876-9537 Fax: (866) 391-2759
License: NC 032-0182-1

MIKE NEELY
MOUNTAINEER CONTRACTORS
P.O. BOX 606
KINGWOOD, WV 26537

Cust ID: 5459

LEAK TEST CERTIFICATE

DEVICE:

Model: 3430 **Serial No:** 66715

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
77-11897	06/05/2012	Cs-137	0.296	8
78-7941	06/29/2012	Am-241:Be	1.48	40

LEAK TEST ANALYSIS:

Sample collected on: 03/13/2017
Sample analyzed on: 03/23/2017 9:58:12 AM **Position:** 15
Analyzed by: STW

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.24E+01	1.96E+01
Background measurement (cpm)	0	25
Sample measurement (cpm)	0	19
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.2E-01	1.3E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi) **Greater than 185 Bq (0.005 uCi)**