

From: [Chris Apperson](#)
To: [Lawyer, Dennis](#); mark.gessel@tritonwv.com
Cc: "Chris Apperson"
Subject: [External_Sender] RE: Removal of location from NRC License
Date: Friday, April 21, 2017 12:59:30 PM
Attachments: [image003.png](#)
[SKM_C554e17042112450.pdf](#)

Mr. Lawyer:

Please find attached leak tests. I apologize for the delay.

Thanks,

Chris

Chris Apperson
(304) 951-7941



From: Lawyer, Dennis [mailto:Dennis.Lawyer@nrc.gov]
Sent: Tuesday, April 18, 2017 6:10 AM
To: Chris.apperson@tritonwv.com; mark.gessel@tritonwv.com
Subject: FW: Removal of location from NRC License

Good Morning Mr. Apperson and Mr. Gessel,

To date, to the best of my knowledge, you have not sent us the requested information to complete the requested amendment (See attached emails). I am quickly running out of time to complete this action within management expectation. I need to have your leak tests by the end of the day on April 17, 2017, in order to meet management expectations. Please send me a signed letter and your leak tests scanned in a PDF attached to an email and mailed to me at dennis.lawyer@nrc.gov. If it is not possible to do this by that date (one week), please send me a letter requesting a withdrawal of the amendment request dated February 14, 2017. You may request for this again, once you have the leak test information. Please send the letter by scanning to a PDF and emailing me.

Please call me if you cannot comply with the above for some reason.

Dennis Lawyer
U.S. NRC Region 1
Health Physicist
610-337-5366

From: Lawyer, Dennis

Sent: Wednesday, April 05, 2017 9:39 AM

To: 'Chris.apperson@tritonwv.com' <Chris.apperson@tritonwv.com>

Subject: Removal of location from NRC License

Mr. Apperson,

I am not sure that I am communicating my request or not as I had not received any communications back from Mr. Gessel. Please email me back that you have received this email.

Attached is the email that I sent on March 2 and a follow up on March 29.

I am looking for a response for my request for information on my email dated March 2. To date, I don't believe that we have received anything.

Dennis Lawyer
U.S. NRC Region 1
Health Physicist
610-337-5366



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
Research Triangle Park, NC 27709
Tel: (877) 876-9537 Fax: (866) 391-2759
License: NC 032-0182-1

WILLIAM GESSEL
TRITON CONSTRUCTION CO
PO BOX 1360
ST. ALBANS, WV 25177

Cust ID: 12321



LEAK TEST CERTIFICATE

DEVICE:

Model: 3401 Serial No: 7717

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
40-4929	09/22/1980	Cs-137	0.296	8
47-4013	06/12/1980	Am-241:Be	1.48	40

LEAK TEST ANALYSIS:

Sample collected on: 01/15/2017
Sample analyzed on: 02/09/2017 10:40:29 A Position: 7
Analyzed by: STW

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.24E+01	1.96E+01
Background measurement (cpm)	0	25
Sample measurement (cpm)	0	27
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.2E-01	1.3E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi) Greater than 185 Bq (0.005 uCi)

W.M.A.
2-16-17



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
Research Triangle Park, NC 27709
Tel: (877) 876-9537 Fax: (866) 391-2759
License: NC 032-0182-1

WILLIAM GESSEL
TRITON CONSTRUCTION CO
PO BOX 1360
ST. ALBANS, WV 25177

Cust ID: 12321



LEAK TEST CERTIFICATE

DEVICE:

Model: 3401 **Serial No:** 16901

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
47-12324	01/05/1989	Am-241:Be	1.48	40
50-6069	01/05/1989	Cs-137	0.296	8

LEAK TEST ANALYSIS:

Sample collected on: 01/15/2017
Sample analyzed on: 02/09/2017 10:41:41 A **Position:** 8
Analyzed by: STW

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.24E+01	1.96E+01
Background measurement (cpm)	0	25
Sample measurement (cpm)	0	24
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.2E-01	1.3E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi) **Greater than 185 Bq (0.005 uCi)**

W.M.G.
2-16-17