

Dykert, Jason

From: Chester Douglas
Sent: Tuesday, May 02, 2017 7:14 AM
To: Dykert, Jason
Subject: NRC document request
Attachments:

Dear Mr. Dykert,

I'm sending this email as a follow up to our conversation on April 20th. Based on the results of the investigation, there doesn't seem to be a clear indication that the source was fully locked out prior to John Grizzell leaving the facility on 6-16-16. Monsanto is looking to have the shield removed soon and it will be coming back to our facility for an evaluation. Once it arrives we will be conducting some hands on training in hopes of preventing any similar future incidents. As soon as we were made aware of the issue, I put into place a suspension of distribution pending the investigation results. If it is determined that the suspension be lifted, there will also be a requirement for all future shield orders to have an additional position indicator attached (see attached photo). Please let me know if you have any questions or require any additional information.

Sincerely,

Chester Douglas, RRPT
Radiation Safety Officer (RSO)
Berthold Technologies USA, LLC
99 Midway Lane
Oak Ridge, TN 37830