

From: [Lanzisera, Penny](#)
To: [Butler, Wayne](#)
Subject: Request for Additional Information
Date: Monday, April 24, 2017 4:05:00 PM

Licensee: Wheeling Hospital
License No. 47-05322-02
Docket No. 03012570
Mail Control No. 594397

To support removal of the licensed material from your license, please submit the following additional information:

1. Documentation from Bionomics showing receipt of the cesium sources.
2. Leak test of the cesium sources prior to transfer.
3. Confirmation that no new cesium sources will be ordered, and hence, 110 mCi is sufficient for a total possession limit.
4. Date of last use of Yttrium 90 microspheres.

To support revision of your spot-check procedures, please submit step-by-step procedures for checks with acceptance criteria (e.g., for door interlock, procedure may be extend active source, push door open, and verify that source retracts to the shielded position).

You may submit a signed pdf to my attention. Please reference Mail Control No. 594397 in your response. Thank you for your assistance.

Penny Lanzisera
Senior Health Physicist
US NRC, Region I