

JUL 19 1995

License No. 37-28518-01  
Docket No. 030-31690  
Control No. 121859

Larry McKinney, P.E.  
Michael Baker, Jr., Inc.  
Construction Management Services  
3600 Var Tan Way  
Second Floor  
Harrisburg, Pennsylvania 17110-9440

Dear Mr. McKinney:

This is in reference to your application dated May 25, 1995 requesting renewal of License No. 37-28518-01. In order to continue our review, we need the following additional information:

1. Confirm that you will notify the NRC, in accordance with 10 CFR 30.50(b)(2), if a sealed source becomes lodged in a hole and it becomes apparent that recovery efforts will be unsuccessful.
2. Submit (1) the name and radiation safety qualifications of the individual who will conduct audits, (2) a description of the scope and extent of the audits, (3) a commitment to conduct audits at intervals not to exceed 12 months and to maintain records of the audits for at least 3 years after the record is made, (4) management's commitment to review the documented results of the audit promptly after the audit's completion, and (5) a commitment to take prompt action to correct deficiencies identified during audits, to inform all personnel (including those at other locations and those working under other licenses) of the deficiencies and the actions management expects its personnel to take to avoid similar deficiencies.
3. State the frequency for calibrating your radiation detection equipment. An annual calibration frequency is acceptable.

We will continue our review upon receipt of this information. Please reply in duplicate to my attention at the Region I office and refer to Mail Control No. 121859. If you have any technical questions regarding this deficiency letter, please call me at (610) 337-5251.

L. McKinney, P.E.  
Michael Baker, Jr., Inc.

-2-

In order to continue prompt review of your application, we request that you submit your response to this letter within 30 calendar days from the date of this letter.

Sincerely,

**Original Signed By:**

Kathleen Dolce  
Nuclear Materials Safety Branch  
Division of Radiation Safety  
and Safeguards

License No. 37-28518-01  
Docket No. 030-31690  
Control No. 121859

Enclosures:

1. 10 CFR Parts 19, 20, 21, 30 and 170.
2. Draft Regulatory Guide DG-0008 entitled "Applications For the Use of Sealed Sources in Portable Gauging Devices"
3. NRC Forms 3 and 313

DOCUMENT NAME: S:\PENDING\BAKER.KD1

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DRSS/RI	N					
NAME	KDolce						
DATE	07/19/95	07/	/95	07/	/95		

OFFICIAL RECORD COPY