

From: [Miller, John](#)
To: jscala@tectonicengineering.com
Subject: Tectonic Engineering Request for Additional Information Mail Control No. 594717
Date: Wednesday, May 17, 2017 3:51:00 PM

Tectonic Engineering & Surveying Consultants P.C.
License No. 06-31339-01
Docket No. 030-37849
Control No. 594717

Mr. Scala

In order to continue the review of your application to amend your NRC license, we need additional information. If you have any questions, please contact me (610-337-5089).

In your letter dated May 4, 2017, you requested that facilities located at 1229 Albany Avenue, Hartford, CT and 1209 Main Street, Coventry, CT be added to your license as places of use/storage. Please provide a facility diagram for each portable gauge storage location. Include on the diagram the use of adjacent areas (including above and below) and information relevant to public dose and security as discussed in sections 8.10.5, "Public Dose," and 8.10.6, "Operating , Emergency, and security Procedures," respectively, in NUREG-1556, Vol. 1, Rev. 2, "Consolidated Guidance About Portable Gauge Licenses." A link to the document has been included for your convenience.

<https://www.nrc.gov/docs/ML1617/ML16175A375.pdf>

Please respond in writing with a signed and dated letter. Please submit the signed response document as an attachment to an e-mail. Please be sure to include the Control No. 594717 on your response. We would appreciate a response within 30 days in order to expedite review of the license amendment request.

Thank you for your cooperation.

John J. Miller
Health Physicist
U.S. Nuclear Regulatory Commission – Region 1
Division of Nuclear Material Safety
2100 Renaissance Boulevard
King of Prussia, PA 19406-2713
Office 610-337-5089
Cell 610-329-0734
Fax 610-337-5269
E-mail john.miller@nrc.gov