



APAC-Missouri, Inc., Missouri Division
P.O. Box 1117 • Columbia, Missouri 65205 • (573) 228-7015

May 22, 2017

Jennifer Bishop
Region III
US Nuclear Regulatory Commission
DNMS/Materials Licensing Branch
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

Re: License No. 24-32769-01

Jennifer,

I have reviewed my file of the Leak Test Certificates sent to me from the buyer, Jefferson Asphalt, the NRC licensed company that purchased the nuclear gauges from APAC-Missouri, Inc. This should be my final response to number 2 on the enclosed NRC Form 699 Conversation Record dated 4/04/17.

I have included the requested the Leak Test Certificates for the two serial numbers machines L652 and 21082. These certificates are for tests performed 4/6/16 which would have been done at the time of the asset sale to Jefferson Asphalt. The sale was finalized 4/29/16.

Sincerely,

A handwritten signature in black ink, appearing to read "Warren L. Harms". The signature is fluid and cursive, with a large initial "W" and a long, sweeping underline.

Warren L. Harms
Construction Controller
APAC-Missouri, Inc.

Seaman Nuclear Corporation

7315 South First Street Oak Creek, WI 53154 USA

Tel 414-762-5100 Fax 414-762-5106

info@seamannuclear.com

Leak Test Certificate

April 12, 2016

A leak test has been performed on meter, serial number **L652**, a model C-200, containing the radionuclide Ra 226

Owned By: APAC-Missouri Inc. (Columbia)
P.O. Box 1117
Columbia, MO 65205


Date Sample Collected: 4/6/16
Collected By: Patrick White
Date Sample Analyzed: 4/12/16
Analyzed By: K. Krysiak

Most regulatory agencies consider a source to be leaking if a leak test reveals the presence of more than 0.005 microcurie of removable contamination.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by Wisconsin license 079-1257-01.

LEAK TEST DUE: 10/6/16



Scott C. Seaman
Radiation Safety Officer
Seaman Nuclear Corporation

Seaman Nuclear Corporation

7315 South First Street Oak Creek, WI 53154 USA

Tel 414-762-5100 Fax 414-762-5106

info@seamannuclear.com

Leak Test Certificate

April 12, 2016

A leak test has been performed on meter, serial number **21082**, a model C-300, containing the radionuclide Ra 226

Owned By: APAC-Missouri Inc. (Columbia)
P.O. Box 1117
Columbia, MO 65205


Date Sample Collected: 4/6/16
Collected By: Patrick White
Date Sample Analyzed: 4/12/16
Analyzed By: K. Krysiak

Most regulatory agencies consider a source to be leaking if a leak test reveals the presence of more than 0.005 microcurie of removable contamination.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by Wisconsin license 079-1257-01.

LEAK TEST DUE: 10/6/16



Scott C. Seaman
Radiation Safety Officer
Seaman Nuclear Corporation

Bishop, Jennifer

From: Harms Warren (APAC Missouri) <Warren.Harms@apac.com>
Sent: Monday, May 22, 2017 3:32 PM
To: Bishop, Jennifer
Subject: [External_Sender] APAC-Missouri Sale to Jefferson Asphalt
Attachments: NRC 2 Leak Test Certs-Machines to Jefferson Asphalt.pdf; Seaman leak test serial 21082 Apr 16.jpg; Seaman leak test serial L652 Apr 16.jpg; NRC Form 699 4_04_17.pdf

Jennifer,

Attached is a cover letter with the two remaining Leak Test Certificates for the machines APAC-Missouri, Inc. sold to Jefferson Asphalt Company on April 29, 2016

You can let me know if there is any other information you need, or I will consider our responses complete to the NRC Form 699 dated 4/04/2017.

Warren L. Harms
Construction Controller
APAC-Kansas City
APAC-Missouri

Warren.Harms@apac.com
573-228-7015 cell