



CONVERSATION RECORD

5/18/17

NAME OF PERSON(S) CONTACTED OR IN CONTACT WITH YOU Victor Hosfeld		DATE OF CONTACT 4/20/17	TYPE OF CONVERSATION <input type="checkbox"/> E-MAIL <input checked="" type="checkbox"/> TELEPHONE <input type="checkbox"/> INCOMING <input type="checkbox"/> OUTGOING
E-MAIL ADDRESS		TELEPHONE NUMBER (989) 839-1407	

ORGANIZATION MidMichigan Medical Center - Alpena	DOCKET NUMBER(S) 030-13274
---	-------------------------------

LICENSE NUMBER(S) 21-17754-01	CONTROL NUMBER(S) 593196
----------------------------------	-----------------------------

SUBJECT  
Conversation Record

SUMMARY

On 4/20/17, reviewer spoke with Victor Hosfeld (RSO for MidMichigan Health, NRC license no. 21-01549-02) and requested a letter describing the history of use (what isotopes were used, quantites, history of spills) and the most recent radiation surveys and weekly wipe tests for the areas where radioactive material is being used at the MidMichigan Medical Center - Alpena. Mr. Hosfeld indicated that he would provide the information. The licensee provided the requested information on May 9, 2017 (ML17130A744).

On 5/10/17, the reviewer spoke with Jennifer, one of the NMTs at the faciltiy who explained there are two teadmll rooms located in the cardiopulmonary department which is not part of the radiology/nuclear medicine area. The licensee takes radiation readings and wipe test of both rooms at the same time. She also provided calibration dated for Bicorn Surveyor 2000B (1/16/17), Fluke 451P (4/30/17). On 5/18/17, M. Andrzejewski called and provided calibration dates for the AtomLab well counter (1/11/17 and 4/27/17).

**Continue on Page 2**

ACTION REQUIRED (IF ANY)

**Continue on Page 3**

NAME OF PERSON DOCUMENTING CONVERSATION  
Magdalena Gryglak

SIGNATURE  
*Magdalena R Gryglak*