

Herrera, Tomas

From: david_bennett@agilent.com
Sent: Monday, June 05, 2017 11:16 AM
To: Herrera, Tomas
Subject: [External_Sender] Addendum request to SS&D Registry Amendment
Attachments: NRC Letter 4_12_17 Amendment Request to Add Little Falls Back to SS&D.pdf; G2397A.png; G2398B.png; G4597A.png; G4598A.png

Good morning Mr. Herrera,

On April 12'th of this year, Agilent submitted an application to amend our SS&D Registry to reflect a change in manufacturing location (first attachment). Just recently, I realized that we did not include upgraded copies of the labels to be attached to our electron capture detectors (ECDs) that should have been included with the amendment request. Can I submit this e-mail as an addendum request submitting the attached photos of the updated labels, with the statement "MADE IN USA" instead of "MADE IN CHINA"? Additionally, as proposed in the letter dated April 12'th, since we do not have an exact date of transition from our manufacturing site in China to here in the US, we request that these photos be added to the registry until the process has been transferred and we can then remove both the China manufacturing location and labels reflecting MADE IN CHINA.

Regards,
Dave



CAUTION
RADIOACTIVE MATERIAL
555 MEGABECQUERELS/
15 MILLICURIES OF NICKEL-63

G2397A ELECTRON CAPTURE DETECTOR

SERIAL NO.	DATE
------------	------



Agilent Technologies

 SEE THE GC AND ECD MANUALS BEFORE OPERATING OR SERVICING THIS DEVICE.

THIS DEVICE MUST BE WIPE (RADIOACTIVE LEAK) TESTED AT INTERVALS NOT GREATER THAN SIX MONTHS.

THE RECEIPT, POSSESSION, USE AND TRANSFER OF THIS DEVICE ARE SUBJECT TO A GENERAL LICENSE OR THE EQUIVALENT, AND THE REGULATIONS OF THE U.S. NRC OR OF A STATE WITH WHICH THE NRC HAS ENTERED INTO AN AGREEMENT FOR THE EXERCISE OF REGULATORY AUTHORITY. OUTSIDE U.S., CONSULT THE APPROPRIATE AGENCIES TO DETERMINE THE EQUIVALENT REGULATIONS.

THE LABEL SHALL BE MAINTAINED ON THIS DEVICE IN A LEGIBLE CONDITION. REMOVAL OF THIS LABEL IS PROHIBITED.

MADE IN USA



CAUTION
RADIOACTIVE MATERIAL
555 MEGABECQUERELS/
15 MILLICURIES OF NICKEL-63

G2398A ELECTRON CAPTURE DETECTOR

SERIAL NO.	DATE
------------	------



⚠ SEE THE GC AND ECD MANUALS BEFORE OPERATING OR SERVICING THIS DEVICE.

THIS DEVICE MUST BE WIPE (RADIOACTIVE LEAK) TESTED AT INTERVALS NOT GREATER THAN SIX MONTHS.

THE RECEIPT, POSSESSION, USE AND TRANSFER OF THIS DEVICE ARE SUBJECT TO A GENERAL LICENSE OR THE EQUIVALENT, AND THE REGULATIONS OF THE U.S. NRC OR OF A STATE WITH WHICH THE NRC HAS ENTERED INTO AN AGREEMENT FOR THE EXERCISE OF REGULATORY AUTHORITY. OUTSIDE U.S., CONSULT THE APPROPRIATE AGENCIES TO DETERMINE THE EQUIVALENT REGULATIONS.

THE LABEL SHALL BE MAINTAINED ON THIS DEVICE IN A LEGIBLE CONDITION. REMOVAL OF THIS LABEL IS PROHIBITED.

MADE IN USA



CAUTION
RADIOACTIVE MATERIAL
555 MEGABECQUERELS/
15 MILLICURIES OF NICKEL-63

G4597A ELECTRON CAPTURE DETECTOR

SERIAL NO.	DATE
------------	------



Agilent Technologies

⚠ SEE THE GC AND ECD MANUALS BEFORE OPERATING OR SERVICING THIS DEVICE.

THIS DEVICE MUST BE WIPE (RADIOACTIVE LEAK) TESTED AT INTERVALS NOT GREATER THAN SIX MONTHS.

THE RECEIPT, POSSESSION, USE AND TRANSFER OF THIS DEVICE ARE SUBJECT TO A GENERAL LICENSE OR THE EQUIVALENT, AND THE REGULATIONS OF THE U.S. NRC OR OF A STATE WITH WHICH THE NRC HAS ENTERED INTO AN AGREEMENT FOR THE EXERCISE OF REGULATORY AUTHORITY. OUTSIDE U.S., CONSULT THE APPROPRIATE AGENCIES TO DETERMINE THE EQUIVALENT REGULATIONS.

THE LABEL SHALL BE MAINTAINED ON THIS DEVICE IN A LEGIBLE CONDITION. REMOVAL OF THIS LABEL IS PROHIBITED.

MADE IN USA



CAUTION

RADIOACTIVE MATERIAL
555 MEGABECQUERELS/
15 MILLICURIES OF NICKEL-63

G4598A ELECTRON CAPTURE DETECTOR

SERIAL NO.	DATE
------------	------



SEE THE GC AND ECD MANUALS BEFORE
 OPERATING OR SERVICING THIS DEVICE.

**THIS DEVICE MUST BE WIPE (RADIOACTIVE LEAK)
TESTED AT INTERVALS NOT GREATER THAN SIX
MONTHS.**

**THE RECEIPT, POSSESSION, USE AND TRANSFER
OF THIS DEVICE ARE SUBJECT TO A GENERAL
LICENSE OR THE EQUIVALENT, AND THE
REGULATIONS OF THE U.S. NRC OR OF A STATE
WITH WHICH THE NRC HAS ENTERED INTO AN
AGREEMENT FOR THE EXERCISE OF REGULATORY
AUTHORITY. OUTSIDE U.S., CONSULT THE
APPROPRIATE AGENCIES TO DETERMINE THE
EQUIVALENT REGULATIONS.**

**THE LABEL SHALL BE MAINTAINED ON THIS
DEVICE IN A LEGIBLE CONDITION. REMOVAL OF
THIS LABEL IS PROHIBITED.**

MADE IN USA