

CONVERSATION RECORD

|TIME |DATE 5/26/17

VISIT

CONFERENCE

x TELEPHONE

X INCOMING
0 OUTGOING

NAME OF PERSON(S) CONTACTED OR IN CONTACT ORGANIZATION (OFFICE, DEPT.ETC.)
John Snider, RSO Anderson Engineering (Docket No.: 030-17919)
TELEPHONE NO.: 417-540-2992

SUBJECT

The licensee asked if it was acceptable to receive a CPN MC Series Gauge (Portaprobe) with Source Models CPN-131 including Cs-137 and Am-241 from another portable gauge licensee

SUMMARY

The inspector discussed 10 CFR 30.41 regarding transfer of the gauges from a portable gauge licensee to the licensee. In addition, the inspector reviewed Anderson Engineering's NRC license to verify that it was authorized for the CPN MC Series Gauge (Portaprobe) with Source Models CPN-131 including Cs-137 and Am-241.

ACTION REQUIRED

NAME OF PERSON DOCUMENTING CONVERSATION
Bob Gattone

SIGNATURE

6/6/17

Robert B. Gattone, Jr.

ACTION TAKEN

SIGNATURE

TITLE

DATE