77)

LICENSEE EVENT REPORT

CONTROL BLOCK: (1) (PLEASE PRINT OR TYPE ALL REQUIRED INFORMATION)

1 | I L D R S 2 (2) | 0 0 - 0 0 0 0 0 - 0 0 (3) | 4 1 1 1 1 (4) | (5)

8 9 LICENSEE CODE 14 15 LICENSE NUMBER 25 26 LICENSE TYPE 30 57 CAT 58

ON'T

1 | REPORT SOURCE L (6) | 0 5 0 0 0 2 3 7 (7) | 0 1 1 4 8 0 (8) | 0 2 1 3 8 0 (9)

8 60 61 DOCKET NUMBER 68 69 EVENT DATE 74 75 REPORT DATE 80

EVENT DESCRIPTION AND PROBABLE CONSEQUENCES (10)

2 During routine surveillance of Reactor Water Level switch LIS-2-263-58A, a setpoint of

3 67.8 inches DP was found. This is above T.S. limit (Table 3.2.1) of 66.8 inches DP.

4 This was of minimal safety significance since redundant switches were operable. There

5 was no effect on public health or safety. Similar events reported in R.O. 76-57/03L-0

6 and 77-61/03L-0, Docket 050-237.

7

8

8 9 80

9 | SYSTEM CODE I A (11) | CAUSE CODE E (12) | CAUSE SUBCODE E (13) | COMPONENT CODE I N S T R U (14) | COMP. SUBCODE S (15) | VALVE SUBCODE Z (16)

8 9 10 11 12 13 18 19 20

(17) LER/RO REPORT NUMBER: EVENT YEAR 8 0 (21 22) | — (23) | SEQUENTIAL REPORT NO. 0 0 3 (24 26) | / (27) | OCCURRENCE CODE 0 3 (28 29) | REPORT TYPE L (30) | — (31) | REVISION NO. 0 (32)

ACTION TAKEN E (18) | FUTURE ACTION Z (19) | EFFECT ON PLANT Z (20) | SHUTDOWN METHOD Z (21) | HOURS (22) 0 0 0 0 (33 34 35 36 37 40) | ATTACHMENT SUBMITTED N (23) | NPRD-4 FORM SUB. Y (24) | PRIME COMP. SUPPLIER N (25) | COMPONENT MANUFACTURER Y 0 0 5 (26)

33 34 35 36 37 40 41 42 43 44 47

CAUSE DESCRIPTION AND CORRECTIVE ACTIONS (27)

0 The cause was due to instrument drift of yarway level instrument switch model 4418C

1 serial number 24699. The switch was inspected with no problems noted. The switch

2 was promptly reset to 65 +/- 1 inch DP. The reactor low water level switches will

3 continue to be tested monthly using DIS 500-2.

4

8 9 80

5 | FACILITY STATUS E (28) | % POWER 0 9 8 (29) | OTHER STATUS (30) NA | METHOD OF DISCOVERY B (31) | DISCOVERY DESCRIPTION (32) Instrument Mechanic Surveillance

8 9 10 12 13 44 45 46 80

6 | ACTIVITY RELEASED Z (33) | CONTENT OF RELEASE Z (34) | AMOUNT OF ACTIVITY (35) NA | LOCATION OF RELEASE (36) NA

8 9 10 11 44 45 80

7 | PERSONNEL EXPOSURES NUMBER 0 0 0 (37) | TYPE Z (38) | DESCRIPTION (39) NA

8 9 11 12 13 80

8 | PERSONNEL INJURIES NUMBER 0 0 0 (40) | DESCRIPTION (41) NA

8 9 11 12 80

9 | LOSS OF OR DAMAGE TO FACILITY TYPE Z (42) | DESCRIPTION (43) NA

8 9 10 80

0 | PUBLICITY ISSUED N (44) | DESCRIPTION (45) NA | NRC USE ONLY

8 9 10 68 69 80

8002200538

B. Hall

Xc-421

NRC 017-926