



APPLICATION FOR MATERIALS LICENSE

Estimated burden per response to comply with this mandatory collection request: 4.3 hours. Submittal of the application is necessary to determine that the applicant is qualified and that adequate procedures exist to protect the public health and safety. Send comments regarding burden estimate to the FOIA, Privacy, and Information Collections Branch (T-5 F53), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by e-mail to Infocollections.Resource@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3150-0120), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

INSTRUCTIONS: SEE THE CURRENT VOLUMES OF THE NUREG-1556 TECHNICAL REPORT SERIES ("CONSOLIDATED GUIDANCE ABOUT MATERIALS LICENSES") FOR DETAILED INSTRUCTIONS FOR COMPLETING THIS FORM: <http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/>. SEND TWO COPIES OF THE COMPLETED APPLICATION TO THE NRC OFFICE SPECIFIED BELOW.

APPLICATION FOR DISTRIBUTION OF EXEMPT PRODUCTS FILE APPLICATIONS WITH:

MATERIALS SAFETY LICENSING BRANCH
DIVISION OF MATERIAL SAFETY, STATE, TRIBAL AND RULEMAKING PROGRAMS
OFFICE OF NUCLEAR MATERIALS SAFETY AND SAFEGUARDS
U.S. NUCLEAR REGULATORY COMMISSION
WASHINGTON, DC 20555-0001

ALL OTHER PERSONS FILE APPLICATIONS AS FOLLOWS:

IF YOU ARE LOCATED IN:

ALABAMA, CONNECTICUT, DELAWARE, DISTRICT OF COLUMBIA, FLORIDA, GEORGIA, KENTUCKY, MAINE, MARYLAND, MASSACHUSETTS, NEW HAMPSHIRE, NEW JERSEY, NEW YORK, NORTH CAROLINA, PENNSYLVANIA, PUERTO RICO, RHODE ISLAND, SOUTH CAROLINA, TENNESSEE, VERMONT, VIRGINIA, VIRGIN ISLANDS, OR WEST VIRGINIA,

SEND APPLICATIONS TO:

LICENSING ASSISTANCE TEAM
DIVISION OF NUCLEAR MATERIALS SAFETY
U.S. NUCLEAR REGULATORY COMMISSION, REGION I
2100 RENAISSANCE BOULEVARD, SUITE 100
KING OF PRUSSIA, PA 19406-2713

IF YOU ARE LOCATED IN:

ILLINOIS, INDIANA, IOWA, MICHIGAN, MINNESOTA, MISSOURI, OHIO, OR WISCONSIN, SEND APPLICATIONS TO:

MATERIALS LICENSING BRANCH
U.S. NUCLEAR REGULATORY COMMISSION, REGION III
2443 WARRENVILLE ROAD, SUITE 210
LISLE, IL 60532-4352

ALASKA, ARIZONA, ARKANSAS, CALIFORNIA, COLORADO, HAWAII, IDAHO, KANSAS, LOUISIANA, MISSISSIPPI, MONTANA, NEBRASKA, NEVADA, NEW MEXICO, NORTH DAKOTA, OKLAHOMA, OREGON, PACIFIC TRUST TERRITORIES, SOUTH DAKOTA, TEXAS, UTAH, WASHINGTON, OR WYOMING,

SEND APPLICATIONS TO:

NUCLEAR MATERIALS LICENSING BRANCH
U.S. NUCLEAR REGULATORY COMMISSION, REGION IV
1600 E. LAMAR BOULEVARD
ARLINGTON, TX 76011-4511

PERSONS LOCATED IN AGREEMENT STATES SEND APPLICATIONS TO THE U.S. NUCLEAR REGULATORY COMMISSION ONLY IF THEY WISH TO POSSESS AND USE LICENSED MATERIAL IN STATES SUBJECT TO U.S. NUCLEAR REGULATORY COMMISSION JURISDICTIONS.

1. THIS IS AN APPLICATION FOR (Check appropriate item)

A. NEW LICENSE

B. AMENDMENT TO LICENSE NUMBER

13-32526

C. RENEWAL OF LICENSE NUMBER

2. NAME AND MAILING ADDRESS OF APPLICANT (Include ZIP code)

Indiana University Health West
Hospital
1111 Ronald Reagan Parkway
AVON, IN 46023

3. ADDRESS WHERE LICENSED MATERIAL WILL BE USED OR POSSESSED

4. NAME OF PERSON TO BE CONTACTED ABOUT THIS APPLICATION

ROBERT T. ANGER, JR., RSO

BUSINESS TELEPHONE NUMBER

317-581-1911

BUSINESS CELLULAR TELEPHONE NUMBER

317-370-2750

BUSINESS EMAIL ADDRESS

ranger@mpcphysius.com

SUBMIT ITEMS 5 THROUGH 11 ON 8-1/2 X 11" PAPER. THE TYPE AND SCOPE OF INFORMATION TO BE PROVIDED IS DESCRIBED IN THE LICENSE APPLICATION GUIDE.

5. RADIOACTIVE MATERIAL

a. Element and mass number; b. chemical and/or physical form; and c. maximum amount which will be possessed at any one time.

6. PURPOSE(S) FOR WHICH LICENSED MATERIAL WILL BE USED.

7. INDIVIDUAL(S) RESPONSIBLE FOR RADIATION SAFETY PROGRAM AND THEIR TRAINING AND EXPERIENCE.

8. TRAINING FOR INDIVIDUALS WORKING IN OR FREQUENTING RESTRICTED AREAS.

9. FACILITIES AND EQUIPMENT.

10. RADIATION SAFETY PROGRAM.

11. WASTE MANAGEMENT.

12. LICENSE FEES (Fees required only for new applications, with few exceptions*)

(See 10 CFR 170 and Section 170.31)

*Amendments/Renewals that increase the scope of the existing license to a new or higher fee category will require a fee.

FEE CATEGORY

AMOUNT ENCLOSED \$

13. CERTIFICATION. (Must be completed by applicant) THE APPLICANT UNDERSTANDS THAT ALL STATEMENTS AND REPRESENTATIONS MADE IN THIS APPLICATION ARE BINDING UPON THE APPLICANT.

THE APPLICANT AND ANY OFFICIAL EXECUTING THIS CERTIFICATION ON BEHALF OF THE APPLICANT, NAMED IN ITEM 2, CERTIFY THAT THIS APPLICATION IS PREPARED IN CONFORMITY WITH TITLE 10, CODE OF FEDERAL REGULATIONS, PARTS 30, 32, 33, 34, 35, 36, 37, 39, AND 40, AND THAT ALL INFORMATION CONTAINED HEREIN IS TRUE AND CORRECT TO THE BEST OF THEIR KNOWLEDGE AND BELIEF.

WARNING: 18 U.S.C. SECTION 1001 ACT OF JUNE 25, 1948 62 STAT. 749 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

CERTIFYING OFFICER -- TYPED/PRINTED NAME AND TITLE

Doug Puckett, CEO

SIGNATURE

D & Puckett

DATE

6/20/17

FOR NRC USE ONLY

TYPE OF FEE	FEE LOG	FEE CATEGORY	AMOUNT RECEIVED	CHECK NUMBER	COMMENTS
			\$		
APPROVED BY				DATE	

RECEIVED JUN 26 2017



Indiana University Health

June 20, 2017

Materials Licensing Branch
U.S. Nuclear Regulatory Commission Region III
2443 Warrenville Road, Suite 210
Lisle, IL 60532

re: Amendment Request for NRC License Number 13-20338-01

Please amend the Indiana University Health West Hospital byproduct materials license 13-20338-01 as follows:


Delete Yun Wang, Ph.D as Radiation Safety Officer

Add Robert T. Anger, Jr., M.S., MPH as Radiation Safety Officer

Mr. Anger was previously listed as the RSO for Indiana University Health West Hospital and is currently the RSO at Indiana University Health North Hospital (see enclosed NRC byproduct materials license number 13-32602-01). Also enclosed is a signed copy of the RSO/Management Letter of Understanding and a copy of the Duties and Responsibilities of the Radiation Safety Officer.

If you have any questions regarding this request, please contact Mr. Anger at 317-370-2750 or ranger@mpcphysics.com.

Sincerely,


Doug Puckett
Chief Executive Officer

IU Health West Hospital
1111 North Ronald Reagan Parkway
Avon, IN 46123

iuhealth.org

Indiana University Health West Hospital
1111 Ronald Reagan Parkway
Avon, IN 46123

RSO / EXECUTIVE MANAGEMENT
LETTER OF UNDERSTANDING

June 20, 2017

Robert T. Anger, Jr., M.S., DABR
Methodist Hospital
1701 N. Senate Blvd.
Indianapolis, Indiana 46202

Re: Radiation Safety Officer / Executive Management
Letter of Understanding

Mr. Anger:

You have been appointed the Radiation Safety Officer (RSO) of this facility for our United States Nuclear Regulatory Commission Materials License. This "Letter of Understanding" is prepared to comply with Title 10 Code of Federal Regulations (CFR) Part 35.24(b). This section of the regulations requires that you agree in writing to the following:


- > Assume responsibility for implementing the radiation protection program.
- > Ensure that radiation safety activities are being performed in accordance with our own approved procedures and all regulatory requirements.

Furthermore, in compliance with 10 CFR 35.24(e),(g), the executive management of this facility agrees to provide you as RSO:


- > Specific written notation of your authority, duties and responsibilities (see attached).
- > Sufficient authority, organizational freedom, time, resources and management prerogative to:
 1. Identify radiation safety problems;
 2. Initiate, recommend, or provide corrective actions;
 3. Stop unsafe operations;
 4. Verify implementation of corrective actions; and,
 5. ensure compliance with regulations.

Our signatures noted below will attest to the issues noted above.

Sincerely,



Chief Executive Officer



Robert T. Anger, Jr., M.S., DABR
Radiation Safety Officer

DUTIES AND RESPONSIBILITIES OF THE RADIATION SAFETY OFFICER

The RSO shall ensure that:

- Unsafe activities involving licensed material are stopped;
- Radiation exposures are ALARA;
- Up-to-date radiation protection procedures in the daily operation of the hospital's byproduct material program are developed, distributed, and implemented;
- Possession, use, and storage of licensed material is consistent with the limitations in the license, the regulations, the SDDR Certificate(s), and the manufacturer's recommendations and instructions;
- Individuals installing, relocating, maintaining, adjusting, or repairing devices containing sealed sources are trained and authorized by an NRC or Agreement State license;
- Personnel training is conducted and is commensurate with the individual's duties regarding licensed material;
- Documentation is maintained to demonstrate that individuals are not likely to receive, in one year, a radiation dose in excess of 10% of the allowable limits or that personnel monitoring devices are provided;
- When necessary, personnel monitoring devices are used and exchanged at the proper intervals, and records of the results of such monitoring are maintained;
- Licensed material is properly secured;
- Documentation is maintained to demonstrate, by measurement or calculation, that the total effective dose equivalent to the individual likely to receive the highest dose from the licensed operation does not exceed the annual limit for members of the public;
- Proper authorities are notified of incidents such as loss or theft of licensed material, damage to or malfunction of sealed sources, and fire;
- Medical events are investigated and reported to the NRC, and cause(s) and appropriate corrective action(s) are identified, and timely corrective action(s) are taken;
- Audits of the radiation safety program are performed at least annually and documented;
- If violations of regulations, license conditions, or program weaknesses are identified, effective corrective actions are developed, implemented, and documented;
- Licensed material is transported, or offered for transport, in accordance with all applicable DOT requirements;
- Licensed material is disposed of properly;
- Appropriate records are maintained;
- An up-to-date license is maintained and amendment and renewal requests are submitted in a timely manner

RADIATION SAFETY OFFICER AUTHORITY, DUTIES AND RESPONSIBILITIES

The Radiation Safety Officer (RSO) shall:

1. Have the authority to implement the Radiation Protection Program as referenced in 10 CFR 20.1101.
2. Have the authority, organizational freedom, time, resources, and management prerogative to:
 - a. Identify radiation safety problems;
 - b. Initiate, recommend or provide corrective actions,
 - c. Stop unsafe operations; and,
 - d. Verify implementation of corrective actions.
3. Investigate deviations from the radiation safety practices approved by facility management and/or the Radiation Safety Committee, if applicable.
4. Collect in a centralized location, executive management approved procedures that can include policy and technical issues which, would makeup the Radiation Protection Program as follows:
 - a. Authorization for the purchase of radioactive material.
 - b. Receipt and opening of packages containing radioactive material.
 - c. Storage of radioactive material.
 - d. Inventory control of radioactive material.
 - e. Safe use of radioactive material.
 - f. Emergency procedures in the event of loss, theft, etc.
 - g. Periodic radiation surveys and wipe tests
 - h. Checks of radiation survey and other radiation safety instruments.
 - i. Disposal of radioactive material.
 - j. Personnel training of those who work in or frequent areas of radioactive material use or storage.
5. Oversee a record system of the Radiation Protection Program per 10 CFR 20.2102 to include at least the following:

The provisions of the Radiation Protection Program until the license is terminated by the NRC such as:

- a. All records, reports, written policies and procedures required by regulatory agencies concerning radioactive material.
- b. A copy of the regulations governing the possession, use and disposal of licensed material, such as Title 10 Code of Federal Regulations.

Audits and other reviews of the Radiation Protection Program content and implementation for a period of three (3) years after the record is made.

6. Periodically evaluate "action levels" for continued appropriateness to ensure compliance with 10 CFR 20.1501 and 1502 for the following:
 - a. Personnel exposure investigation levels
 - b. Area surveys dose rate and contamination levels
 - c. Bioassays, if necessary
 - d. Radioactive effluent concentrations, if necessary
7. Review the following Radiation Protection Program records, if applicable:
 - a. Sealed source inventories
 - b. Sealed source leak tests
 - c. Dose calibrator linearity tests
 - d. Dose calibrator accuracy tests
 - e. Dose calibrator geometrical variation tests
 - f. Occupational radiation exposure reports
 - g. Medical event documentation
 - h. Spill / incident reports for cause and corrective action
 - i. Dose rate and contamination survey results
 - j. Changes in the radiation safety program
8. Ensure the use of reasonable practices and controls to strive to maintain doses to workers and to the public are ALARA, in compliance with 10 CFR 20.1101(b).
9. Review with facility management at least annually of the content of the Radiation Protection Program and determine if the written program is being implemented in compliance with 10 CFR 20.1101(c).
10. Ensure as a part of the ALARA effort that individual members of the public shall not receive a Total Effective Dose Equivalent (TEDE) of more than 10 mrem (0.1 mSv) per year from airborne radioactive material releases as per 10 CFR 20.1101(d) as necessary.
11. Be a member of the Radiation Safety Committee (RSC), if applicable, that will oversee all uses of byproduct material permitted by the license as per 10 CFR 35.24(f).

MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 37, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

<p style="text-align: center;">Licensee</p> <p>1. Indiana University Health North Hospital</p> <p>2. 11700 North Meridian Street Carmel, IN 46032</p>	<p>In accordance with the letter dated September 10, 2015,</p> <p>3. License number 13-32602-01 is renewed in its entirety to read as follows:</p> <hr/> <p>4. Expiration date February 28, 2026</p> <hr/> <p>5. Docket No. 030-37059 Reference No.</p>
---	---

<p>6. Byproduct, source, and/or special nuclear material</p> <p>A. Any byproduct material permitted by 10 CFR 35.100</p> <p>B. Any byproduct material permitted by 10 CFR 35.200</p> <p>C. Any byproduct material permitted by 10 CFR 35.300</p>	<p>7. Chemical and/or physical form</p> <p>A. Any</p> <p>B. Any</p> <p>C. Any</p>	<p>8. Maximum amount that licensee may possess at any one time under this license</p> <p>A. As needed</p> <p>B. As needed</p> <p>C. 300 millicuries</p>
--	---	---

9. Authorized use:
- A. Any uptake, dilution and excretion study permitted by 10 CFR 35.100.
 - B. Any imaging and localization study permitted by 10 CFR 35.200.
 - C. Any diagnostic study or therapy procedure permitted by 10 CFR 35.300.

CONDITIONS

- 10. Licensed material may be used or stored only at the licensee's facilities located at 11700 North Meridian Street, Carmel, Indiana.
- 11. The Radiation Safety Officer for this license is Robert T. Anger, Jr.
- 12. Licensed material is only authorized for use by, or under the supervision of:
 - A. Individuals permitted to work as an authorized user, authorized nuclear pharmacist, and/or authorized medical physicist in accordance with 10 CFR 35.13 and 35.14.

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License Number
13-32602-01

Docket or Reference Number
030-37059

Amendment No. 04

B. The following individuals are authorized users for medical use as indicated:

Authorized Users

Material and Use

James Fletcher, M.D.

10 CFR 35.100, 35.200, and 35.300.

Mark Tann, M.D.

10 CFR 35.100, 35.200, and 35.300.

Steven Westphal, M.D.

10 CFR 35.100, 35.200, and 35.300.

Justin B. Sims, M.D.

10 CFR 35.100, 35.200, and 35.300, limited to the oral administration of sodium iodide I-131.

Yazid Fadl, M.D.

10 CFR 35.200.

Ronald Mastouri, M.D.

10 CFR 35.200.

Mark Estrada, M.D.

10 CFR 35.100, 35.200, and 35.300.

Vasantha Aaron, M.D.

10 CFR 35.100, 35.200, and 35.300.

13. The licensee is authorized to transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."

14. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. This license condition applies only to those procedures that are required to be submitted in accordance with the regulations. Additionally, this license condition does not limit the licensee's ability to make changes to the radiation protection program as provided for in 10 CFR 35.26. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.

A. Letter dated September 10, 2015. (ML15260B195)

B. Letter dated February 10, 2016. (ML16043A275)

FOR THE U.S. NUCLEAR REGULATORY COMMISSION

Date FEB 23 2016

By



Jennifer L. Bishop
Materials Licensing Branch
Region III



Indiana University Health

IU Health West Hospital
Diagnostic Imaging
1111 North Ronald Reagan Parkway
Avon, IN 46123

INDIANAPOLIS

111-460

21 JUN '17

PM 5:1

neopost[®]

06/21/2017

US POSTAGE

FIRST-CLASS MAIL

\$000.67⁰



ZIP 46241
041M1251029

Materials Licensing Branch
U.S. Nuclear Regulatory Commission, Region III
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

RECEIVED JUN 26 2017

60532-435260

