U. S. NUCLEAR REGULATORY COMMISSION

LICENSEE EVENT REPORT

CONTROL BLOCK: (1) (PLEASE PRINT OR TYPE ALL REQUIRED INFORMATION)

1 | ILDRS3 (2) | 0 0 - 0 0 (3) | 4 1 1 1 1 (4) | (5)
LICENSEE CODE | LICENSE NUMBER | LICENSE TYPE | CAT

CON'T

REPORT SOURCE: L (6) | 05000249 (7) | 02 18 81 (8) | 03 13 81 (9)
DOCKET NUMBER | EVENT DATE | REPORT DATE

EVENT DESCRIPTION AND PROBABLE CONSEQUENCES (10)

During unit normal operation, while conducting an instrument surveillance, refueling floor ARM Ch 'A' tripped at 105 mr/hr. This exceeded 100 mr/hr limit as listed by Tech. Spec. 3.2.E.3. The ARM setpoint was adjusted. There was no hazard to the public since redundant monitor was operable. Similar events were reported by Docket 50-237 RO 80-41/03L and Docket 50-249 RO 80-12/03L.

SYSTEM CODE (11)	CAUSE CODE (12)	CAUSE SUBCODE (13)	COMPONENT CODE (14)	COMP. SUBCODE (15)	VALVE SUBCODE (16)
MC	X	Z	INSTRUU	S	Z

(17) LER/RO REPORT NUMBER: EVENT YEAR 81 | — | SEQUENTIAL REPORT NO. 007 | / | OCCURRENCE CODE 03 | REPORT TYPE L | — | REVISION NO. 0

ACTION TAKEN (18)	FUTURE ACTION (19)	EFFECT ON PLANT (20)	SHUTDOWN METHOD (21)	HOURS (22)	ATTACHMENT SUBMITTED (23)	NPRD-4 FORM SUB. (24)	PRIME COMP. SUPPLIER (25)	COMPONENT MANUFACTURER (26)
E	Z	Z	Z	0000	N	N	N	G080

CAUSE DESCRIPTION AND CORRECTIVE ACTIONS (27)

The cause of this event was instrument drift. The setpoint was adjusted to 95 mr/hr and operability verified. Area radiation monitors will continue to be calibrated quarterly per DCP 2700-12. No further corrective action is deemed necessary.

FACILITY STATUS (28)	% POWER (29)	OTHER STATUS (30)	METHOD OF DISCOVERY (31)	DISCOVERY DESCRIPTION (32)
E	099	NA	B	Surveillance Test

ACTIVITY RELEASED (33)	CONTENT OF RELEASE (34)	AMOUNT OF ACTIVITY (35)	LOCATION OF RELEASE (36)
Z	Z	NA	NA

PERSONNEL EXPOSURES

NUMBER (37)	TYPE (38)	DESCRIPTION (39)
000	Z	NA

PERSONNEL INJURIES

NUMBER (40)	DESCRIPTION (41)
000	NA

LOSS OF OR DAMAGE TO FACILITY (43)

TYPE (42)	DESCRIPTION
Z	NA

PUBLICITY

ISSUED (44)	DESCRIPTION (45)
Z	NA

NRC USE ONLY

8103240450

NAME OF PREPARER: Sang J. Rhee PHONE: 815-942-2920 Ext 526

GPO 917-926