



July 12, 2017

Materials Licensing Branch
U.S. Nuclear Regulatory Commission, Region III
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

13-35179-02

Dear Mr. Tran,

This letter is a response to your phone conversation with Matthew Trusner on July 11, 2017.

Zevacor has multiple area monitors placed in close proximity to the outside of the hot cells and are used and maintained according to manufacturer's recommendations.

In the event that an area monitor malfunctions inside of the hot cell, Zevacor will utilize internal procedures to remove the malfunctioning area monitor and place it into an appropriate decay-in-storage area.

Please contact me at the phone number or email address below with any questions.

Sincerely,

A handwritten signature in black ink, appearing to read "J. A. Zehner", written over a horizontal line.

John A. Zehner
COO, RSO
Zevacor Molecular
317-459-8315
jzehner@zevacor.com

Tran, Frank

From: Matthew Trusner <mtrusner@zevacor.com>
Sent: Wednesday, July 12, 2017 3:16 PM
To: Tran, Frank
Subject: [External_Sender] Follow Up Letter about Area Monitors - Zevacor
Attachments: 170712 - Follow Up Letter to NRC about Area Monitors - Submitted via email.pdf

Mr. Tran,

Please see the attached letter regarding our phone call conversation that we had on 7/11/17.

If you have any additional questions, please don't hesitate to contact me.

Thank You,
Matthew Trusner
Zevacor
217-433-7902
mtrusner@zevacor.com