

International Paper/Newport Mill
2585 E. 200 N.
Cayuga, IN 47928
Date: June 12, 2017

To: Director of Nuclear Material Safety & Safeguards
ATTN: GLTS
U.S. Nuclear Regulatory Commission
Washington DC 20555-0001

This report is being submitted pursuant to 10 CFR 31.5(c)(5), or equivalent State Regulatory Code, information required to be reported in the case of a Generally Licensed Device damage or shutter failure.

Reportable Event: *Malfunctioning shutter control mechanism*
GL Device: *Valmet USA, Inc., Model BWM-T. Serial Number*
26941419

Remedial Action: *Device was removed from operation and replaced with backup*
spare, Serial Number 19851401

The byproduct material has been transferred to this device at this same location.

Transfer Date: June 7, 2017

Sincerely,



Korey Balding
Process Specialist
765-492-3341 x211