

MATERIALS LICENSE

Amendment No. 01

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 40 and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

<p style="text-align: center;">Licensee</p> <p>1. Valley Forge Laboratories, Inc.</p> <p>2. 6 Berkeley Road Devon, Pennsylvania 19333</p>	<p>In accordance with application dated August 7, 1987,</p> <p>3. License number 37-20584-01 is amended in its entirety to read as follows:</p> <hr/> <p>4. Expiration date January 31, 1993</p> <hr/> <p>5. Docket or Reference No. 030-19945</p>
---	--

<p>6. Byproduct, source, and/or special nuclear material</p> <p>A. Cesium 137</p> <p>B. Americium 241</p>	<p>7. Chemical and/or physical form</p> <p>A. Sealed sources (Troxler Drawing No. A-102112)</p> <p>B. Sealed source (Troxler Drawing No. A-102451)</p>	<p>8. Maximum amount that licensee may possess at any one time under this license</p> <p>A. Not to exceed 10 millicuries per source</p> <p>B. Not to exceed 50 millicuries per source</p>
---	--	---

9. Authorized use

A. and B. For use in Troxler Model Series 3400 gauges to measure properties of materials.

CONDITIONS

- 10. Licensed material may be used at 6 Berkeley Road, Devon, Pennsylvania and at temporary job sites of the licensee anywhere in the United States where the U.S. Nuclear Regulatory Commission maintains jurisdiction for regulating the use of licensed material.
- 11. Licensed material shall be used by, or under the supervision and in the physical presence of, Ward McMasters or individuals who have been trained as specified in application dated August 7, 1987. The licensee shall maintain records of individuals designated as users.
- 12. Sealed sources or detector cells containing licensed material shall not be opened or sources removed from source holders or detector cells by the licensee.
- 13. A. Any sealed sources or detector cells specified in Items 7.A. and 7.B. shall be tested for leakage and/or contamination at intervals not to exceed 6 months. Any source or detector cell received from another person which is not accompanied by a certificate indicating that a test was performed within 6 months before the transfer shall not be put into use until tested.

"OFFICIAL RECORD COPY"

ML18

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License number	37-20584-01
Docket or Reference number	030-19945
Amendment No. 01	

(13. continued)

CONDITIONS

- B. Any sealed source or detector cell in storage and not being used need not be tested. When the source or detector cell is removed from storage for use or transfer to another person, it shall be tested before use or transfer.
 - C. The test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. If the test reveals the presence of 0.005 microcurie or more of removable contamination, the source or detector cell shall be removed from service and decontaminated, repaired, or disposed of in accordance with Commission regulations. A report shall be filed within 5 days of the date the leak test result is known with the U.S. Nuclear Regulatory Commission, Region I, ATTN: Chief, Nuclear Materials Safety and Safeguards Branch, 631 Park Avenue, King of Prussia, Pennsylvania 19406. The report shall specify the source involved, the test results, and corrective action taken. Records of leak test results shall be kept in units of microcuries and shall be maintained for inspection by the Commission. Records may be disposed of following Commission inspection.
 - D. The licensee is authorized to collect leak test samples for analysis by Troxler Electronics Laboratories, Inc. or tests for leakage and/or contamination shall be performed by persons specifically licensed by the Commission or an Agreement State to perform such services.
14. The licensee may transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material".
15. The licensee shall conduct a physical inventory every 6 months to account for all sources and/or devices received and possessed under the license. Records of inventories shall be maintained for 2 years from the date of each inventory.
16. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents including any enclosures, listed below. The Nuclear Regulatory Commission's regulations shall govern unless the statements, representations and procedures in the licensee's application and correspondence are more restrictive than the regulations.
- A. Application dated August 7, 1987

For the U.S. Nuclear Regulatory Commission

Date 14 SEP 1987

Original Signed By:
 By John E. Glenn
 Nuclear Materials Safety and
 Safeguards Branch, Region I
 King of Prussia, Pennsylvania 19406