



CONVERSATION RECORD

06/14/2017

NAME OF PERSON(S) CONTACTED OR IN CONTACT WITH YOU

Mark Beanblossom

DATE OF CONTACT

06/14/2017

TYPE OF CONVERSATION

E-MAIL

TELEPHONE

INCOMING

OUTGOING

E-MAIL ADDRESS

m.beanblossom@ampmedicalphysics.com

TELEPHONE NUMBER

(636) 798-2250

ORGANIZATION

Progress West Healthcare Center

DOCKET NUMBER(S)

030-37397

LICENSE NUMBER(S)

24-32642-01

CONTROL NUMBER(S)

592915

SUBJECT

Additional Information Request for Renewal Application

SUMMARY

Dear Mr. Beanblossom,

After initial review of your renewal application dated January 19, 2017, additional information and/or clarification is needed to process the request. Please see the following requested information in preparation for a phone call scheduled for June 15, 2017:

1) Page 1, Type of Application (Page 1 missing):

Page 1 referenced in the table of contents is not included in your application. Please provide or confirm the use of NRC Form 313 as Page 1.

**Continue on Page 2**

ACTION REQUIRED (IF ANY)

Provide updated renewal application submit in its entirety and on or before the discussed date.

**Continue on Page 3**

NAME OF PERSON DOCUMENTING CONVERSATION

Daniel Strohmeyer

SIGNATURE

CONVERSATION RECORD (continued)

SUMMARY: (Continued from page 1)

2) Items #5 and 6, Radioactive Material Use (Page 5-1):

a) In the Radioactive Material Use table there is a line item for 10 CFR 35.300 and Iodine-131. In the 10 CFR 35.300 item there is a request for "As Needed" for the activity limit. Please note that a maximum activity limit must be provided. Additionally, there should be only one line item for both 10 CFR 35.300 and Iodine-131 (i.e., I-131 should be included in 10 CFR 35.300). Also, confirm if there should be an activity limit change.

b) Provide the units for the activity listed in 10 CFR 31.11.

3) Item #7.1, Delegation of Authority (Page 7-2):

The signatures on the document do not have name and title. Please type the full name and title of all signatures.

4) Item #7.2, Authorized Users (AU) for Medical Use (Page 7-3):

We noted that at least one AU on the renewal application are not the same as the current license. Dr Halls-Daniel whose authorization changed from 10 CFR 35.300 (<33 mCi) to 10 CFR 35.300. Please confirm if there are changes being requested for Dr. Halls-Daniel.

5) Item #9.1, Facility Description (Page 9-1):

Provide a more detailed map. See NUREG-1556 Vol. 9, Rev. 2 Section 8.16 for additional details.

6) Item 10, Radiation Safety Program:

This section is needing commitments contained in our guidance. Please revise this section to be consistent with NUREG-1556 Vol. 9 Rev. 2.

Once the issues are addressed, please send the revised renewal application with signed and dated cover letter on or before June 30, 2017. You may submit your response with the reference control number 592915 via email (letter must be signed and scanned into pdf format). We will continue the review upon receipt of the requested information. If you have any question, please feel free to contact me at 630.829.9689 or [daniel.strohmeyer@nrc.gov](mailto:daniel.strohmeyer@nrc.gov).

Respectfully,

Daniel C. Strohmeyer, CHP  
Materials Licensing Branch  
US NRC Region III  
630.829.9689