



CONVERSATION RECORD

7/5/17

NAME OF PERSON(S) CONTACTED OR IN CONTACT WITH YOU

Stojan Trajkovski

DATE OF CONTACT

06/19/2017

TYPE OF CONVERSATION

E-MAIL

TELEPHONE

INCOMING

OUTGOING

E-MAIL ADDRESS

Stojan.Trajkovski@franciscanalliance.org

TELEPHONE NUMBER

(219) 922-4274

ORGANIZATION

Franciscan Health Munster

DOCKET NUMBER(S)

030-36594

LICENSE NUMBER(S)

13-32519-01

CONTROL NUMBER(S)

594657

SUBJECT

Conversation Record - Request for Additional Information

SUMMARY

See attached email

**Continue on Page 2**

ACTION REQUIRED (IF ANY)

**Continue on Page 3**

NAME OF PERSON DOCUMENTING CONVERSATION

MAGDALENA GRYGALAK

SIGNATURE

Magdalena B. Grygala

## **Gryglak, Magdalena**

---

**From:** Gryglak, Magdalena  
**Sent:** Monday, June 19, 2017 11:28 AM  
**To:** 'Stojan.Trajkovski@franciscanalliance.org'  
**Subject:** NRC License no. 13-32519-01 Amendment Request

Mr. Trajkovski,

After review of your request to amend NRC license no. 13-32519-01 for use of Florine 18, additional information is needed to complete my review. A number of deficiencies was identified during review of the facility diagram. Specifically, as discussed with you on June 19, 2017, please provide the following:

- 1) Provide enlarged facility diagram/s describing the rooms of interest (the nuclear medicine suite: hot lab, PET/CT room, PET Injection rooms, pet cool down room, bathroom) as well as areas immediately above, below, adjacent to the rooms specified above;
- 2) Verify the correct layout of the rooms of interest and provide description (e.g. sinks, waste container, work bench, instrumentation etc.);
- 3) Verify room dimensions used in the calculations are correct; indicate those on the facility diagram;
- 4) Clarify the discrepancies in room numbers in the facility diagram and the room numbers as indicated in the shielding calculations.

Please follow guidance in NUREG 1556, Rev 2, Section 8.16, Facility Diagram, to provide a revised response.

Please submit your response in a signed and dated letter by June 30, 2016. Please do not hesitate to call if you have questions.

Thank you  
Magdalena R. Gryglak  
U.S. NRC Region III  
630-829-9875