

Docket File Information

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Antech Diagnostics 1187 E. Northfield Drive Brownsburg, IN 46112 REPORT NUMBER(S) 2017001	2. NRC/REGIONAL OFFICE Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352
---	---

3. DOCKET NUMBER(S) 030-39009	4. LICENSE NUMBER(S) 13-35388-01	5. DATE(S) OF INSPECTION July 28, 2017
--------------------------------------	---	---

6. INSPECTION PROCEDURES USED 87126	7. INSPECTION FOCUS AREAS 03.01 - 03.07
--	--

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 03620	2. PRIORITY 5	3. LICENSEE CONTACT Meridith Houk, RSO	4. TELEPHONE NUMBER (317) 286-3149
---------------------------------	----------------------	---	---

Main Office Inspection Next Inspection Date: July 28, 2022

Field Office Inspection _____

Temporary Job Site Inspection _____

PROGRAM SCOPE

This was an initial, announced inspection of a small diagnostic laboratory authorized to use iodine-125 for diagnostic testing. The licensee initiated its use of RAM in June 2017. The licensee used prepackaged in-vitro kits for thyroid diagnostic studies on veterinary patients. The licensee employed 15 individuals at its site, with two individuals named as authorized users of RAM. The licensee used its material as needed (at least weekly) based on incoming testing requests by the veterinary customer.

This inspection consisted of interviews with licensee personnel, a review of select records, a tour of the laboratory and waste storage room, and independent measurements. The inspection included observations of surveys, security of byproduct material, use of byproduct material for a diagnostic testing assay.

No violations of NRC requirements were identified during this inspection