

ACCELERATED DISTRIBUTION DEMONSTRATION SYSTEM
 REGULATORY INFORMATION DISTRIBUTION SYSTEM (RIDS)

ACCESSION NBR: 8903220479 DOC. DATE: 89/03/14 NOTARIZED: NO DOCKET #
 FACIL: 50-335 St. Lucie Plant, Unit 1, Florida Power & Light Co. 05000335
 50-389 St. Lucie Plant, Unit 2, Florida Power & Light Co. 05000389
 AUTH. NAME AUTHOR AFFILIATION
 CONWAY, W.F. Florida Power & Light Co.
 RECIPIENT NAME RECIPIENT AFFILIATION
 Document Control Branch (Document Control Desk)

SUBJECT: Responds to 890106 request re unknown spiked samples received on 890113.

DISTRIBUTION CODE: A009D COPIES RECEIVED: LTR 1 ENCL 1 SIZE: 2
 TITLE: OR/Licensing Submittal: Appendix I

NOTES:

	RECIPIENT ID CODE/NAME	COPIES LTTR ENCL	RECIPIENT ID CODE/NAME	COPIES LTTR ENCL
	PD2-2 LA	1 0	PD2-2 PD	5 5
	NORRIS, J	1 1		
INTERNAL:	ACRS	3 3	ARM/DAF/LFMB	1 0
	NRR/DREP DIR10E	1 0	NRR/DREP/RPB 10	2 2
	NRR/PMAS/PMSB12	1 1	NUDOCS-ABSTRACT	1 1
	OGC/HDS2	1 0	<u>REG FILE</u> 01	1 1
	RGN2 DRSS/RPB	2 2		
EXTERNAL:	EG&G SIMPSON, F	2 2	LPDR	1 1
	NRC PDR	1 1		

NOTE TO ALL "RIDS" RECIPIENTS:

PLEASE HELP US TO REDUCE WASTE! CONTACT THE DOCUMENT CONTROL DESK,
 ROOM P1-37 (EXT. 20079) TO ELIMINATE YOUR NAME FROM DISTRIBUTION
 LISTS FOR DOCUMENTS YOU DON'T NEED!

TOTAL NUMBER OF COPIES REQUIRED: LTR 24 ENCL 20

R
I
D
S
/
A
D
D
S

R
I
D
S
/
A
D
D
S



MARCH 14 1989

L-89-101

U. S. Nuclear Regulatory Commission
Attn: Document Control Desk
Washington, D. C. 20555

Gentlemen:


Re: St. Lucie Units 1 and 2
Docket Nos. 50-335 and 50-389
Confirmatory Measurements Program

In response to NRC request dated January 6, 1989, Florida Power & Light Company (FPL) has analyzed the unknown spiked samples received on January 13, 1989, at the St. Lucie Plant Chemistry Department.

As instructed, the attached results were back-decay corrected to January 3, 1989.

Should there be any additional information needed, please contact us.

Very truly yours,


W. F. Conway
Senior Vice President - Nuclear

WFC/JRH/cm

Attachment

cc: Stewart D. Ebnetter, Regional Administrator, Region II, USNRC
Mr. J. B. Kahle, Chief Radiological Effluents and Chemistry
Section, Region II, USNRC
Senior Resident Inspector, USNRC, St. Lucie Plant

8903220479 890314
PDR ADDCK 05000335
P PNU

A009
1/d

ST. LUCIE PLANT
NRC SPIKED SAMPLE RESULTS

<u>Nuclide</u>	<u>micro-Curies/milli-Liter</u>
H - 3	1.37 E-05
Fe - 55	1.53 E-05
Sr - 89	6.28 E-05
Sr - 90	3.4 E-06